

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GROVELAND PD
Time of Inspection: 22:09

Date of Inspection: 04/17/2022

Serial Number: 80-007252
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted	Yes	
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:15421080A1 Exp: 08/05/2023
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.050	0.079	0.198	0.080

Number of Simulators Used: 5

Remarks:

Time-Date changed. *TO MARK CADS/DISPATCH TIME.*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RUSSEL S PENA

RUSSEL S PENA

Signature and Printed Name

04/17/2022
Date