

**From:** [Hendrick Martinez](#)  
**To:** [Melissa A. McLeod](#); [Soto, Israel](#); [PensacolaBAR@flhsmv.gov](mailto:PensacolaBAR@flhsmv.gov)  
**Cc:** [Drew Glass](#)  
**Subject:** June 2022 AI Report  
**Date:** Tuesday, June 21, 2022 9:48:14 AM  
**Attachments:** [DOC031.pdf](#)

---

**CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.**

Good Morning,

Attached are the June 2022 AI Report. There are two due to the first test resulting in Range Exceeded in the Mouth Alcohol Testing part. A new test was conducted and is attached.

If you have questions or concerns, please feel free to contact me. Have a great day and stay safe.

Very Respectfully,

Police Officer Hendrick J Martinez #B118/201  
Uniform Patrol Division / Agency Inspector  
Pensacola Police Department  
711 North Hayne Street  
Pensacola, FL 32501  
Phone Number: 850-435-1901  
Cell Phone Number: 850-637-6642  
Email: [hmartinez@cityofpensacola.com](mailto:hmartinez@cityofpensacola.com)

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PENSACOLA PD  
Time of Inspection: 07:33

Date of Inspection: 06/21/2022

Serial Number: 80-007362  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1408275 Exp: 11/19/2023
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.200	0.081
0.000	0.049	0.079	0.200	0.081

Number of Simulators Used: 5

Remarks:  
A F / M A: Range Exceeded.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



HENDRICK J MARTINEZ

Signature and Printed Name

06/21/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PENSACOLA PD  
Time of Inspection: 08:19

Date of Inspection: 06/21/2022

Serial Number: 80-007362  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1408275 Exp: 11/19/2023
0.000	0.048	0.079	0.200	0.081
0.000	0.048	0.079	0.199	0.081
0.000	0.048	0.079	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

HENDRICK J MARTINEZ

Signature and Printed Name

06/21/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PENSACOLA PD  
Time of Inspection: 07:54

Date of Inspection: 07/27/2022

Serial Number: 80-007362  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1349789 Exp: 07/15/2023
0.000	0.048	0.078	0.199	0.081
0.000	0.049	0.078	0.200	0.081
0.000	0.049	0.079	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PO  \*175/B118

HENDRICK J MARTINEZ

Signature and Printed Name

07/27/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PENSACOLA PD  
Time of Inspection: 06:26

Date of Inspection: 08/16/2022

Serial Number: 80-007362  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1349789 Exp: 07/15/2023
0.000	0.049	0.078	0.200	0.081
0.000	0.049	0.078	0.200	0.081
0.000	0.049	0.078	0.200	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PO [Signature] #175/B18 HENDRICK J MARTINEZ  
Signature and Printed Name

08/16/2022  
Date