

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO  
Time of Inspection: 11:12

Date of Inspection: 04/19/2022

Serial Number: 80-007481  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:13921080A2 Exp: 07/05/2023
0.000	0.053	0.075	0.188 / 0.193	0.083
0.000	0.051	0.076	0.192 / 0.193	0.082
0.000	0.050	0.076	0.194 / 0.194	0.081

Number of Simulators Used: 5

Remarks:

20: Control Outside Tolerance. RESEAL SIMULATOR. AI F MASON #51920 present for inspection.  
FM 51920

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Hailey Dendy*

HALEY M DENDY

Signature and Printed Name

04/19/2022  
Date