

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO  
Time of Inspection: 06:51

Date of Inspection: 04/27/2022

Serial Number: 80-007507  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000				
0.000				
0.000				

Number of Simulators Used: 5

**Remarks:**

INT DET NOT READ ON SECOND SAMPLE, . Non-compliance: .

*Per ran Acetone Sample a second time and again on the second sample there was no interferent detection. It did read it on the 1st & 3rd samples.*

**A TRUE COPY**  
**PAUL BLACKMAN, SHERIFF**  
**HIGHLANDS COUNTY, FLORIDA**  
BY Shon W. Abeln 2184 D.S.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shon W. Abeln 2184 SHON W ABELN  
Signature and Printed Name

04/27/2022  
Date