

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 23:12

Date of Inspection: 01/25/2022

Serial Number: 80-007508
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211543 Exp: 01/05/2023
0.000	0.047	0.078	0.197	0.080
0.000	0.048	0.079	0.198	0.080
0.000	0.048	0.079	0.198	0.080

Dry gas Lot# 22521080A1
Exp: 01/05/2023

Number of Simulators Used: 5

Remarks:

Utilized wrong lot number - M B 6/20/22
Incorrect date due to typo M B 4/20/22

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name MICHAEL D BELLFLOWER

01/25/2022
Date

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 13:31

Date of Inspection: 02/17/2022

Serial Number: 80-007508
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211543 Exp: 01/05/2023
0.000	0.046	0.077	0.194	0.080
0.000	0.047	0.077	0.196	0.080
0.000	0.048	0.078	0.196	0.080

Number of Simulators Used: 5

Dry Gas Lot# 2252108041
Exp: 10/05/2023

Remarks:

Utilized wrong lot number - MB 6/20/22
Incorrect date due to type - MB 4/20/22

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

MICHAEL D BELLFLOWER

02/17/2022
Date

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 11:26

Date of Inspection: 05/17/2022

Serial Number: 80-007508
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211543 Exp: 10/05/2023
0.000	0.047	0.077	0.194	0.079
0.000	0.047	0.077	0.195	0.079
0.000	0.048	0.078	0.197	0.079

Number of Simulators Used: 5

Dry Gas Lot# 22521080A1

Remarks:

Utilized wrong lot number - M.B.G. 6/20/22

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M.B.G. JB

MICHAEL D BELLFLOWER
Signature and Printed Name

05/17/2022
Date

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 08:44

Date of Inspection: 04/30/2022

Serial Number: 80-007508
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211543 Exp: 10/05/2023
0.000	0.047	0.077	0.196	0.079
0.000	0.047	0.077	0.196	0.080
0.000	0.048	0.078	0.197	0.079

Number of Simulators Used: 5

Dry Gas Lot# 22521086A1
Exp: 10/05/2023

Remarks:

Utilized wrong lot number - MBS 4/20/22
Incorrect date due to typo - MBS 4/20/22

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

MICHAEL D BELLFLOWER

04/30/2022
Date