

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 13:28

Date of Inspection: 02/17/2022

Serial Number: 80-007509
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211543 Exp: 01/05/2023
0.000	0.046	0.076	0.192	0.080
0.000	0.046	0.076	0.193	0.080
0.000	0.046	0.077	0.194	0.080

Number of Simulators Used: 5

Remarks:

Utilized wrong lot number - MBG 6/20/22
Incorrect date due to typo - MBG 4/20/22

Dry Gas Lot# 22521080A1
Exp: 10/05/2023

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

MICHAEL D BELLFLOWER

02/17/2022
Date

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 11:13

Date of Inspection: 03/24/2022

Serial Number: 80-007509
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2021	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2021	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211543 Exp: 10/05/2023
0.000	0.046	0.075	0.192	0.079
0.000	0.046	0.076	0.193	0.079
0.000	0.046	0.076	0.194	0.079

Number of Simulators Used: 5

Dry Gas Lot# 22521080A1

Remarks:

Utilized wrong lot number - MB 6/20/22

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

MICHAEL D BELLFLOWER

03/24/2022
Date

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 08:41

Date of Inspection: 04/29/2022

Serial Number: 80-007509
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211543 Exp: 10/05/2023
0.000	0.047	0.076	0.194	0.079
0.000	0.047	0.076	0.194	0.079
0.000	0.047	0.077	0.195	0.080

Number of Simulators Used: 5

Dry Gas Lot# 22521880A1

Remarks:

Utilized wrong lot number - MB 6/20/22

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MICHAEL D BELLFLOWER

Signature and Printed Name

04/29/2022
Date

AMENDED

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 11:52

Date of Inspection: 05/16/2022

Serial Number: 80-007509
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:211543 Exp: 10/05/2023
0.000	0.047	0.077	0.195	0.080
0.000	0.047	0.077	0.196	0.080
0.000	0.048	0.077	0.197	0.080

Number of Simulators Used: 5

Dry Gas lot# 22521080A1

Remarks:

Utilized wrong lot number. MBJ 6/20/22

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MBJ JS

MICHAEL D BELLFLOWER
Signature and Printed Name

05/16/2022
Date