

## Gutschow, Taylor

---

**From:** Gutschow, Taylor  
**Sent:** Friday, December 9, 2022 1:37 PM  
**To:** Damon Laney  
**Cc:** Malcolm Deane  
**Subject:** 80-001003 Sept AI- Noncompliant  
**Attachments:** 80-001003 Form 42.pdf; 80-001003 Affected Breath Tests.pdf

Good morning,

I confirmed with our new Program Manager, but since the repeated Alcohol Free part of the AF/MA Test does not read "Yes" (on the September AI for Intox 1003), it is not a passing Agency Inspection. Please update the "Compliance" portion of the original inspection paperwork to "No". You can also include a remark about what happened and the corrective action you took. Please scan the amended paperwork to me so I can add it to our website.

I've attached a list of affected breath tests for the instrument. You'll want to provide that to your local State Attorney's Office and let them know that the instruments were run on an instrument that didn't have a compliant inspection for the month. If they have any questions, you can provide them with my contact info to discuss further.

### Taylor Gutschow

Department Inspector for the Alcohol Testing Program  
Florida Department of Law Enforcement  
Fort Myers Regional Operations Center  
4700 Terminal Drive, Suite 3  
Fort Myers, FL 33907

Desk #239-335-7500  
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.