

**From:** [Gutschow, Taylor](#)  
**To:** ["Chris Johnston"](#)  
**Cc:** [Graham, Thomas](#)  
**Subject:** 80-007506 Noncompliance in May  
**Date:** Tuesday, June 21, 2022 11:41:00 AM  
**Attachments:** [80-007506 Form 42 May.pdf](#)  
[Affected Breath Tests.pdf](#)

---

Hi,

It appears that during last month's Agency Inspection on Intoxilyzer 80-007506, a passing result was not obtained on the Alcohol Free Subject Test (it is marked NO instead of YES). This renders the instrument noncompliant with Chapter 11D-8, FAC, for the month of May. Please see the attached Form 42 and make the necessary corrections on the original Form 40- Agency Inspection paperwork, including marking it as non-compliant and adding a remark about what occurred during the AF/MA Test.

I've attached a list of affected breath tests. Please distribute the list to your local State Attorney's Office and inform them that the breath tests were run on a non-compliant instrument. If they have any questions, they can contact me to discuss.

Because the instrument has already received a passing June inspection, it can remain in evidential use.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program  
Florida Department of Law Enforcement  
Fort Myers Regional Operations Center  
4700 Terminal Drive, Suite 3  
Fort Myers, FL 33907

Desk #239-335-7500  
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.