



INSTRUMENT PROCESSING SHEET

Agency Perry Police Department

S/N 80-000811

Florida Department of Law Enforcement

Date In 01-27-2022 DI Completion Date 01-31-2022 Ship P/U H/D CMI EE

Intake, Quality Checks, Flow Calibration, Maintenance, and Simulator data tables.

Calibration Adjustment and Department Inspection sections with tables for gauge and simulator data.

Notes/Suggested Service: [Blank lines for notes]

Compliance checkboxes: Instrument Complies with Chapter 11D-8, FAC, etc.

Tech Review / Date and Admin Review / Date with digital signature and timestamp.

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PERRY POLICE DEPT  
Time of Inspection: 11:45

Date of Inspection: 01/31/2022

Serial Number: 80-000811  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.080	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.079

Standard Deviations	0.0004	0.0003	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Israel Soto*

ISRAEL SOTO

Signature and Printed Name

01/31/2022  
Date

# stability checks

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
01/28/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:57
Control Test	0.048	14:57
Air Blank	0.000	14:58
Control Test	0.048	14:59
Air Blank	0.000	14:59
Control Test	0.048	15:00
Air Blank	0.000	15:00
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
01/28/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:02
Control Test	0.078	15:02
Air Blank	0.000	15:03
Control Test	0.078	15:04
Air Blank	0.000	15:04
Control Test	0.079	15:05
Air Blank	0.000	15:05
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Operator's Signature

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
01/28/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:07
Control Test	0.197	15:07
Air Blank	0.000	15:08
Control Test	0.197	15:09
Air Blank	0.000	15:09
Control Test	0.197	15:10
Air Blank	0.000	15:10
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

wet

PERRY POLICE DEPT  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000811  
01/28/2022  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:12
Control Test	0.078	15:13
Air Blank	0.000	15:13
Control Test	0.079	15:14
Air Blank	0.000	15:14
Control Test	0.079	15:14
Air Blank	0.000	15:15
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Operator's Signature

Dry



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2729 Fort Knox Blvd.  
Bldg. 2, Suite 1300  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000811, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000811</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PERRY POLICE DEPT</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/31/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:45</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. Digitally signed by Israel Soto  
Date: 2022.01.31 13:25:51 -05'00'

**Israel Soto**

Date: 2022.01.31  
13:25:51 -05'00'

01/31/2022

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 January 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality