

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000883, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

| | | | |
|-------------------|----------------------|--------------------------------|-------|
| Serial Number: | <u>80-000883</u> | UNCERTAINTY* ± | |
| Owning Agency: | <u>MIAMI-DADE PD</u> | 0.050 g/ 210 L | 0.004 |
| Calibration Date: | <u>07/21/2022</u> | 0.080 g/ 210 L | 0.004 |
| Calibration Time: | <u>11:13</u> | 0.200 g/ 210 L | 0.007 |
| | | 0.080 g/ 210 L Dry Gas Control | 0.005 |

All results are reported in g/210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.
Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Date 07/21/2022


DAVID E REYES-RIVERA,
Department Inspector

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD
Time of Inspection: 11:13

Date of Inspection: 07/21/2022

Serial Number: 80-000883
Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024 | 0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024 | 0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023 |
|----------------------------|---|---|---|---|
| 0.000 | 0.050 | 0.079 | 0.198 | 0.079 |
| 0.000 | 0.050 | 0.079 | 0.200 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.079 | 0.201 | 0.080 |
| 0.000 | 0.050 | 0.079 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.079 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.079 | 0.200 | 0.079 |
| 0.000 | 0.050 | 0.079 | 0.200 | 0.080 |
| 0.000 | 0.050 | 0.079 | 0.199 | 0.079 |
| 0.000 | 0.050 | 0.079 | 0.200 | 0.079 |


| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0003 | 0.0000 | 0.0007 | 0.0004 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

DAVID E REYES-RIVERA

07/21/2022
 Date

| | | | | |
|--------------|---------------|------------------------------|-----------|-----------------|
| Type of Test | Serial Number | Agency | Date | Performed By |
| Stabilities | 80-000883 | Miami-Dade Police Department | 7/21/2022 | DERR <i>MLL</i> |

| 0.05g/210L | 0.08g/210L | 0.20g/210L | DGS 0.08g/210L |
|---|---|---|---|
| <p>MIAMI-DADE PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000883 07/21/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:17 Control Test 0.048 07:18 Air Blank 0.000 07:19 Control Test 0.049 07:19 Air Blank 0.010 07:20 Control Test 0.049 07:21 Air Blank 0.000 07:21</p> <p>Control Test Stats Average 0.0487 Std Dev 0.0006 Rel Std Dev(%) 1.1853</p> <p><i>MLL</i> Operator's Signature</p> | <p>MIAMI-DADE PD Intoxilizer - Alcohol Analyzer Model 8001 SN 80-000883 07/21/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:22 Control Test 0.077 07:23 Air Blank 0.000 07:24 Control Test 0.073 07:24 Air Blank 0.000 07:25 Control Test 0.079 07:26 Air Blank 0.000 07:26</p> <p>Control Test Stats Average 0.0780 Std Dev 0.0010 Rel Std Dev(%) 1.2821</p> <p><i>MLL</i> Operator's Signature</p> | <p>MIAMI-DADE PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000883 07/21/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:28 Control Test 0.197 07:28 Air Blank 0.000 07:29 Control Test 0.197 07:30 Air Blank 0.000 07:30 Control Test 0.198 07:31 Air Blank 0.000 07:31</p> <p>Control Test Stats Average 0.1973 Std Dev 0.0006 Rel Std Dev(%) 0.2925</p> <p><i>MLL</i> Operator's Signature</p> | <p>MIAMI-DADE PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-000883 07/21/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:33 Control Test 0.080 07:33 Air Blank 0.000 07:34 Control Test 0.080 07:34 Air Blank 0.000 07:34 Control Test 0.080 07:35 Air Blank 0.000 07:35</p> <p>Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p><i>MLL</i> Operator's Signature</p> |
| 0.047 to 0.053 <input checked="" type="checkbox"/> | 0.077 to 0.083 <input checked="" type="checkbox"/> | 0.194 to 0.206 <input checked="" type="checkbox"/> | 0.077 to 0.083 <input checked="" type="checkbox"/> |

MIAMI-PAGE.PG
Intoxilyzer - Eticonol Analyzer
Model 8000 SN 80-000883
07/21/2022
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
 SORT(DIFF)) = 6.855
2: Rate (Liters/min) = 15
 SORT(DIFF)) = 11.660
3: Rate (Liters/min) = 30
 SORT(DIFF)) = 20.613
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 703
Rounded Intercept = -673959
Correlation = 0.99835

| Flow Calibration | |
|------------------|------------------------------|
| SN: | 80-000883 |
| Agency: | Miami-Dade Police Department |
| Date: | 7/21/2022 |
| By: | DERR <i>MM</i> |