

INSTRUMENT PROCESSING SHEET

Agency Pasco CSO

S/N 80-001081

Florida Department of
Law Enforcement

Date In 3/7/2022DI Completion Date 3/7/2022☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u>		Quality Checks By <u>TDG</u>		Date <u>3/7/2022</u>		Flow Calibration By _____ Date _____																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>232</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks				Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
		<table border="1" style="width:100%; border-collapse: collapse;"><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr><tr><td>0.050</td><td>MP6286</td><td>202010A 10/05/2022</td></tr><tr><td>0.080</td><td>MP6287</td><td>202010B 10/05/2022</td></tr><tr><td>0.200</td><td>MP6288</td><td>202010D 10/06/2022</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A 10/05/2022	0.080	MP6287	202010B 10/05/2022	0.200	MP6288	202010D 10/06/2022	0.080 DGS	N/A	AG115904 06/08/2023			Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____														
Simulator	Serial #	Lot #/Exp																																	
0.050	MP6286	202010A 10/05/2022																																	
0.080	MP6287	202010B 10/05/2022																																	
0.200	MP6288	202010D 10/06/2022																																	
0.080 DGS	N/A	AG115904 06/08/2023																																	
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																															
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28199</u>																															
<table border="1" style="width:100%; border-collapse: collapse;"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1020</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2021-B</u> Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks				<table border="1" style="width:100%; border-collapse: collapse;"><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																		
0.000	MP6284																																		
Interferent	MP6285																																		
0.050	MP6286																																		
0.080	MP6287																																		
0.200	MP6288																																		
<table border="1" style="width:100%; border-collapse: collapse;"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <table border="1" style="width:100%; border-collapse: collapse;"><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></table>				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____						
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																		
Notes/Suggested Service: <u>Instrument has power-up issue that causes it to shut off during the start-up sequence. Was eventually able to turn on after multiple attempts. Recommended sending to repair after the Department Inspection. (TDG)</u> _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
				 Date: 2022.03.07 15:57:11 -05'00' 2022.03.07 16:01:40 Tech Review / Date Admin Review / Date																															

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001081	Pasco CSO	03/07/2022	TDG ML

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																																																																																																																																	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083																																																																																																																																																	
✓		✓		✓		✓																																																																																																																																																	
PASCO COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/07/2022 Software: 8100.27 SN 80-001081 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr><tr><td>Control Test</td><td>0.048</td><td>11:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:09</td></tr><tr><td>Control Test</td><td>0.048</td><td>11:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:10</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0483</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr></tbody></table> Operator's Signature <i>ML</i>		Test	g/210L	Time	Air Blank	0.000	11:07	Control Test	0.049	11:07	Air Blank	0.000	11:08	Control Test	0.048	11:09	Air Blank	0.000	11:09	Control Test	0.048	11:10	Air Blank	0.000	11:10	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		PASCO COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/07/2022 Software: 8100.27 SN 80-001081 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table> Operator's Signature <i>ML</i>		Test	g/210L	Time	Air Blank	0.000	11:14	Control Test	0.078	11:14	Air Blank	0.000	11:15	Control Test	0.079	11:16	Air Blank	0.000	11:16	Control Test	0.079	11:17	Air Blank	0.000	11:17	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		PASCO COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/07/2022 Software: 8100.27 SN 80-001081 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr><tr><td>Control Test</td><td>0.199</td><td>11:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr><tr><td>Control Test</td><td>0.199</td><td>11:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr><tr><td>Control Test</td><td>0.198</td><td>11:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1987</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr></tbody></table> Operator's Signature <i>ML</i>		Test	g/210L	Time	Air Blank	0.000	11:20	Control Test	0.199	11:21	Air Blank	0.000	11:21	Control Test	0.199	11:22	Air Blank	0.000	11:23	Control Test	0.198	11:23	Air Blank	0.000	11:24	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		PASCO COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 03/07/2022 Software: 8100.27 SN 80-001081 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr><tr><td>Control Test</td><td>0.081</td><td>11:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr><tr><td>Control Test</td><td>0.082</td><td>11:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr><tr><td>Control Test</td><td>0.082</td><td>11:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0817</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7070</td><td></td></tr></tbody></table> Operator's Signature <i>ML</i>		Test	g/210L	Time	Air Blank	0.000	11:26	Control Test	0.081	11:27	Air Blank	0.000	11:27	Control Test	0.082	11:28	Air Blank	0.000	11:28	Control Test	0.082	11:28	Air Blank	0.000	11:29	Control Test Stats			Average	0.0817		Std Dev	0.0006		Rel Std Dev(%)	0.7070	
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	11:07																																																																																																																																																					
Control Test	0.049	11:07																																																																																																																																																					
Air Blank	0.000	11:08																																																																																																																																																					
Control Test	0.048	11:09																																																																																																																																																					
Air Blank	0.000	11:09																																																																																																																																																					
Control Test	0.048	11:10																																																																																																																																																					
Air Blank	0.000	11:10																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0483																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	1.1945																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	11:14																																																																																																																																																					
Control Test	0.078	11:14																																																																																																																																																					
Air Blank	0.000	11:15																																																																																																																																																					
Control Test	0.079	11:16																																																																																																																																																					
Air Blank	0.000	11:16																																																																																																																																																					
Control Test	0.079	11:17																																																																																																																																																					
Air Blank	0.000	11:17																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0787																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.7339																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	11:20																																																																																																																																																					
Control Test	0.199	11:21																																																																																																																																																					
Air Blank	0.000	11:21																																																																																																																																																					
Control Test	0.199	11:22																																																																																																																																																					
Air Blank	0.000	11:23																																																																																																																																																					
Control Test	0.198	11:23																																																																																																																																																					
Air Blank	0.000	11:24																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.1987																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.2906																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	11:26																																																																																																																																																					
Control Test	0.081	11:27																																																																																																																																																					
Air Blank	0.000	11:27																																																																																																																																																					
Control Test	0.082	11:28																																																																																																																																																					
Air Blank	0.000	11:28																																																																																																																																																					
Control Test	0.082	11:28																																																																																																																																																					
Air Blank	0.000	11:29																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0817																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.7070																																																																																																																																																						

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO

Serial Number: 80-001081

Time of Inspection: 14:38

Date of Inspection: 03/07/2022

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#: 202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#: 202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG115904 Exp: 06/08/2023
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.199	0.081
0.000	0.048	0.078	0.200	0.081
0.000	0.048	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.079	0.199	0.080
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.080	0.199	0.079
0.000	0.049	0.079	0.199	0.080

Standard Deviations	0.0005	0.0005	0.0003	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks: A small crimp in the external printer adapter prevented physical connection to the instrument. Due to power issue, did not want to move instrument to other printer. Reprinted from COBRA. Crimp is on FDLE adapter not the instrument.
TNG 3/7/2022

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/07/2022

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001081, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001081</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PASCO COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/07/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:38</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/07/2022

Date



TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality