



Agency **Bradford Co. SO**

S/N 80-001291

Date In 12/21/2021 DI Completion Date 1/3/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

FDLE/ATP Form 48 January 2021
Issuing Authority: Alcohol Testing Program

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BRADFORD COUNTY SO
Time of Inspection: 13:36

Date of Inspection: 01/03/2022

Serial Number: 80-001291
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG113403 Exp: 05/14/2023
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.050	0.081	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.079


Standard Deviations	0.0004	0.0003	0.0000	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name RICHARD A WILLIAMS

01/03/2022
Date

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001291
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.049	09:59
Air Blank	0.000	09:59
Control Test	0.050	10:00
Air Blank	0.000	10:00
Control Test	0.050	10:01
Air Blank	0.000	10:01
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

05

PAW

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001291
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:03
Control Test	0.080	10:04
Air Blank	0.000	10:04
Control Test	0.080	10:05
Air Blank	0.000	10:05
Control Test	0.081	10:06
Air Blank	0.000	10:06
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

08

wet

PAW

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001291
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:08
Control Test	0.200	10:09
Air Blank	0.000	10:09
Control Test	0.200	10:10
Air Blank	0.000	10:11
Control Test	0.199	10:11
Air Blank	0.000	10:12
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

20

PAW

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001291
01/03/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:17
Control Test	0.079	10:18
Air Blank	0.000	10:18
Control Test	0.079	10:19
Air Blank	0.000	10:19
Control Test	0.080	10:19
Air Blank	0.000	10:20
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

08
DGS

PAW

Operator's Signature

Stability Checks

80-001291

01/03/22

PAW



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2729 Fort Knox Blvd.
Bldg. 2, Suite 1300
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001291, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001291</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>BRADFORD COUNTY SO</u>	0.050 g/ 210 L	0.005
Calibration Date:	<u>01/03/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:36</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Richard A
Williams
Digitally signed by Richard A
Williams
Date: 2022.01.04 14:28:57
-05'00'

01/03/2022

Date

RICHARD A WILLIAMS,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



Agency **Bradford County Sheriff's Office**

S/N 80-001291

Date In 12/12/2022 DI Completion Date 12/15/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

FDLE/ATP Form 48 January 2022
Issuing Authority: Alcohol Testing Program

Page 1 of 1

STABILITY CHECKS

80-001291

12/14/2022

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001291
12/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:59
Control Test	0.048	13:59
Air Blank	0.000	14:00
Control Test	0.048	14:00
Air Blank	0.000	14:01
Control Test	0.047	14:02
Air Blank	0.000	14:02
Control Test Stats		
Average	0.0477	
Std Dev	0.0006	
Rel Std Dev(%)	1.2112	

0.05 g/210L

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001291
12/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:04
Control Test	0.077	14:05
Air Blank	0.000	14:05
Control Test	0.077	14:06
Air Blank	0.000	14:06
Control Test	0.077	14:07
Air Blank	0.000	14:07
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.08 g/210L

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001291
12/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:09
Control Test	0.199	14:10
Air Blank	0.000	14:10
Control Test	0.199	14:11
Air Blank	0.000	14:11
Control Test	0.198	14:12
Air Blank	0.000	14:12
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

0.20 g/210L

Operator's Signature

BRADFORD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001291
12/14/2022
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:14
Control Test	0.078	14:14
Air Blank	0.000	14:15
Control Test	0.077	14:15
Air Blank	0.000	14:15
Control Test	0.077	14:16
Air Blank	0.000	14:16
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

D65 0.08 g/210L

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BRADFORD COUNTY SO
Time of Inspection: 11:11

Date of Inspection: 12/15/2022

Serial Number: 80-001291
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.077	0.199	0.077
0.000	0.048	0.077	0.200	0.077
0.000	0.048	0.076	0.199	0.077
0.000	0.048	0.077	0.199	0.077
0.000	0.048	0.077	0.199	0.077
0.000	0.048	0.077	0.199	0.077
0.000	0.047	0.076	0.199	0.077
0.000	0.048	0.077	0.198	0.076
0.000	0.048	0.077	0.199	0.077
0.000	0.048	0.077	0.198	0.077
Standard Deviations	0.0003	0.0004	0.0005	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

12/15/2022
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001291, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001291</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>BRADFORD COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/15/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:11</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/15/2022

Date

Phil Nicodemo

Digitally signed by Phil
Nicodemo
Date: 2022.12.15 14:51:00 -05'00'

PHIL NICODEMO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1