

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001344, manufactured by CML, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001344</u>	UNCERTAINTY* \pm
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/26/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:07</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/26/2022

Date


DAVID E REYES-RIVERA,
Department Inspector

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 12:07

Date of Inspection: 07/26/2022

Serial Number: 80-001344
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.078	0.199	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.050	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.050	0.078	0.198	0.078

Standard Deviations	0.0004	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

07/26/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Post Stabilities	80-001344	Sarasota County Sheriff's Office	7/26/2022	DERR <i>MLL</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>SARASOTA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001344 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:37 Control Test 0.050 09:37 Air Blank 0.000 09:39 Control Test 0.049 09:39 Air Blank 0.000 09:39 Control Test 0.049 09:40 Air Blank 0.000 09:40 Control Test 0.000 09:40</p> <p>Control Test Stats Average 0.0493 Std Dev 0.0006 Rel Std Dev(%) 1.1703</p> <p>Operator's Signature <i>MLL</i></p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>SARASOTA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001344 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:42 Control Test 0.079 09:43 Air Blank 0.000 09:44 Control Test 0.080 09:44 Air Blank 0.000 09:45 Control Test 0.078 09:45 Air Blank 0.000 09:46 Control Test 0.000 09:46</p> <p>Control Test Stats Average 0.0790 Std Dev 0.0010 Rel Std Dev(%) 1.2658</p> <p>Operator's Signature <i>MLL</i></p>	<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>SARASOTA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001344 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:48 Control Test 0.200 09:48 Air Blank 0.000 09:49 Control Test 0.199 09:50 Air Blank 0.000 09:50 Control Test 0.199 09:51 Air Blank 0.000 09:51 Control Test 0.000 09:51</p> <p>Control Test Stats Average 0.1993 Std Dev 0.0006 Rel Std Dev(%) 0.2895</p> <p>Operator's Signature <i>MLL</i></p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>SARASOTA COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 SN 80-001344 07/26/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:53 Control Test 0.078 09:53 Air Blank 0.000 09:54 Control Test 0.079 09:54 Air Blank 0.000 09:54 Control Test 0.079 09:55 Air Blank 0.000 09:55 Control Test 0.000 09:55</p> <p>Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339</p> <p>Operator's Signature <i>MLL</i></p>

SARASOTA COUNTY SO
Intelligence - Alcohol Analyzer
Model 8050
SN 86-001344
08-31-25

AUTO Calibration
New Power Res Value = 57
Auto Range Res Value = 57

Soil Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/L ****
Samples Taken = 4, Discarded = 1
Sum To = 12587, Sum To = 12587

Channel 1 >>>>
Sample #1 = 0.0000 (% Abs Ref)
Sample #2 = 0.0000 (% Abs Ref)
Sample #3 = 0.0000 (% Abs Ref)
Sample #4 = 0.0000 (% Abs Ref)
Avg % Abs = 0.0000 (0.0000)
STD Dev = 0.0000 (0.0000)
REL STD Dev = 0.0000 (0.0000)

Channel 2 >>>>
Sample #1 = 0.1200 (% Abs Ref)
Sample #2 = 0.1200 (% Abs Ref)
Sample #3 = 0.0950 (% Abs Ref)
Sample #4 = 0.0860 (% Abs Ref)
Avg % Abs = 0.0940 (0.0187)
STD Dev = 0.0075 (0.0145)
REL STD Dev = 8.032 (77.694)

Soil Value = 0.030 g/210L ***
Fit Value = 0.1905 mg/L ****
Samples Taken = 4, Discarded = 1
Sum To = 12581, Sum To = 12581

Channel 1 >>>>
Sample #1 = 0.0350 (% Abs Ref)
Sample #2 = 0.0400 (% Abs Ref)
Sample #3 = 0.0130 (% Abs Ref)
Sample #4 = 0.0400 (% Abs Ref)
Avg % Abs = 0.0823 (0.0123)
STD Dev = 0.0153 (0.0147)
REL STD Dev = 1.862 (119.532)

Channel 2 >>>>
Sample #1 = 1.5160 (% Abs Ref)
Sample #2 = 1.4690 (% Abs Ref)
Sample #3 = 1.4970 (% Abs Ref)
Sample #4 = 1.4940 (% Abs Ref)
Avg % Abs = 1.4943 (0.0143)
STD Dev = 0.0046 (0.0012)
REL STD Dev = 0.309 (8.056)

Soil Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/L ****
Samples Taken = 4, Discarded = 1
Sum To = 12576, Sum To = 12576

Channel 1 >>>>
Sample #1 = 1.0990 (% Abs Ref)
Sample #2 = 1.0990 (% Abs Ref)
Sample #3 = 1.0990 (% Abs Ref)
Sample #4 = 1.0990 (% Abs Ref)
Avg % Abs = 1.0990 (0.0000)
STD Dev = 0.0000 (0.0000)
REL STD Dev = 0.341 (330.454)

Channel 2 >>>>
Sample #1 = 3.5570 (% Abs Ref)
Sample #2 = 3.5510 (% Abs Ref)
Sample #3 = 3.5500 (% Abs Ref)
Sample #4 = 3.5540 (% Abs Ref)
Avg % Abs = 3.5533 (0.0123)
STD Dev = 0.0136 (0.0065)
REL STD Dev = 0.384 (52.755)

Soil Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/L ****
Samples Taken = 4, Discarded = 1
Sum To = 12575, Sum To = 12575

Channel 1 >>>>
Sample #1 = 3.5190 (% Abs Ref)
Sample #2 = 3.5110 (% Abs Ref)
Sample #3 = 2.9960 (% Abs Ref)
Sample #4 = 3.6170 (% Abs Ref)
Avg % Abs = 3.6583 (0.0140)
STD Dev = 0.0108 (0.0171)
REL STD Dev = 0.300 (122.057)

Channel 2 >>>>
Sample #1 = 6.8170 (% Abs Ref)
Sample #2 = 6.7910 (% Abs Ref)
Sample #3 = 6.7740 (% Abs Ref)
Sample #4 = 6.7870 (% Abs Ref)
Avg % Abs = 6.7840 (0.0110)
STD Dev = 0.0089 (0.0122)
REL STD Dev = 0.131 (57.931)

Soil Value = 0.300 g/210L ***
Fit Value = 1.4206 mg/L ****
Samples Taken = 4, Discarded = 1
Sum To = 12573, Sum To = 12573

Channel 1 >>>>
Sample #1 = 5.3100 (% Abs Ref)
Sample #2 = 5.3080 (% Abs Ref)
Sample #3 = 5.2850 (% Abs Ref)
Sample #4 = 5.2890 (% Abs Ref)
Avg % Abs = 5.2873 (0.0057)
STD Dev = 0.0196 (0.0172)
REL STD Dev = 0.371 (67.069)

Channel 2 >>>>
Sample #1 = 9.8780 (% Abs Ref)
Sample #2 = 9.8470 (% Abs Ref)
Sample #3 = 9.8580 (% Abs Ref)
Sample #4 = 9.8670 (% Abs Ref)
Avg % Abs = 9.8607 (0.0117)
STD Dev = 0.0118 (0.0025)
REL STD Dev = 0.120 (3.512)

Soil Value = 0.400 g/210L ***
Fit Value = 0.9524 mg/L ****
Samples Taken = 4, Discarded = 1
Sum To = 12573, Sum To = 12573

Channel 1 >>>>
Sample #1 = 3.5190 (% Abs Ref)
Sample #2 = 3.5110 (% Abs Ref)
Sample #3 = 2.9960 (% Abs Ref)
Sample #4 = 3.6170 (% Abs Ref)
Avg % Abs = 3.6583 (0.0140)
STD Dev = 0.0108 (0.0171)
REL STD Dev = 0.300 (122.057)

AUTO CAL DATA *****
Channel 1 >>>>
Soil Val = 0.0000 mg/L or 0.000 g/210L
% Abs = 0.095
Std Dev = 0.01 Rel Std Dev = 10.84
Soil Val = 0.1905 mg/L or 0.040 g/210L
% Abs = 0.822
Std Dev = 0.02 Rel Std Dev = 1.86
Soil Val = 0.4762 mg/L or 0.100 g/210L
% Abs = 1.885

Std Dev = 0.01 Rel Std Dev = 0.34
Soil Val = 0.9524 mg/L or 0.200 g/210L
% Abs = 3.608
Std Dev = 0.01 Rel Std Dev = 0.38
Soil Val = 1.4206 mg/L or 0.300 g/210L
% Abs = 5.287
Std Dev = 0.02 Rel Std Dev = 0.37
Zero Order Coef = -263.38
First Order Coef = 2622.06
Second Order Coef = 24.60
Standard Deviation = 5.114029

Channel 2 >>>>
Soil Val = 0.0000 mg/L or 0.000 g/210L
% Abs = 0.094
Std Dev = 0.01 Rel Std Dev = 0.03
Soil Val = 0.1905 mg/L or 0.040 g/210L
% Abs = 1.494
Std Dev = 0.00 Rel Std Dev = 0.31
Soil Val = 0.4762 mg/L or 0.100 g/210L
% Abs = 3.558
Std Dev = 0.01 Rel Std Dev = 0.36
Soil Val = 0.9524 mg/L or 0.200 g/210L
% Abs = 5.784
Std Dev = 0.01 Rel Std Dev = 0.13
Soil Val = 1.4206 mg/L or 0.300 g/210L
% Abs = 9.861
Std Dev = 0.01 Rel Std Dev = 0.12

Zero Order Coef = -123.69
First Order Coef = 1336.17
Second Order Coef = 12.68
Standard Deviation = 2.121642

Optical Calibration
SN: 80-001344
Agency: Sarasota CSO
Date: 7/26/2022
Quadratic Fit: +/- 0.002g/210L
By: DERR

Solution Stats Quadratic Fit Chan 1:
Act Fit Residual:
g/210L g/210L g/210L
0.000 0.000 -0.000
0.040 0.040 0.001
0.100 0.100 -0.000
0.200 0.200 -0.000
0.300 0.300 0.000

Soil Value = 0.000 g/210L ***
Fit Value = 0.3910 mg/L ****
Samples Taken = 4, Discarded = 1
Channel 1 >>>>
Sample #1 = 3144.00
Sample #2 = 3069.00
Sample #3 = 3196.00
Sample #4 = 3076.00
Average Result = 3113.6667
STD Dev = 71.3886
REL STD Dev = 2.293

Channel 2 >>>>
Sample #1 = 3432.00
Sample #2 = 3426.00
Sample #3 = 3435.00
Sample #4 = 3436.00
Average Result = 3432.3333
STD Dev = 5.5076
REL STD Dev = 0.160
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1018
3 um H2O Adjust (mg/L*10,000) = 595
9 um H2O Adjust (mg/L*10,000) = 377
***** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 2:
Act Fit Residual:
g/210L g/210L g/210L
0.000 0.000 -0.000
0.040 0.040 0.001
0.100 0.100 -0.000
0.200 0.200 -0.000
0.300 0.300 0.000

Solution Stats Quadratic Fit Chan 1:
Act Fit Residual:
g/210L g/210L g/210L
0.000 0.000 -0.000
0.040 0.040 -0.001
0.100 0.100 -0.001
0.200 0.200 0.001
0.300 0.300 -0.000

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001344	Sarasota County Sheriff's Office	7/26/2022	DERR <i>DERR</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>Ver 1.0 7/26/2022</p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p>	<p>0.194 to 0.206 <input checked="" type="checkbox"/></p>	<p>0.077 to 0.083 <input checked="" type="checkbox"/></p>
<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/26/2022 SN 80-001344 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:34 Control Test 0.048 07:35 Air Blank 0.000 07:35 Control Test 0.048 07:36 Air Blank 0.000 07:36 Control Test 0.045 07:37 Air Blank 0.001 07:38 Control Test Stats Average 0.0473 Std Dev 0.0012 Rel Std Dev(%) 2.4395</p> <p>Operator's Signature <i>DERR</i></p>	<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/26/2022 SN 80-001344 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:40 Control Test 0.077 07:40 Air Blank 0.000 07:41 Control Test 0.076 07:42 Air Blank 0.000 07:42 Control Test 0.076 07:43 Air Blank 0.000 07:43 Control Test Stats Average 0.0763 Std Dev 0.0006 Rel Std Dev(%) 0.7564</p> <p>Operator's Signature <i>DERR</i></p>	<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/26/2022 SN 80-001344 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:45 Control Test 0.197 07:46 Air Blank 0.000 07:47 Control Test 0.196 07:47 Air Blank 0.000 07:48 Control Test 0.196 07:49 Air Blank 0.000 07:49 Control Test Stats Average 0.1963 Std Dev 0.0006 Rel Std Dev(%) 0.2941</p> <p>Operator's Signature <i>DERR</i></p>	<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/26/2022 SN 80-001344 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 07:51 Control Test 0.078 07:51 Air Blank 0.000 07:52 Control Test 0.079 07:52 Air Blank 0.000 07:53 Control Test 0.079 07:53 Air Blank 0.000 07:53 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339</p> <p>Operator's Signature <i>DERR</i></p>

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO

Time of Inspection: 06:54

Date of Inspection: 07/26/2022

Serial Number: 80-001344

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____


Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DAVID E REYES-RIVERA

 Signature and Printed Name

07/26/2022
 Date