

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-002988, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-002988</u>	UNCERTAINTY* ±	
Owning Agency:	<u>MIAMI GARDENS PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/01/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:23</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/01/2022

Date


DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI GARDENS PD
Time of Inspection: 10:23

Date of Inspection: 06/01/2022

Serial Number: 80-002988
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.078	0.197	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.047	0.078	0.197	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.081
0.000	0.049	0.078	0.197	0.080


Standard Deviations	0.0006	0.0004	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

06/01/2022
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-002988	Miami Gardens Police Department	6/1/2022	DERR

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
<p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>MIAMI GARDENS PD Analytizer - Alcohol Analyzer Model: 8000 SN: 80-002988 05/01/2022 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:15</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:17</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:19</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:19</td></tr> <tr><td>Control Test</td><td>0.048</td><td>08:19</td></tr> <tr><td>Average</td><td>0.048</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Rel. Std Dev (%)</td><td>1.1945</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	08:15	Control Test	0.048	08:16	Air Blank	0.000	08:17	Control Test	0.049	08:18	Air Blank	0.000	08:19	Control Test	0.048	08:19	Air Blank	0.000	08:19	Control Test	0.048	08:19	Average	0.048		Std Dev	0.000		Rel. Std Dev (%)	1.1945		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI GARDENS PD Analytizer - Alcohol Analyzer Model: 8000 SN: 80-002988 06/01/2022 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:16</td></tr> <tr><td>Control Test</td><td>0.078</td><td>08:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:17</td></tr> <tr><td>Control Test</td><td>0.077</td><td>08:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:18</td></tr> <tr><td>Control Test</td><td>0.077</td><td>08:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:19</td></tr> <tr><td>Control Test</td><td>0.077</td><td>08:19</td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0005</td><td></td></tr> <tr><td>Rel. Std Dev (%)</td><td>0.7466</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	08:16	Control Test	0.078	08:16	Air Blank	0.000	08:17	Control Test	0.077	08:18	Air Blank	0.000	08:18	Control Test	0.077	08:19	Air Blank	0.000	08:19	Control Test	0.077	08:19	Average	0.0773		Std Dev	0.0005		Rel. Std Dev (%)	0.7466		<p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>MIAMI GARDENS PD Analytizer - Alcohol Analyzer Model: 8000 SN: 80-002988 05/01/2022 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:20</td></tr> <tr><td>Control Test</td><td>0.196</td><td>08:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:21</td></tr> <tr><td>Control Test</td><td>0.197</td><td>08:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:23</td></tr> <tr><td>Control Test</td><td>0.197</td><td>08:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:24</td></tr> <tr><td>Control Test</td><td>0.197</td><td>08:24</td></tr> <tr><td>Average</td><td>0.1967</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev (%)</td><td>0.2956</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	08:20	Control Test	0.196	08:21	Air Blank	0.000	08:21	Control Test	0.197	08:22	Air Blank	0.000	08:23	Control Test	0.197	08:23	Air Blank	0.000	08:24	Control Test	0.197	08:24	Average	0.1967		Std Dev	0.0006		Rel. Std Dev (%)	0.2956		<p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>MIAMI GARDENS PD Analytizer - Alcohol Analyzer Model: 8000 SN: 80-002988 05/01/2022 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>08:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:26</td></tr> <tr><td>Control Test</td><td>0.078</td><td>08:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:27</td></tr> <tr><td>Control Test</td><td>0.078</td><td>08:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:28</td></tr> <tr><td>Control Test</td><td>0.078</td><td>08:28</td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev (%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>[Signature]</i></p>	Test	g/210L	Time	Air Blank	0.000	08:25	Control Test	0.079	08:26	Air Blank	0.000	08:26	Control Test	0.078	08:26	Air Blank	0.000	08:27	Control Test	0.078	08:27	Air Blank	0.000	08:28	Control Test	0.078	08:28	Average	0.0783		Std Dev	0.0006		Rel. Std Dev (%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:15																																																																																																																																																	
Control Test	0.048	08:16																																																																																																																																																	
Air Blank	0.000	08:17																																																																																																																																																	
Control Test	0.049	08:18																																																																																																																																																	
Air Blank	0.000	08:19																																																																																																																																																	
Control Test	0.048	08:19																																																																																																																																																	
Air Blank	0.000	08:19																																																																																																																																																	
Control Test	0.048	08:19																																																																																																																																																	
Average	0.048																																																																																																																																																		
Std Dev	0.000																																																																																																																																																		
Rel. Std Dev (%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:16																																																																																																																																																	
Control Test	0.078	08:16																																																																																																																																																	
Air Blank	0.000	08:17																																																																																																																																																	
Control Test	0.077	08:18																																																																																																																																																	
Air Blank	0.000	08:18																																																																																																																																																	
Control Test	0.077	08:19																																																																																																																																																	
Air Blank	0.000	08:19																																																																																																																																																	
Control Test	0.077	08:19																																																																																																																																																	
Average	0.0773																																																																																																																																																		
Std Dev	0.0005																																																																																																																																																		
Rel. Std Dev (%)	0.7466																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:20																																																																																																																																																	
Control Test	0.196	08:21																																																																																																																																																	
Air Blank	0.000	08:21																																																																																																																																																	
Control Test	0.197	08:22																																																																																																																																																	
Air Blank	0.000	08:23																																																																																																																																																	
Control Test	0.197	08:23																																																																																																																																																	
Air Blank	0.000	08:24																																																																																																																																																	
Control Test	0.197	08:24																																																																																																																																																	
Average	0.1967																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev (%)	0.2956																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:25																																																																																																																																																	
Control Test	0.079	08:26																																																																																																																																																	
Air Blank	0.000	08:26																																																																																																																																																	
Control Test	0.078	08:26																																																																																																																																																	
Air Blank	0.000	08:27																																																																																																																																																	
Control Test	0.078	08:27																																																																																																																																																	
Air Blank	0.000	08:28																																																																																																																																																	
Control Test	0.078	08:28																																																																																																																																																	
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev (%)	0.7370																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI GARDENS PD
Time of Inspection: 07:56

Date of Inspection: 06/01/2022

Serial Number: 80-002988
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


DAVID E REYES-RIVERA

 Signature and Printed Name

06/01/2022
 Date