



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006471, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006471</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MONROE COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/09/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:57</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/09/2022

Date

David Reyes-Rivera
DAVID E REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY S.O.
Time of Inspection: 09:57

Date of Inspection: 03/09/2022

Serial Number: 80-006471
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.047	0.078	0.197	0.078
0.000	0.047	0.077	0.198	0.077
0.000	0.047	0.078	0.197	0.077
0.000	0.047	0.077	0.198	0.077
0.000	0.047	0.078	0.198	0.078
0.000	0.047	0.078	0.198	0.077
0.000	0.047	0.078	0.198	0.077
0.000	0.047	0.078	0.198	0.077
0.000	0.048	0.078	0.198	0.077
0.000	0.047	0.078	0.198	0.076


Standard Deviations	0.0003	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

DAVID E REYES-RIVERA

03/09/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006471	Monroe County Sheriff's Office	3/9/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 03/09/2022 Software: 8100.27 SN 80-006471 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>07:58</td></tr><tr><td>Control Test</td><td>0.048</td><td>07:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:59</td></tr><tr><td>Control Test</td><td>0.048</td><td>08:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:01</td></tr><tr><td>Control Test</td><td>0.048</td><td>08:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0480</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	07:58	Control Test	0.048	07:59	Air Blank	0.000	07:59	Control Test	0.048	08:00	Air Blank	0.000	08:01	Control Test	0.048	08:01	Air Blank	0.000	08:02	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 03/09/2022 Software: 8100.27 SN 80-006471 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:03</td></tr><tr><td>Control Test</td><td>0.078</td><td>08:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:04</td></tr><tr><td>Control Test</td><td>0.077</td><td>08:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:05</td></tr><tr><td>Control Test</td><td>0.078</td><td>08:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:07</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0777</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	08:03	Control Test	0.078	08:04	Air Blank	0.000	08:04	Control Test	0.077	08:05	Air Blank	0.000	08:05	Control Test	0.078	08:06	Air Blank	0.000	08:07	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 03/09/2022 Software: 8100.27 SN 80-006471 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:08</td></tr><tr><td>Control Test</td><td>0.197</td><td>08:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:09</td></tr><tr><td>Control Test</td><td>0.197</td><td>08:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:10</td></tr><tr><td>Control Test</td><td>0.197</td><td>08:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:11</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1970</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	08:08	Control Test	0.197	08:08	Air Blank	0.000	08:09	Control Test	0.197	08:10	Air Blank	0.000	08:10	Control Test	0.197	08:11	Air Blank	0.000	08:11	Control Test Stats			Average	0.1970		Std Dev	0.0000		Rel Std Dev(%)	0.0000		MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 03/09/2022 Software: 8100.27 SN 80-006471 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>08:16</td></tr><tr><td>Control Test</td><td>0.077</td><td>08:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:17</td></tr><tr><td>Control Test</td><td>0.077</td><td>08:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:18</td></tr><tr><td>Control Test</td><td>0.077</td><td>08:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>08:19</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0770</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	08:16	Control Test	0.077	08:17	Air Blank	0.000	08:17	Control Test	0.077	08:18	Air Blank	0.000	08:18	Control Test	0.077	08:18	Air Blank	0.000	08:19	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:58																																																																																																																																																	
Control Test	0.048	07:59																																																																																																																																																	
Air Blank	0.000	07:59																																																																																																																																																	
Control Test	0.048	08:00																																																																																																																																																	
Air Blank	0.000	08:01																																																																																																																																																	
Control Test	0.048	08:01																																																																																																																																																	
Air Blank	0.000	08:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:03																																																																																																																																																	
Control Test	0.078	08:04																																																																																																																																																	
Air Blank	0.000	08:04																																																																																																																																																	
Control Test	0.077	08:05																																																																																																																																																	
Air Blank	0.000	08:05																																																																																																																																																	
Control Test	0.078	08:06																																																																																																																																																	
Air Blank	0.000	08:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:08																																																																																																																																																	
Control Test	0.197	08:08																																																																																																																																																	
Air Blank	0.000	08:09																																																																																																																																																	
Control Test	0.197	08:10																																																																																																																																																	
Air Blank	0.000	08:10																																																																																																																																																	
Control Test	0.197	08:11																																																																																																																																																	
Air Blank	0.000	08:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1970																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:16																																																																																																																																																	
Control Test	0.077	08:17																																																																																																																																																	
Air Blank	0.000	08:17																																																																																																																																																	
Control Test	0.077	08:18																																																																																																																																																	
Air Blank	0.000	08:18																																																																																																																																																	
Control Test	0.077	08:18																																																																																																																																																	
Air Blank	0.000	08:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
 Operator's Signature	 Operator's Signature	 Operator's Signature	 Operator's Signature																																																																																																																																																