

Return Material Authorization

Ship to: ☒ CMI, Inc.
☐ Enforcement Electronics

Shipment to repair facility authorized by: Anthony Dobosiewicz on 2/24/2022

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-006638

Bill To Address:

Florida Highway Patrol Troop C

Attn: Anthony Dobosiewicz

11305 N. McKinley Drive

Tampa, Florida 33612

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

During startup, instrument gives EEPROM Fail and will not come out of Standby Mode.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Anthony Dobosiewicz

Phone #: 813-558-1800 Email: AnthonyDobosiewicz@flhsmv.gov

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us



INSTRUMENT PROCESSING SHEET

Agency FHP Troop C

S/N 80-006638

Florida Department of
Law Enforcement

Date In 2/24/2022

DI Completion Date _____

☒ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By TDG	Quality Checks	By	Date	Flow Calibration	By	Date															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☐ Breath Tube Screen ☐ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value _____ ☐ Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) ☐ Barometric Pressure Check Gauge ID # _____ ☐ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																	
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____		
Simulator	Serial #	Lot #/Exp																				
0.050																						
0.080																						
0.200																						
0.080 DGS	N/A																					

Calibration Adjustment	By	Department Inspection	By																																																													
Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: <u>EEPROM Fail during startup.</u> <u>Will not come out of Standby Mode. Will send to CMI</u> <u>for repair (TDG)</u> _____ _____ _____ _____ _____	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.050</td> <td></td> </tr> <tr> <td>0.080</td> <td></td> </tr> <tr> <td>0.200</td> <td></td> </tr> </tbody> </table> Attachments ☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 51</u> ☐ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use Tech Review / Date _____ Admin Review / Date _____	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200			
Simulator	Serial #	Lot #	Expiration																																																													
0.000		N/A	N/A																																																													
0.040																																																																
0.100																																																																
0.200																																																																
0.300																																																																
0.080 DGS	N/A																																																															
Simulator	Serial #	Lot #	Expiration																																																													
0.050																																																																
0.080																																																																
0.200																																																																
0.080 DGS	N/A																																																															
Simulator	Serial Number																																																															
0.000																																																																
Interferent																																																																
0.050																																																																
0.080																																																																
0.200																																																																



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Troop CS/N 80-006638Florida Department of
Law EnforcementDate In 7/11/2022DI Completion Date 7/12/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>DERR</u>	Quality Checks	By <u>DERR</u>	Date <u>7/12/2022</u>	Flow Calibration	By	Date														
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>204</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP4863</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP4864</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5097</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP4863	202201C 01/11/2024	0.080	MP4864	202201D 01/18/2024	0.200	MP5097	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023				
Simulator	Serial #	Lot #/Exp																			
0.050	MP4863	202201C 01/11/2024																			
0.080	MP4864	202201D 01/18/2024																			
0.200	MP5097	202201E 01/18/2024																			
0.080 DGS	N/A	AG115904 06/08/2023																			

Maintenance	By
☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____	_____ _____ _____ _____ _____

Calibration Adjustment	By	Department Inspection	By <u>DERR</u>																																																												
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Barometric Pressure ID# <u>26932</u> Gauge <u>1014</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>SD1014</td></tr><tr><td>Interferent</td><td>SD1017</td></tr><tr><td>0.050</td><td>MP4863</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP5097</td></tr></tbody></table>	Simulator	Serial Number	0.000	SD1014	Interferent	SD1017	0.050	MP4863	0.080	MP4864	0.200	MP5097	
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	SD1014																																																														
Interferent	SD1017																																																														
0.050	MP4863																																																														
0.080	MP4864																																																														
0.200	MP5097																																																														
		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																																													

Notes/Suggested Service: _____

_____☒ Instrument Complies with Chapter 11D-8, FAC
☐ Instrument Does Not Comply with Chapter 11D-8, FAC
☒ Return to/Place into Evidentiary Use
☐ Remain Out of Evidentiary Use
☒ Conduct an Agency Inspection Before Evidentiary UseIsrael Soto
Digitally signed by Israel Soto
Date: 2022.07.12 11:21:48
+0400


Tech Review / Date

Admin Review / Date

2022.07.12

14:57:53

0400



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006638, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006638</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP C</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/12/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:28</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

07/12/2022

Date

DAVID E REYES-RIVERA,
Department Inspector

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP C
Time of Inspection: 10:28

Date of Inspection: 07/12/2022

Serial Number: 80-006638
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.049	0.079	0.198	0.000 / 0.080
0.000	0.049	0.080	0.199	0.000 / 0.080
0.000	0.050	0.079	0.199	0.000 / 0.080
0.000	0.050	0.079	0.199	0.000 / 0.080
0.000	0.050	0.080	0.199	0.000 / 0.080
0.000	0.050	0.079	0.199	0.000 / 0.080
0.000	0.050	0.080	0.199	0.000 / 0.080
0.000	0.050	0.079	0.199	0.000 / 0.080
0.000	0.050	0.079	0.199	0.000 / 0.080
0.000	0.050	0.080	0.200	0.000 / 0.080

Standard Deviations	0.0004	0.0005	0.0004	0.0000 / 0.0000
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance REPLACED EMPTY DGS.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

07/12/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006638	Florida Highway Patrol Troop C	7/12/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
FHP TROOP C Intoxilyzer - Alcohol Analyzer Model 8000 07/12/2022 Software: 8100.27 SN 80-006638 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>07:34</td></tr><tr><td>Control Test</td><td>0.049</td><td>07:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:36</td></tr><tr><td>Control Test</td><td>0.049</td><td>07:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:37</td></tr><tr><td>Control Test</td><td>0.048</td><td>07:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:38</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0487</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	07:34	Control Test	0.049	07:35	Air Blank	0.000	07:36	Control Test	0.049	07:36	Air Blank	0.000	07:37	Control Test	0.048	07:37	Air Blank	0.000	07:38	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		FHP TROOP C Intoxilyzer - Alcohol Analyzer Model 8000 07/12/2022 Software: 8100.27 SN 80-006638 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>07:39</td></tr><tr><td>Control Test</td><td>0.078</td><td>07:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:41</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:42</td></tr><tr><td>Control Test</td><td>0.078</td><td>07:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:43</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0016</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	07:39	Control Test	0.078	07:40	Air Blank	0.000	07:41	Control Test	0.079	07:41	Air Blank	0.000	07:42	Control Test	0.078	07:42	Air Blank	0.000	07:43	Control Test Stats			Average	0.0783		Std Dev	0.0016		Rel Std Dev(%)	0.7370		FHP TROOP C Intoxilyzer - Alcohol Analyzer Model 8000 07/12/2022 Software: 8100.27 SN 80-006638 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.001</td><td>07:44</td></tr><tr><td>Control Test</td><td>0.198</td><td>07:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:45</td></tr><tr><td>Control Test</td><td>0.198</td><td>07:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:46</td></tr><tr><td>Control Test</td><td>0.198</td><td>07:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:48</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1980</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.001	07:44	Control Test	0.198	07:45	Air Blank	0.000	07:45	Control Test	0.198	07:46	Air Blank	0.000	07:46	Control Test	0.198	07:47	Air Blank	0.000	07:48	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	1.0000		FHP TROOP C Intoxilyzer - Alcohol Analyzer Model 8000 07/12/2022 Software: 8100.27 SN 80-006638 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>07:49</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:49</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:50</td></tr><tr><td>Control Test</td><td>0.079</td><td>07:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:51</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	07:49	Control Test	0.079	07:49	Air Blank	0.000	07:49	Control Test	0.079	07:50	Air Blank	0.000	07:50	Control Test	0.079	07:51	Air Blank	0.000	07:51	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:34																																																																																																																																																	
Control Test	0.049	07:35																																																																																																																																																	
Air Blank	0.000	07:36																																																																																																																																																	
Control Test	0.049	07:36																																																																																																																																																	
Air Blank	0.000	07:37																																																																																																																																																	
Control Test	0.048	07:37																																																																																																																																																	
Air Blank	0.000	07:38																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:39																																																																																																																																																	
Control Test	0.078	07:40																																																																																																																																																	
Air Blank	0.000	07:41																																																																																																																																																	
Control Test	0.079	07:41																																																																																																																																																	
Air Blank	0.000	07:42																																																																																																																																																	
Control Test	0.078	07:42																																																																																																																																																	
Air Blank	0.000	07:43																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.001	07:44																																																																																																																																																	
Control Test	0.198	07:45																																																																																																																																																	
Air Blank	0.000	07:45																																																																																																																																																	
Control Test	0.198	07:46																																																																																																																																																	
Air Blank	0.000	07:46																																																																																																																																																	
Control Test	0.198	07:47																																																																																																																																																	
Air Blank	0.000	07:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	1.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:49																																																																																																																																																	
Control Test	0.079	07:49																																																																																																																																																	
Air Blank	0.000	07:49																																																																																																																																																	
Control Test	0.079	07:50																																																																																																																																																	
Air Blank	0.000	07:50																																																																																																																																																	
Control Test	0.079	07:51																																																																																																																																																	
Air Blank	0.000	07:51																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																																



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Troop C

S/N 80-006638

Florida Department of
Law Enforcement

Date In 08/30/2022 DI Completion Date 09/01/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DERR	Date <u>09/01/2022</u>	Flow Calibration	By	Date															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>209</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																	
				<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>SD3963</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>SD1017</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>SD3968</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2 02/05/2023</td> </tr> </tbody> </table>				Simulator	Serial #	Lot #/Exp	0.050	SD3963	202201C 01/11/2024	0.080	SD1017	202201D 01/18/2024	0.200	SD3968	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023
Simulator	Serial #	Lot #/Exp																				
0.050	SD3963	202201C 01/11/2024																				
0.080	SD1017	202201D 01/18/2024																				
0.200	SD3968	202201E 01/18/2024																				
0.080 DGS	N/A	00521080A2 02/05/2023																				
				<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Maintenance</th> <th>By</th> </tr> </thead> <tbody> <tr> <td colspan="2"> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ </td> <td>_____</td> </tr> </tbody> </table>				Maintenance		By	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____		_____									
Maintenance		By																				
☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____		_____																				

Calibration Adjustment	Department Inspection																																																																
DEER <u>9/1/22</u> DEER <u>9/1/22</u> By DERR _____ Barometric Pressure Gauge <u>4046</u> ID # <u>26932</u> DEER <u>9/1/22</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: <u>Instrument was sent back due to the gas regulator had separated during shipment, it was rescrewed and instrument tested to see if there was anything else wrong. DERR</u> <u>Technical Review identified that I had made some notations in the calibration adjustment section DEER 9/1/22</u>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>26932</u> Gauge <u>1016</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD1014</td> </tr> <tr> <td>Interferent</td> <td>MP5097</td> </tr> <tr> <td>0.050</td> <td>SD3963</td> </tr> <tr> <td>0.080</td> <td>SD1017</td> </tr> <tr> <td>0.200</td> <td>SD3968</td> </tr> </tbody> </table> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment </td> <td> ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____ </td> </tr> </tbody> </table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Phil Nicodemo Quality signed by Phil Nicodemo Date: 2022/09/12 18:46:13 -0400 13 14:27:03 -0400 Tech Review / Date _____ Admin Review / Date _____	Simulator	Serial Number	0.000	SD1014	Interferent	MP5097	0.050	SD3963	0.080	SD1017	0.200	SD3968	Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	SD1014																																																																
Interferent	MP5097																																																																
0.050	SD3963																																																																
0.080	SD1017																																																																
0.200	SD3968																																																																
Attachments																																																																	
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																																																																



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006638, manufactured by CML, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006638</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FHP TROOP C</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/01/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:04</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration. *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full,

without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/01/2022

Date

David Reyes-Rivera
DAVID E REYES-RIVERA,
Department Inspector

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP C

Time of Inspection: 13:04

Date of Inspection: 09/01/2022

Serial Number: 80-006638

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.079	0.198	0.081
0.000	0.050	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.081
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.080	0.199	0.080

Standard Deviations	0.0005	0.0005	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

09/01/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006638	Florida Highway Patrol Troop C	9/1/2022	DERR <i>pk</i>

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	D65 0.08g/210L 0.077 to 0.083
FHP TROOP C Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006638 09/01/2022 Software: 8100.27 Test g/210L Time Air Blank 0.010 08:38 Control Test 0.050 08:39 Air Blank 0.000 08:40 Control Test 0.050 08:40 Air Blank 0.000 08:41 Control Test 0.050 08:42 Air Blank 0.000 08:42 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>pk</i>	FHP TROOP C Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006638 09/01/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:44 Control Test 0.079 08:45 Air Blank 0.000 08:45 Control Test 0.080 08:46 Air Blank 0.000 08:46 Control Test 0.080 08:47 Air Blank 0.000 08:47 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247 Operator's Signature <i>pk</i>	FHP TROOP C Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006638 09/01/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:49 Control Test 0.199 08:50 Air Blank 0.000 08:51 Control Test 0.198 08:51 Air Blank 0.000 08:52 Control Test 0.198 08:53 Air Blank 0.000 08:53 Control Test Stats Average 0.1983 Std Dev 0.0006 Rel Std Dev(%) 0.2911 Operator's Signature <i>pk</i>	FHP TROOP C Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006638 09/01/2022 Software: 8100.27 Test g/210L Time Air Blank 0.000 08:55 Control Test 0.090 08:56 Air Blank 0.000 08:56 Control Test 0.091 08:56 Air Blank 0.000 08:57 Control Test 0.090 08:57 Air Blank 0.000 08:58 Control Test Stats Average 0.0903 Std Dev 0.0006 Rel Std Dev(%) 0.7187 Operator's Signature <i>pk</i>

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP C

Time of Inspection: 07:09

Date of Inspection: 09/01/2022

Serial Number: 80-006638

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____


Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

09/01/2022
Date