

**Return Material Authorization**

**Ship to:**  CMI, Inc.  
 Enforcement Electronics

Shipment to repair facility authorized by: Captain Strickland \_\_\_\_\_ on 1/10/2022

**Items Returned:** Instrument  Supplies  Other  Describe: \_\_\_\_\_  
Instrument Model: Intoxilyzerv 8000 Serial Number: 80-006763

Bill to Address: <u>Florida Highway Patrol Troop L</u> <u>ATTN: Captain Strickland</u> Lake Worth Svc. Plz. MM94 Law Enforcement Bldg. #9330 Lake Worth, FL 33467 _____	Ship to address: <u>FDLE FMROC</u> <u>4700 Terminal Drive Suite 1</u> <u>Fort Myers, Florida 33991</u>
---	--

**Reason for Return:** I received 80-006763 (Florida Highway Patrol Troop L). The instrument is not working properly. During the stabilities testing the instrument had test exemptions (interferent detect, ambient fail) on all of the values. DVM also shows a great spike on the 3 filter, as high as S: 54.

**Please choose one of the following options:**

- 1. I \_\_\_\_\_, authorize all repairs.
- 2. I \_\_\_\_\_, authorize repairs up to \$\_\_\_\_\_.
- 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Captain Strickland  
Phone #: 561-357-4040 Email: kevinstrickland@flhsmv.gov

ATP Contact Name: David Reyes ATP Email: davidreyes@fdle.state.fl.us