



INSTRUMENT PROCESSING SHEET

Agency Charlotte County Sheriff's OfficeS/N 80-007083Florida Department of
Law EnforcementDate In 4/6/2022DI Completion Date 4/20/2022☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By DERR	Quality Checks	By DERR	Date <u>4/19/2022</u>	Flow Calibration	By	Date																																						
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Instrument was received with a note indication RAM fail. Instrument was turned on and performed quality checks without any fails. DERR</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>250</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28199</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																								
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP6287</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904 06/08/2023</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	AG115904 06/08/2023																												
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6286	202201C 01/11/2024																																											
0.080	MP6287	202201D 01/18/2024																																											
0.200	MP6288	202201E 01/18/2024																																											
0.080 DGS	N/A	AG115904 06/08/2023																																											
Calibration Adjustment		By _____		Department Inspection By DERR _____																																									
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>68639</u> Gauge <u>1020</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>																																									
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6284																																												
Interferent	MP6285																																												
0.050	MP6286																																												
0.080	MP6287																																												
0.200	MP6288																																												
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				Attachments																																									
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																					
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: <u>Instrument performed very well during the quality checks. Moved on to the DI. DERR</u>				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																									
				2022.04. 22 13:40:22 Israel Soto Tech Review / Date Admin Review / Date																																									



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007083, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007083</u>	UNCERTAINTY* \pm
Owning Agency:	<u>CHARLOTTE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/20/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>09:21</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/20/2022

Date

DAVID E'REYES-RIVERA,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CHARLOTTE COUNTY SO
Time of Inspection: 09:21

Date of Inspection: 04/20/2022

Serial Number: 80-007083
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.051	0.080	0.200	0.080
0.000	0.050	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.080	0.200	0.080

Standard Deviations	0.0003	0.0000	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

David E Reyes-Rivera DAVID E REYES-RIVERA
Signature and Printed Name

04/20/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007083	Charlotte County Sheriff's Office	4/19/2022	DERR <i>[Signature]</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<div>CHARLOTTE COUNTY SO Intoxilzyer - Alcohol Analyzer Model 8000 04/19/2022 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:29</td></tr><tr><td>Control Test</td><td>0.050</td><td>13:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr><tr><td>Control Test</td><td>0.050</td><td>13:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr><tr><td>Control Test</td><td>0.050</td><td>13:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0500</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <div><i>[Signature]</i> Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	13:29	Control Test	0.050	13:30	Air Blank	0.000	13:30	Control Test	0.050	13:31	Air Blank	0.000	13:31	Control Test	0.050	13:32	Air Blank	0.000	13:33	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<div>CHARLOTTE COUNTY SO Intoxilzyer - Alcohol Analyzer Model 8000 04/19/2022 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:34</td></tr><tr><td>Control Test</td><td>0.079</td><td>13:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:35</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr><tr><td>Control Test</td><td>0.079</td><td>13:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:37</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0016</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></tbody></table> <div><i>[Signature]</i> Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	13:34	Control Test	0.079	13:34	Air Blank	0.000	13:35	Control Test	0.078	13:35	Air Blank	0.000	13:36	Control Test	0.079	13:37	Air Blank	0.000	13:37	Control Test Stats			Average	0.0787		Std Dev	0.0016		Rel Std Dev(%)	0.7339		<div>CHARLOTTE COUNTY SO Intoxilzyer - Alcohol Analyzer Model 8000 04/19/2022 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr><tr><td>Control Test</td><td>0.198</td><td>13:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr><tr><td>Control Test</td><td>0.199</td><td>13:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr><tr><td>Control Test</td><td>0.199</td><td>13:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1987</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr></tbody></table> <div><i>[Signature]</i> Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	13:38	Control Test	0.198	13:39	Air Blank	0.000	13:40	Control Test	0.199	13:40	Air Blank	0.000	13:41	Control Test	0.199	13:42	Air Blank	0.000	13:42	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		<div>CHARLOTTE COUNTY SO Intoxilzyer - Alcohol Analyzer Model 8000 04/19/2022 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:44</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:45</td></tr><tr><td>Control Test</td><td>0.079</td><td>13:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:46</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></tbody></table> <div><i>[Signature]</i> Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	13:43	Control Test	0.080	13:43	Air Blank	0.000	13:44	Control Test	0.080	13:44	Air Blank	0.000	13:45	Control Test	0.079	13:45	Air Blank	0.000	13:46	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:29																																																																																																																																																	
Control Test	0.050	13:30																																																																																																																																																	
Air Blank	0.000	13:30																																																																																																																																																	
Control Test	0.050	13:31																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test	0.050	13:32																																																																																																																																																	
Air Blank	0.000	13:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:34																																																																																																																																																	
Control Test	0.079	13:34																																																																																																																																																	
Air Blank	0.000	13:35																																																																																																																																																	
Control Test	0.078	13:35																																																																																																																																																	
Air Blank	0.000	13:36																																																																																																																																																	
Control Test	0.079	13:37																																																																																																																																																	
Air Blank	0.000	13:37																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:38																																																																																																																																																	
Control Test	0.198	13:39																																																																																																																																																	
Air Blank	0.000	13:40																																																																																																																																																	
Control Test	0.199	13:40																																																																																																																																																	
Air Blank	0.000	13:41																																																																																																																																																	
Control Test	0.199	13:42																																																																																																																																																	
Air Blank	0.000	13:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2906																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:43																																																																																																																																																	
Control Test	0.080	13:43																																																																																																																																																	
Air Blank	0.000	13:44																																																																																																																																																	
Control Test	0.080	13:44																																																																																																																																																	
Air Blank	0.000	13:45																																																																																																																																																	
Control Test	0.079	13:45																																																																																																																																																	
Air Blank	0.000	13:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		



Agency Charlotte County Sheriff's Office

S/N 80-007083

Date In 11/17/2022 DI Completion Date 11/21/2022 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

30 DEER 11/30/22

FDLE/ATP Form 48 January 2022
Issuing Authority: Alcohol Testing Program

Page 1 of 1



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007083, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007083</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>CHARLOTTE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/30/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:29</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/30/2022

Date


DAVID E REYES-RIVERA,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CHARLOTTE COUNTY SO
Time of Inspection: 08:29

Date of Inspection: 11/30/2022

Serial Number: 80-007083
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.050	0.078	0.197	0.081
0.000	0.050	0.078	0.199	0.081
0.000	0.050	0.079	0.198	0.080
0.000	0.051	0.079	0.198	0.081
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.079	0.199	0.080


Standard Deviations	0.0003	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DAVID E REYES-RIVERA

Signature and Printed Name

11/30/2022
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007083	Charlotte County Sheriff's Office	11/30/2022	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007083 11/30/2022 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>06:46</td></tr><tr><td>Control Test</td><td>0.049</td><td>06:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:47</td></tr><tr><td>Control Test</td><td>0.049</td><td>06:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr><tr><td>Control Test</td><td>0.050</td><td>06:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:49</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr></tbody></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:46	Control Test	0.049	06:46	Air Blank	0.000	06:47	Control Test	0.049	06:48	Air Blank	0.000	06:48	Control Test	0.050	06:49	Air Blank	0.000	06:49	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007083 11/30/2022 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>06:50</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:53</td></tr><tr><td>Control Test</td><td>0.079</td><td>06:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:54</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:50	Control Test	0.079	06:51	Air Blank	0.000	06:52	Control Test	0.079	06:52	Air Blank	0.000	06:53	Control Test	0.079	06:54	Air Blank	0.000	06:54	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007083 11/30/2022 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>06:55</td></tr><tr><td>Control Test</td><td>0.198</td><td>06:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:57</td></tr><tr><td>Control Test</td><td>0.198</td><td>06:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:58</td></tr><tr><td>Control Test</td><td>0.198</td><td>06:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>06:59</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.1980</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:55	Control Test	0.198	06:56	Air Blank	0.000	06:57	Control Test	0.198	06:57	Air Blank	0.000	06:58	Control Test	0.198	06:59	Air Blank	0.000	06:59	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007083 11/30/2022 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr><tr><td>Control Test</td><td>0.081</td><td>07:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:01</td></tr><tr><td>Control Test</td><td>0.081</td><td>07:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:02</td></tr><tr><td>Control Test</td><td>0.080</td><td>07:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>07:03</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0807</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr></tbody></table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	07:00	Control Test	0.081	07:00	Air Blank	0.000	07:01	Control Test	0.081	07:01	Air Blank	0.000	07:02	Control Test	0.080	07:02	Air Blank	0.000	07:03	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:46																																																																																																																																																	
Control Test	0.049	06:46																																																																																																																																																	
Air Blank	0.000	06:47																																																																																																																																																	
Control Test	0.049	06:48																																																																																																																																																	
Air Blank	0.000	06:48																																																																																																																																																	
Control Test	0.050	06:49																																																																																																																																																	
Air Blank	0.000	06:49																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:50																																																																																																																																																	
Control Test	0.079	06:51																																																																																																																																																	
Air Blank	0.000	06:52																																																																																																																																																	
Control Test	0.079	06:52																																																																																																																																																	
Air Blank	0.000	06:53																																																																																																																																																	
Control Test	0.079	06:54																																																																																																																																																	
Air Blank	0.000	06:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:55																																																																																																																																																	
Control Test	0.198	06:56																																																																																																																																																	
Air Blank	0.000	06:57																																																																																																																																																	
Control Test	0.198	06:57																																																																																																																																																	
Air Blank	0.000	06:58																																																																																																																																																	
Control Test	0.198	06:59																																																																																																																																																	
Air Blank	0.000	06:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test	0.081	07:00																																																																																																																																																	
Air Blank	0.000	07:01																																																																																																																																																	
Control Test	0.081	07:01																																																																																																																																																	
Air Blank	0.000	07:02																																																																																																																																																	
Control Test	0.080	07:02																																																																																																																																																	
Air Blank	0.000	07:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		