



# INSTRUMENT PROCESSING SHEET

Agency FWCCS/N 80-007166

Florida Department of Law Enforcement

Date In 11/16/2022DI Completion Date 11/17/2022 Ship  P/U  H/D  CMI  EE~~DEC 11/17/2022~~  
11/17/2022 ~~DEC 11/17/2022~~

Intake	By DERR	Quality Checks	By DERR	Date	Flow Calibration	By	Date															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>268</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28199</u> <input checked="" type="checkbox"/> Stability Checks		<u>11/17/2022</u>	Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																	
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2 02/05/2023</td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP6287	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023	<b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____ _____		
Simulator	Serial #	Lot #/Exp																				
0.050	MP6286	202201C 01/11/2024																				
0.080	MP6287	202201D 01/18/2024																				
0.200	MP6288	202201E 01/18/2024																				
0.080 DGS	N/A	00521080A2 02/05/2023																				

Calibration Adjustment	By	Department Inspection	By DERR																																																												
Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Barometric Pressure ID# <u>26932</u> Gauge <u>1020</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP6284																																																														
Interferent	MP6285																																																														
0.050	MP6286																																																														
0.080	MP6287																																																														
0.200	MP6288																																																														
		<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																													

Notes/Suggested Service: <u>Tech review identified that I had the wrong date for the quality checks. DEC 11/17/2022</u>  <u>Mistakenly corrected the wrong date DE inspection date was correct. Quality check date was corrected DEC 11/17/22</u>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2022.11.17 12:43:56 -05'00'</small>	Israel Soto <small>Digitally signed by Israel Soto Date: 2022.11.17 14:26:42 -05'00'</small>
Tech Review / Date	Admin Review / Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007166 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007166</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/17/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:28</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/17/2022

Date

  
DAVID E REYES-RIVERA,  
Department Inspector

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FWCC

Time of Inspection: 08:28

Date of Inspection: 11/17/2022

Serial Number: 80-007166

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080


Standard Deviations	0.0004	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 \_\_\_\_\_  
 Signature and Printed Name

DAVID E REYES-RIVERA

11/17/2022  
 Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007166	FFWCC	11/16/2022	DERR <i>[Signature]</i>

0.05g/210L <input checked="" type="checkbox"/>	0.08g/210L <input checked="" type="checkbox"/>	0.20g/210L <input checked="" type="checkbox"/>	DGS 0.08g/210L <input checked="" type="checkbox"/>																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<p>FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007166 11/16/2022 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:27</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:28</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	13:27	Control Test	0.048	13:28	Air Blank	0.000	13:28	Control Test	0.048	13:29	Air Blank	0.000	13:30	Control Test	0.048	13:30	Air Blank	0.000	13:31	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007166 11/16/2022 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:34</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	13:32	Control Test	0.078	13:33	Air Blank	0.000	13:33	Control Test	0.078	13:34	Air Blank	0.000	13:34	Control Test	0.078	13:35	Air Blank	0.000	13:36	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007166 11/16/2022 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:37</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	13:37	Control Test	0.198	13:37	Air Blank	0.000	13:38	Control Test	0.198	13:39	Air Blank	0.000	13:39	Control Test	0.198	13:40	Air Blank	0.000	13:40	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007166 11/16/2022 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:44</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	13:42	Control Test	0.080	13:42	Air Blank	0.000	13:42	Control Test	0.079	13:43	Air Blank	0.000	13:43	Control Test	0.079	13:43	Air Blank	0.000	13:44	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:27																																																																																																																																																	
Control Test	0.048	13:28																																																																																																																																																	
Air Blank	0.000	13:28																																																																																																																																																	
Control Test	0.048	13:29																																																																																																																																																	
Air Blank	0.000	13:30																																																																																																																																																	
Control Test	0.048	13:30																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:32																																																																																																																																																	
Control Test	0.078	13:33																																																																																																																																																	
Air Blank	0.000	13:33																																																																																																																																																	
Control Test	0.078	13:34																																																																																																																																																	
Air Blank	0.000	13:34																																																																																																																																																	
Control Test	0.078	13:35																																																																																																																																																	
Air Blank	0.000	13:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:37																																																																																																																																																	
Control Test	0.198	13:37																																																																																																																																																	
Air Blank	0.000	13:38																																																																																																																																																	
Control Test	0.198	13:39																																																																																																																																																	
Air Blank	0.000	13:39																																																																																																																																																	
Control Test	0.198	13:40																																																																																																																																																	
Air Blank	0.000	13:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:42																																																																																																																																																	
Control Test	0.080	13:42																																																																																																																																																	
Air Blank	0.000	13:42																																																																																																																																																	
Control Test	0.079	13:43																																																																																																																																																	
Air Blank	0.000	13:43																																																																																																																																																	
Control Test	0.079	13:43																																																																																																																																																	
Air Blank	0.000	13:44																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		