



## INSTRUMENT PROCESSING SHEET

Agency Hillsborough CSO

S/N 80-007484

Florida Department of  
Law Enforcement

Date In 2/11/2022 DI Completion Date 2/21/2022

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

<b>Intake</b> By TDG _____ <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____	<b>Quality Checks</b> By TDG _____ Date 2/15/2022 <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value 225 <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # ATP104 32 mm 0.156 (.139 - .169) 36 mm 0.171 (.156 - .190) 53 mm 0.242 (.228 - .278) 103 mm 0.500 (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # 68639 <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202010A 10/05/2022</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202010B 10/05/2022</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202010D 10/06/2022</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG115904 06/08/2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202010A 10/05/2022	0.080	MP6287	202010B 10/05/2022	0.200	MP6288	202010D 10/06/2022	0.080 DGS	N/A	AG115904 06/08/2023	<b>Flow Calibration</b> By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____
Simulator	Serial #	Lot #/Exp															
0.050	MP6286	202010A 10/05/2022															
0.080	MP6287	202010B 10/05/2022															
0.200	MP6288	202010D 10/06/2022															
0.080 DGS	N/A	AG115904 06/08/2023															

<b>Calibration Adjustment</b> By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> By TDG _____ Barometric Pressure ID# 28199 Gauge 1025 Instrument 1024 Mouth Alcohol Solution Lot # 2021-B Acetone Stock Solution Lot # 2021-C <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>SD3963</td> </tr> <tr> <td>Interferent</td> <td>SD3966</td> </tr> <tr> <td>0.050</td> <td>MP5092</td> </tr> <tr> <td>0.080</td> <td>MP4864</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between;"> <div>             David Eliezer              Reyes-Rivera  <small>Digitally signed by David Eliezer Reyes-Rivera              Date: 2022.02.27 20:30:11 -05'00'</small> </div> <div style="text-align: right;">             01              10:22:00              05'00'           </div> </div> <div style="display: flex; justify-content: space-between;"> <div>Tech Review / Date</div> <div>Admin Review / Date</div> </div>	Simulator	Serial Number	0.000	SD3963	Interferent	SD3966	0.050	MP5092	0.080	MP4864	0.200	MP5094
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	SD3963																																																												
Interferent	SD3966																																																												
0.050	MP5092																																																												
0.080	MP4864																																																												
0.200	MP5094																																																												

Notes/Suggested Service: The first 0.05 Stability Test was unable to be printed due to a jam in the internal printer paper. Repeated the 0.05 Stability Test (TDG)  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 7484	Hillsborough CSO	02/15/2022	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
<p>HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 02/15/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:53</p> <p>Control Test 0.049 14:54</p> <p>Air Blank 0.000 14:54</p> <p>Control Test 0.049 14:55</p> <p>Air Blank 0.000 14:55</p> <p>Control Test 0.050 14:56</p> <p>Air Blank 0.000 14:57</p> <p>Control Test Stats</p> <p>Average 0.0493</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1703</p>	<p>HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 02/15/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:00</p> <p>Control Test 0.079 15:01</p> <p>Air Blank 0.000 15:02</p> <p>Control Test 0.080 15:02</p> <p>Air Blank 0.000 15:03</p> <p>Control Test 0.079 15:03</p> <p>Air Blank 0.000 15:04</p> <p>Control Test Stats</p> <p>Average 0.0793</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7277</p>	<p>HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 02/15/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:06</p> <p>Control Test 0.199 15:07</p> <p>Air Blank 0.000 15:08</p> <p>Control Test 0.198 15:08</p> <p>Air Blank 0.000 15:09</p> <p>Control Test 0.198 15:10</p> <p>Air Blank 0.000 15:10</p> <p>Control Test Stats</p> <p>Average 0.1983</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2911</p>	<p><i>065</i></p> <p>HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 02/15/2022 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:11</p> <p>Control Test 0.080 15:12</p> <p>Air Blank 0.000 15:12</p> <p>Control Test 0.080 15:12</p> <p>Air Blank 0.000 15:13</p> <p>Control Test 0.081 15:13</p> <p>Air Blank 0.000 15:14</p> <p>Control Test Stats</p> <p>Average 0.0803</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7187</p>
<p><i>ML</i></p> <p>Operator's Signature</p>	<p><i>ML</i></p> <p>Operator's Signature</p>	<p><i>ML</i></p> <p>Operator's Signature</p>	<p><i>ML</i></p> <p>Operator's Signature</p>

Comments:



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO  
Time of Inspection: 12:27

Date of Inspection: 02/21/2022

Serial Number: 80-007484  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202010A Exp: 10/05/2022	0.08g/210L Test (g/210L) Lot#:202010B Exp: 10/05/2022	0.20g/210L Test (g/210L) Lot#:202010D Exp: 10/06/2022	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.050	0.079	0.198	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.080	0.197	0.080
0.000	0.050	0.080	0.198	0.081
0.000	0.049	0.080	0.198	0.081

Standard Deviations	0.0004	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 TAYLOR D GUTSCHOW  
Signature and Printed Name

02/21/2022  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007484, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007484</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/21/2022</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:27</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/21/2022

Date

  
TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality







Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 7484	Hillsborough CSO	11/08/2022	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	≤0.003 of Wet																																																																																																																																																
<div>HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007484 11/08/2022 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr><tr><td>Control Test</td><td>0.047</td><td>12:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr><tr><td>Control Test</td><td>0.048</td><td>12:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr><tr><td>Control Test</td><td>0.047</td><td>12:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0473</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2198</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	12:49	Control Test	0.047	12:49	Air Blank	0.000	12:50	Control Test	0.048	12:51	Air Blank	0.000	12:51	Control Test	0.047	12:52	Air Blank	0.000	12:52	Control Test Stats			Average	0.0473		Std Dev	0.0006		Rel Std Dev(%)	1.2198		<div>HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007484 11/08/2022 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr><tr><td>Control Test</td><td>0.077</td><td>13:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr><tr><td>Control Test</td><td>0.077</td><td>13:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr><tr><td>Control Test</td><td>0.077</td><td>13:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0770</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	13:01	Control Test	0.077	13:02	Air Blank	0.000	13:02	Control Test	0.077	13:03	Air Blank	0.000	13:04	Control Test	0.077	13:04	Air Blank	0.000	13:05	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<div>HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007484 11/08/2022 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr><tr><td>Control Test</td><td>0.197</td><td>13:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr><tr><td>Control Test</td><td>0.197</td><td>13:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr><tr><td>Control Test</td><td>0.196</td><td>13:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1967</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2936</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	13:07	Control Test	0.197	13:08	Air Blank	0.000	13:08	Control Test	0.197	13:09	Air Blank	0.000	13:09	Control Test	0.196	13:10	Air Blank	0.000	13:11	Control Test Stats			Average	0.1967		Std Dev	0.0006		Rel Std Dev(%)	0.2936		<div>DGS #1</div> <div>HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007484 11/08/2022 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr><tr><td>Control Test</td><td>0.000</td><td>13:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr><tr><td>Control Test</td><td>0.000</td><td>13:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0260</td><td></td></tr><tr><td>Std Dev</td><td>0.0450</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>173.2051</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	13:11	Control Test	0.000	13:12	Air Blank	0.000	13:12	Control Test	0.000	13:13	Air Blank	0.000	13:13	Control Test	0.078	13:13	Air Blank	0.000	13:14	Control Test Stats			Average	0.0260		Std Dev	0.0450		Rel Std Dev(%)	173.2051	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.047	12:49																																																																																																																																																	
Air Blank	0.000	12:50																																																																																																																																																	
Control Test	0.048	12:51																																																																																																																																																	
Air Blank	0.000	12:51																																																																																																																																																	
Control Test	0.047	12:52																																																																																																																																																	
Air Blank	0.000	12:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0473																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2198																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test	0.077	13:02																																																																																																																																																	
Air Blank	0.000	13:02																																																																																																																																																	
Control Test	0.077	13:03																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.077	13:04																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:07																																																																																																																																																	
Control Test	0.197	13:08																																																																																																																																																	
Air Blank	0.000	13:08																																																																																																																																																	
Control Test	0.197	13:09																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test	0.196	13:10																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1967																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2936																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test	0.000	13:12																																																																																																																																																	
Air Blank	0.000	13:12																																																																																																																																																	
Control Test	0.000	13:13																																																																																																																																																	
Air Blank	0.000	13:13																																																																																																																																																	
Control Test	0.078	13:13																																																																																																																																																	
Air Blank	0.000	13:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0260																																																																																																																																																		
Std Dev	0.0450																																																																																																																																																		
Rel Std Dev(%)	173.2051																																																																																																																																																		

Comments:

The DGS hose was not attached to the instrument. I attached it and reran the 0.08 DGS Stability Test. MG 11/08/2022

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 7484	Hillsborough CSO	11/08/2022	TDG <i>TMG</i>

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																					
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083	≤0.003 of Wet																																				
						<div>DGS #2</div> <div>HILLSBOROUGH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007484 11/08/2022 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <div>Operator's Signature</div>		Test	g/210L	Time	Air Blank	0.000	13:15	Control Test	0.078	13:15	Air Blank	0.000	13:16	Control Test	0.078	13:16	Air Blank	0.000	13:16	Control Test	0.078	13:17	Air Blank	0.000	13:17	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																									
Air Blank	0.000	13:15																																									
Control Test	0.078	13:15																																									
Air Blank	0.000	13:16																																									
Control Test	0.078	13:16																																									
Air Blank	0.000	13:16																																									
Control Test	0.078	13:17																																									
Air Blank	0.000	13:17																																									
Control Test Stats																																											
Average	0.0780																																										
Std Dev	0.0000																																										
Rel Std Dev(%)	0.0000																																										

Comments: Nominal values obtained when DGS hose was attached.  
*TMG* 11/08/2022

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO  
Time of Inspection: 15:27

Date of Inspection: 11/08/2022

Serial Number: 80-007484  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.047	0.076	0.196	0.078
0.000	0.047	0.077	0.196	0.078
0.000	0.048	0.077	0.196	0.078
0.000	0.048	0.077	0.196	0.078
0.000	0.048	0.077	0.196	0.079
0.000	0.047	0.077	0.196	0.079
0.000	0.047	0.077	0.196	0.078
0.000	0.048	0.077	0.196	0.078
0.000	0.048	0.077	0.196	0.078
0.000	0.048	0.077	0.196	0.078


Standard Deviations	0.0005	0.0003	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/08/2022  
Date





## Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007484, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007484</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>HILLSBOROUGH CO SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/08/2022</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:27</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/08/2022

Date

  
TAYLOR D GUTSCHOW,  
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1