

## AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Instrument Serial Number: 80-000951 Agency: Hendry CSO AGENCY INSPECTION DISCREPENCY Inspection Date: 10/09/2023 **Agency Inspector:** Darrin McNeil A test was repeated without reason for repeat or the corrective action given ■ Alcohol Free/Mouth Alcohol ☐Alcohol Free Test Interferent Detect ☐ 0.080 g/210L Dry Gas □ 0.050 g/210L ☐ Agency Inspection not conducted for calendar month ☐ 0.080 g/210L Dry Gas □ 0.080 g/210L □ 0.200 g/210L □ 0.050 g/210L ☐ Lot Number and/or Expiration Date for Standards entered incorrectly □ 0.050 g/210L ☐0.080 g/210L □0.200 g/210L □ 0.080 g/210L Dry Gas ■ Other Electronic Data Review/Comments: Please include a remark describing why the tests were repeated and what corrective actions were taken prior to repeating them. REQUESTED CORRECTIVE ACTION ■ Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector ☐ Provide written explanation regarding referenced item(s) to Department Inspector ☐ Remove instrument from evidentiary use until otherwise directed by the Department Other: Digitally signed by Taylor Taylor Gutschow Gutschow Date: 2023.11.13 12:00:45 -05'00' 11/13/2023 **Signature of Alcohol Testing Program Staff Member Date** 

· Admended

### Florida Department of Law Enforcement **Alcohol Testing Program**

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY S.O. Time of Inspection: 19:49

Date of Inspection: 10/09/2023

Serial Number: 80-000951

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		***
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.046	0.073 / 0.077		
0.000	0.047	0.077 (RFI)		
0.000	0.047	0.077		

Number of Simulators Used: 5

Remarks:

A F / M A: . 08: Control Outside Tolerance, RFI Detect. Non-compliance: .

RFI Fail

ReTest

Remoded from room to prevent 1/27/23

Red. o Remoded from room to prevent 1/27/23

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

> DARRIN J MCNEIL Signature and Printed Name

10/09/2023

# Florida Department of Law Enforcement Alcohol Testing Program

#### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY S.O. Time of Inspection: 20:49

Date of Inspection: 10/09/2023

Serial Number: 80-000951

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1474747 Exp: 04/12/2024
0.000	0.046	0.074 / 0.076	0.194	0.081
0.000	0.048	0.077 / 0.077	0.196	0.081.
0.000	0.048	0.078 / 0.077	0.196	0.081

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

Simulator Lid Tishtened refest@11/27/25

The above instrument complies (  $\,$  X  $\,$  ) does not comply (  $\,$  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DARRIN J MCNEIL

Signature and Printed Name

10/09/2023 Date