

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE COUNTY SO
Time of Inspection: 10:34

Date of Inspection: ~~12/06/2022~~ ^{01/22/23}

Serial Number: 80-001321
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1474747 Exp: 04/12/2024
0.000	0.047	0.077	0.197	0.079
0.000	0.047	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079

Number of Simulators Used: 4

Remarks: *Wrong Dates*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CAI M Wright

MARK A WRIGHT

Signature and Printed Name

~~12/06/2022~~ *1/22/2023*
Date

FDLE/ATP Form 40 - March 2004