



Florida Department of  
Law Enforcement

## AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: **Florida Highway Patrol**

Instrument Serial Number: **80-006771**

### AGENCY INSPECTION DATA REVIEW

Agency Inspector: **William Coleman**

Date of Inspection: **07/18/2022**

Time of Inspection: **10:19:07**

Agency Inspection Discrepancy: ☐ Incomplete ☐ Untimely/Not Received ☐ Erroneous Information  
☐ Procedural ☒ Other (Missing Required Information)

☐ Agency Inspection Not Conducted or Records regarding Agency Inspection have not been uploaded.

☐ Lot Number ☐ Expiration Date for g/210L ☐ Alcohol Reference Solution ☐ Dry Gas Standard is ☐ Incorrect ☐ Expired.

☒ FDLE/ATP Form 39 states in part, "If a test must be repeated, the REASON must be entered when prompted and recorded in the Remarks section of FDLE/ATP Form 40 Agency Inspection Report – Intoxilyzer 8000. The ☒ REASON for repeating the following test(s); OR the ☒ Possible Cause and Corrective Action Taken on the following test(s) was not recorded:

☐ Alcohol Free Subject Test ☐ Mouth Alcohol Test ☐ Alcohol Free Test ☐ Interferent Detect Test  
☒ 0.05 g/210L Test ☐ 0.08 g/210L Test ☐ 0.20 g/210L Test ☐ 0.08 g/210L Dry Gas Standard Test

☐ FDLE/ATP Form 39 states in part, "If the instrument does not comply with the requirements of Chapter 11D-8, FAC, remove the instrument from service and notify the Department Inspector."

- ☐ The Department Inspector was not notified. However, the issue was satisfactorily corrected and the repeated Agency Inspection complies with the requirements of Chapter 11D-8, FAC.
- ☐ The Department Inspector was not notified. However, the repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was correctly removed from evidentiary use.
- ☐ The Department Inspector was not notified. The repeated Agency Inspection does not comply with the requirements of Chapter 11D-8, FAC and the instrument was not removed from evidentiary use.

☐ The Agency Inspection is noted as "Complies" when it does not comply with the requirements of Chapter 11D-8, FAC.

☐ Other: \_\_\_\_\_

### OTHER ELECTRONIC DATA REVIEW

☐ Login Records  
Date: \_\_\_\_\_

☐ Cylinder Change Records  
Date: \_\_\_\_\_

☐ Control Test Records  
Date: \_\_\_\_\_

☐ Diagnostic Check Records  
Date: \_\_\_\_\_

#### Comments:

The 0.05 ARS Test was repeated. The reason for repeating the test, as well as the corrective action taken prior to repeating the test, must be included in the Remarks.

### CORRECTIVE ACTION

☒ Record hand-written amendments on the FDLE/ATP Form 40, Agency Inspection Report, initial and date the amendments, mark the report "AMENDED", and forward a copy to the Department Inspector by **01/05/2023** (Date).

- ☐ Provide a written explanation regarding the referenced item(s) to the Department Inspector by \_\_\_\_\_ (Date).
- ☐ Upload the Agency Inspection(s).
- ☐ Remove the instrument from evidentiary use until otherwise directed by the Department.
- ☐ No action required
- ☐ Other: \_\_\_\_\_

*Taylor Gutachow*

Signature of Alcohol Testing Program Staff Member

12/5/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL

Serial Number: 80-006771

Time of Inspection: 10:19

Date of Inspection: 07/18/2022

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#: 202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 402357111 Exp: 02/10/2025
0.000	/ 0.049	0.079	0.197	0.080
0.000	/ 0.049	0.079	0.197	0.080
0.000	/ 0.049	0.079	0.197	0.080

Number of Simulators Used: 5

### Remarks:

05: RFI Detect. RFI Source was located, removed, and test re-done. (WDC)

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WILLIAM D COLEMAN

Signature and Printed Name

07/18/2022

Date