



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Desoto CSO

Instrument Serial Number: 80-007508

AGENCY INSPECTION DISCREPENCY

Agency Inspector: Michael Bellflower

Inspection Date: 01/25/2023

☐ A test was repeated without reason for repeat or the corrective action given

☐ Alcohol Free/Mouth Alcohol ☐ Alcohol Free Test ☐ Interferent Detect

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ Agency Inspection not conducted for calendar month

☐ Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☒ Lot Number and/or Expiration Date for Standards entered incorrectly

☒ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ Other Electronic Data Review/Comments:

REQUESTED CORRECTIVE ACTION

☒ Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector

☐ Provide written explanation regarding referenced item(s) to Department Inspector

☐ Remove instrument from evidentiary use until otherwise directed by the Department

☐ Other:

Taylor Gutschow

Digitally signed by Taylor

Gutschow

Date: 2023.05.17 09:23:12 -04'00'

05/17/2023

Signature of Alcohol Testing Program Staff Member

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.

Time of Inspection: 22:25

Date of Inspection: 01/25/2023

Serial Number: 80-007508

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22521080A1 Exp: 10/05/2023
0.000	0.046	0.077	0.195	0.080
0.000	0.047	0.077	0.195	0.080
0.000	0.047	0.077	0.196	0.080

Number of Simulators Used: 5

Remarks:

This lot # was a typographical error, the lot # should have been 202112A. - Lt. Michael Bellflower
M Bell

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

M Bell 20041/3413

MICHAEL D BELLFLOWER

Signature and Printed Name

01/25/2023
Date



Florida Department of
Law Enforcement

AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Desoto CSO

Instrument Serial Number: 80-007508

AGENCY INSPECTION DISCREPANCY

Agency Inspector: Michael Bellflower

Inspection Date: 02/09/2023

☒ **A test was repeated without reason for repeat or the corrective action given**

☒ Alcohol Free/Mouth Alcohol ☐ Alcohol Free Test ☐ Interferent Detect

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ **Agency Inspection not conducted for calendar month**

☐ **Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection**

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ **Lot Number and/or Expiration Date for Standards entered incorrectly**

☐ 0.050 g/210L ☐ 0.080 g/210L ☐ 0.200 g/210L ☐ 0.080 g/210L Dry Gas

☐ **Other Electronic Data Review/Comments:**

REQUESTED CORRECTIVE ACTION

☒ **Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector**

☐ **Provide written explanation regarding referenced item(s) to Department Inspector**

☐ **Remove instrument from evidentiary use until otherwise directed by the Department**

☐ **Other:**

Taylor Gutschow

Digitally signed by Taylor

Gutschow

Date: 2023.03.20 14:46:14 -04'00'

03/20/2023

Signature of Alcohol Testing Program Staff Member

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DESOTO COUNTY S.O.
Time of Inspection: 18:37

Date of Inspection: 02/09/2023

Serial Number: 80-007508
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:225108A1 Exp: 10/05/2023
0.000	0.047	0.076	0.195	0.080
0.000	0.047	0.076	0.196	0.080
0.000	0.047	0.077	0.197	0.080

Number of Simulators Used: 5

Remarks:

A F / M A: RFI Detect. - Handheld radio was in proximity to instrument.
Remove radio and test was completed.

MBF

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MBF

MICHAEL D BELLFLOWER

Signature and Printed Name

02/09/2023
Date