

# Florida Department of Law Enforcement

## Alcohol Testing Program

### AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Miami PD

Serial Number: 80-007516

Time of Inspection: 16:44

Date of Inspection: 03/14/2023

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		X
Diagnostic Check (Pre-Inspection): OK	X	
Alcohol Free Subject Test: 0.000		X
Mouth Alcohol Test: Slope Not Met		X
Interferent Detect Test: Interferent Detect		X
Diagnostic Check (Post-Inspection): OK		X


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: 5

Remarks: LOGIN ATTEMPT NOR THE INSPECTION COULD BE PERFORMED DUE TO A LOOSE KEYBOARD CONNECTOR ON THE INTOXILYZER AND THE KEYBOARD WOULD NOT FUNCTION. IT WAS TAKEN OUT OF SERVICE AND SENT TO CMI FOR REPAIR OF THE KEYBOARD CONNECTOR. IT WAS ALSO REPORTED THAT THERE WAS A TEMPATURE REGULATOR TEST FAIL ON MARCH 2, 2023 DURING A SUBJECT BREATH TEST.

The above instrument complies ( ) does not comply ( ☒ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ANGEL FERNANDEZ #28394

Signature and Printed Name

09/08/2023  
Date