

Gutschow, Taylor

From: Gutschow, Taylor
Sent: Thursday, July 6, 2023 3:17 PM
To: 'Holloway, Devarious'
Subject: 80-001067 June AI

Hello,

I noticed during my routine data review that you weren't able to obtain a passing Agency Inspection for June due to an unknown source of RFI that occurred during your initial inspection and also the follow-up inspection. If you need assistance identifying the source, please give me a call to discuss further.

Whenever a test within the Agency Inspection is repeated, you should always include a remark about the corrective action you took prior to repeating the test (aka "identified potential source of RFI and removed it"). If you aren't able to get an instrument to pass its monthly inspection, you should contact me per the procedure. This way, I can help troubleshoot, confirm if the instrument should be kept out of evidential use, and direct you to send the instrument to FDLE or a repair facility as needed. Ending the month with a failed inspection can jeopardize any breath tests on the instrument (this is just an FYI, since it appears there were no tests run on it last month).

Since the instrument failed its June inspection, it should already be out of evidential use and it should stay out of evidential use until you can get a passing inspection. If you can't get the instrument to pass its next inspection, please give me a call. Ideally, this is immediately afterward while you still have the alcohol reference solutions and simulators on hand.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.