

## Gutschow, Taylor

---

**From:** Dempster, Paul <pdempster@ccsofl.net>  
**Sent:** Monday, July 10, 2023 9:39 PM  
**To:** Gutschow, Taylor  
**Cc:** Kern, Michael; Chandler, Vanessa  
**Subject:** Charlotte County 946 & 7083

**CAUTION:** This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

The following intoxylizers are being sent to you for:

946 is being sent for annual and Purge Fail.  
7083 is being sent for Breath Tube.

Thank you,  
Cpl. Paul Dempster #0776



**Cpl. Paul R. Dempster #0776**

D-Squad Housing Supervisor  
Bureau of Detention  
Charlotte County Sheriff's Office  
Office: (941) 833-6371

[pdempster@ccsofl.net](mailto:pdempster@ccsofl.net)

[www.ccsso.org](http://www.ccsso.org)

*Integrity • Professionalism • Trust*

## Gutschow, Taylor

---

**From:** Gutschow, Taylor  
**Sent:** Friday, July 7, 2023 12:17 PM  
**To:** 'Kern, Michael'  
**Subject:** RE: Missed April AI 80-007083

Now that you mention it, I believed you also indicated the instrument was experiencing RAM Fails during Diagnostics and would not come out of Standby Mode. Please fill out the blank Form 40 in my previous email and include a remark about why the inspection couldn't be conducted, then send it back to me so I can update our records. Thanks for the prompt response.

### Taylor Gutschow

Department Inspector for the Alcohol Testing Program  
Florida Department of Law Enforcement  
Fort Myers Regional Operations Center  
4700 Terminal Drive, Suite 3  
Fort Myers, FL 33907

Desk #239-335-7500  
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

---

**From:** Kern, Michael <mkern@ccsofl.net>  
**Sent:** Friday, July 7, 2023 12:12 PM  
**To:** Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>  
**Subject:** RE: Missed April AI 80-007083

**CAUTION:** This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

7083 was sent to FDLE for an issue with the sample hose and was not received back until may.

---

**From:** Gutschow, Taylor <[TaylorGutschow@fdle.state.fl.us](mailto:TaylorGutschow@fdle.state.fl.us)>  
**Sent:** Friday, July 7, 2023 12:04 PM  
**To:** Kern, Michael <[mkern@ccsofl.net](mailto:mkern@ccsofl.net)>  
**Subject:** Missed April AI 80-007083

[EXTERNAL] This email originated outside of the CCSO. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Good morning,

It looks like instrument 7083 was used to conduct breath tests back in April (see Affected Breath Tests attachment) but never received an Agency Inspection for that month. If an Agency Inspection was conducted, please send me the Form 40. If one was not conducted, please fill out the Blank Form 40 that has been attached, include a remark about why the inspection did not occur, and then scan it back to me.

If the inspection did not occur, you should also contact your local State Attorney's Office, provide them with the list of affected breath tests, and let them know that these tests were run on a non-compliant instrument. They can contact me if they have any additional questions.

## **Taylor Gutschow**

Department Inspector for the Alcohol Testing Program  
Florida Department of Law Enforcement  
Fort Myers Regional Operations Center  
4700 Terminal Drive, Suite 3  
Fort Myers, FL 33907

Desk #239-335-7500  
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

## Gutschow, Taylor

---

**From:** Kern, Michael <mkern@ccsofl.net>  
**Sent:** Monday, April 4, 2022 2:58 AM  
**To:** Gutschow, Taylor  
**Subject:** Instruments 1739 and 7083

**CAUTION:** This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

The following instruments are being sent to FDLE Fort Myers for further testing.

1739- System failure contact CMI error

7083- Ram failure on diagnostic will not come out of standby mode.

Sgt. Kern  
Charlotte County Sheriff's Office.