



INSTRUMENT PROCESSING SHEET

Agency Columbia Sheriff's OfficeS/N 80-000775

Florida Department of Law Enforcement

Date In 10-09-2023 DI Completion Date 10/12/2023 Ship P/U H/D CMI EE

Intake By ALL _____ <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By BS _____ Date <u>10/11/2023</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>248</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202303K</td> </tr> <tr> <td>3/29/2025</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202303L</td> </tr> <tr> <td>3/29/2025</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202304C</td> </tr> <tr> <td>4/5/2025</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG325603</td> </tr> <tr> <td>9/13/2025</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K	3/29/2025	0.080	MP5089	202303L	3/29/2025	0.200	MP5090	202304C	4/5/2025	0.080 DGS	N/A	AG325603	9/13/2025	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp																			
0.050	MP5088	202303K																			
		3/29/2025																			
0.080	MP5089	202303L																			
		3/29/2025																			
0.200	MP5090	202304C																			
		4/5/2025																			
0.080 DGS	N/A	AG325603																			
		9/13/2025																			
		Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____																			

Calibration Adjustment By BS _____	Department Inspection By BS _____
---	--

Barometric Pressure Gauge <u>998</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>22460</td> <td>12/28/2024</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>22430</td> <td>11/30/2024</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>22400</td> <td>10/12/2024</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>23070</td> <td>3/6/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>06723080A5</td> <td>4/5/2025</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	22460	12/28/2024	0.100	MP6296	22430	11/30/2024	0.200	MP6297	22400	10/12/2024	0.300	MP6298	23070	3/6/2025	0.080 DGS	N/A	06723080A5	4/5/2025	Barometric Pressure ID# <u>28662 / 30793</u> Gauge <u>1010 / 1000</u> Instrument <u>998 / 998</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																						
0.000	MP6294	N/A	N/A																																						
0.040	MP6295	22460	12/28/2024																																						
0.100	MP6296	22430	11/30/2024																																						
0.200	MP6297	22400	10/12/2024																																						
0.300	MP6298	23070	3/6/2025																																						
0.080 DGS	N/A	06723080A5	4/5/2025																																						
Simulator	Serial Number																																								
0.000	MP6289																																								
Interferent	MP6290																																								
0.050	MP6291																																								
0.080	MP6292																																								
0.200	MP6293																																								

<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202303K</td> <td>3/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202303L</td> <td>3/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202304C</td> <td>4/5/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG325603</td> <td>9/13/2025</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202303K	3/29/2025	0.080	MP6292	202303L	3/29/2025	0.200	MP6293	202304C	4/5/2025	0.080 DGS	N/A	AG325603	9/13/2025	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____
Simulator	Serial #	Lot #	Expiration																		
0.050	MP6291	202303K	3/29/2025																		
0.080	MP6292	202303L	3/29/2025																		
0.200	MP6293	202304C	4/5/2025																		
0.080 DGS	N/A	AG325603	9/13/2025																		

Notes/Suggested Service: <u>At department inspection, barometric pressure was over 1% different from gauge, requires a calibration adjustment (BS 10/11/2023)</u> <u>Tech review: corrected post calibration stability 0.20 lot # (BS 10/12/2023)</u> _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <table style="width:100%;"> <tr> <td style="width:50%;"> Israel Soto <small>Digitally signed by Israel Soto Date: 2023.10.12 14:16:24 -0400</small> </td> <td style="width:50%;"> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.10.12 14:27:30 -0400</small> </td> </tr> <tr> <td>Tech Review / Date _____</td> <td>Admin Review / Date _____</td> </tr> </table>	Israel Soto <small>Digitally signed by Israel Soto Date: 2023.10.12 14:16:24 -0400</small>	Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.10.12 14:27:30 -0400</small>	Tech Review / Date _____	Admin Review / Date _____
Israel Soto <small>Digitally signed by Israel Soto Date: 2023.10.12 14:16:24 -0400</small>	Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.10.12 14:27:30 -0400</small>				
Tech Review / Date _____	Admin Review / Date _____				

Stability Checks

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000775
10/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.009	09:58
Control Test	0.048	09:59
Air Blank	0.000	10:00
Control Test	0.049	10:00
Air Blank	0.000	10:01
Control Test	0.049	10:01
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Benjamin Siddeeg
Operator's Signature

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000775
10/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.078	10:05
Air Blank	0.000	10:06
Control Test	0.078	10:07
Air Blank	0.000	10:07
Control Test	0.078	10:08
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddeeg
Operator's Signature

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000775
10/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.196	10:13
Air Blank	0.000	10:14
Control Test	0.196	10:14
Air Blank	0.000	10:15
Control Test	0.197	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

Benjamin Siddeeg
Operator's Signature

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000775
10/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.080	10:20
Air Blank	0.000	10:21
Control Test	0.080	10:21
Air Blank	0.000	10:22
Control Test	0.080	10:22
Air Blank	0.000	10:23
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Benjamin Siddeeg
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLUMBIA COUNTY SO
Time of Inspection: 10:38

Date of Inspection: 10/11/2023

Serial Number: 80-000775
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK		No	Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK		No
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:
BAROMETRIC PRESSURE >1 PERCENT DIFFERENCE
RCENT DIFFERENCE .
Non-compliance: BAROMETRIC PRESSURE >1 PE

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin W Siddoway BENJAMIN W SIDDOWAY
Signature and Printed Name

10/11/2023
Date

COLUMBIA COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

10/12/2023

08:05:14

SN 80-000775

Auto Calibration

Max Power Res Value = 104

Auto Range Res Value = 65

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5180 (-0.0150)
 Sample #2 = 1.5030 (0.0000)
 Sample #3 = 1.5130 (-0.0100)
 Sample #4 = 1.5100 (-0.0000)
 Avg % Abs = 1.5087 (-0.0033)
 STD DEV = 0.0051 (0.0058)
 REL STD DEV = 0.340 (173.205)

Sol Value = 0.000 g/210L ***

Fit Value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12525, Sum Io = 13476

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.1150 (-0.0200)

Sample #2 = 0.1270 (0.0100)

Sample #3 = 0.1140 (0.0150)

Sample #4 = 0.1290 (-0.0020)

Avg % Abs = 0.1233 (-0.0077)

STD DEV = 0.0081 (0.0087)

REL STD DEV = 6.604 (113.960)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.0860 (0.0030)

Sample #2 = 0.0860 (0.0050)

Sample #3 = 0.1070 (-0.0190)

Sample #4 = 0.1090 (-0.0330)

Avg % Abs = 0.1007 (-0.0157)

STD DEV = 0.0127 (0.0192)

REL STD DEV = 12.657 (122.668)

Sol Value = 0.040 g/210L ***

Fit Value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12518, Sum Io = 13475

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 0.8460 (-0.0160)

Sample #2 = 0.8330 (-0.0010)

Sample #3 = 0.8180 (-0.0120)

Sample #4 = 0.7970 (0.0170)

Avg % Abs = 0.8160 (0.0013)

STD DEV = 0.0181 (0.0146)

REL STD DEV = 2.216 (1098.009)

<<<<< CHANNEL 2 >>>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.7510 (-0.0070)
 Sample #2 = 6.7740 (-0.0120)
 Sample #3 = 6.7810 (-0.0190)
 Sample #4 = 6.7790 (-0.0100)
 Avg % Abs = 6.7780 (-0.0137)
 STD DEV = 0.0036 (0.0947)
 REL STD DEV = 0.053 (34.573)

Sol Value = 0.300 g/210L ***

Fit Value = 1.4286 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12504, Sum Io = 13466

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 5.1400 (-0.0060)

Sample #2 = 5.1510 (-0.0040)

Sample #3 = 5.1620 (0.0120)

Sample #4 = 5.1660 (-0.0080)

Avg % Abs = 5.1597 (-0.0000)

STD DEV = 0.0078 (0.0106)

REL STD DEV = 0.151 (68.804536.000)

<<<<< CHANNEL 2 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 9.8490 (-0.0100)

Sample #2 = 9.8280 (0.0160)

Sample #3 = 9.8240 (0.0340)

Sample #4 = 9.8050 (0.0010)

Avg % Abs = 9.8190 (0.0170)

STD DEV = 0.0123 (0.0165)

REL STD DEV = 0.125 (97.192)

Sol Value = 0.200 g/210L ***

Fit Value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12508, Sum Io = 13466

<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)

Sample #1 = 3.5170 (-0.0080)

Sample #2 = 3.5610 (-0.0200)

Sample #3 = 3.5370 (-0.0120)

Sample #4 = 3.5530 (-0.0070)

Avg % Abs = 3.5503 (-0.0130)

STD DEV = 0.0122 (0.0066)

REL STD DEV = 0.344 (50.442)

***** AUTO CAL DATA *****
 <<<<< CHANNEL 1 >>>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.123
 Std Dev = 0.01 Rel Std Dev = 6.60
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.816
 Std Dev = 0.02 Rel Std Dev = 2.22
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.958
 Std Dev = 0.01 Rel Std Dev = 0.35

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.550

Std Dev = 0.01 Rel Std Dev = 0.34

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.160

Std Dev = 0.01 Rel Std Dev = 0.15

Zero Order Coef = -311.43

First Order Coef = 2665.73

Second Order Coef = 31.31

Standard Deviation = 19.492929

<<<<< CHANNEL 2 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.101

Std Dev = 0.01 Rel Std Dev = 12.66

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.509

Std Dev = 0.01 Rel Std Dev = 0.34

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.553

Std Dev = 0.03 Rel Std Dev = 0.85

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.778

Std Dev = 0.00 Rel Std Dev = 0.05

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 9.819

Std Dev = 0.01 Rel Std Dev = 0.13

Zero Order Coef = -132.51

First Order Coef = 1326.77

Second Order Coef = 14.42

Standard Deviation = 1.904962

Solution Stats Quadratic Fit Chan 1
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.000 -0.0004
 0.040 0.040 0.0004
 0.100 0.100 0.0003
 0.200 0.200 -0.0005
 0.300 0.300 0.0002

Solution Stats Quadratic Fit Chan 2
 Act Fit Residual
 g/210L g/210L g/210L
 0.000 0.000 -0.0000
 0.040 0.040 0.0001
 0.100 0.100 -0.0000
 0.200 0.200 0.0000
 0.300 0.300 0.0000

Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 3064.00
 Sample #2 = 2927.00
 Sample #3 = 2880.00
 Sample #4 = 2921.00
 Average Result = 2909.3333
 STD DEV = 25.5799
 REL STD DEV = 0.879

***** CHANNEL 2 *****
 Sample #1 = 3409.00
 Sample #2 = 3345.00
 Sample #3 = 3336.00
 Sample #4 = 3343.00
 Average Result = 3341.3333
 STD DEV = 4.7258
 REL STD DEV = 0.141

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 998
 3 um H2O Adjust (mg/l*10.000) = 900
 9 um H2O Adjust (mg/l*10.000) = 458
 ***** AUTO CAL PASS *****

Calibration Adjustment

Post Calibration Adjustment Stability Checks

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000775
10/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:28
Control Test	0.049	09:29
Air Blank	0.000	09:29
Control Test	0.049	09:30
Air Blank	0.000	09:31
Control Test	0.049	09:31
Air Blank	0.000	09:32
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddle
Operator's Signature

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000775
10/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	0.077	09:34
Air Blank	0.000	09:34
Control Test	0.078	09:35
Air Blank	0.000	09:36
Control Test	0.078	09:36
Air Blank	0.000	09:37
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

Benjamin Siddle
Operator's Signature

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000775
10/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:38
Control Test	0.196	09:39
Air Blank	0.000	09:39
Control Test	0.196	09:40
Air Blank	0.000	09:40
Control Test	0.196	09:41
Air Blank	0.000	09:42
Control Test Stats		
Average	0.1960	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddle
Operator's Signature

COLUMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000775
10/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:43
Control Test	0.079	09:43
Air Blank	0.000	09:44
Control Test	0.080	09:44
Air Blank	0.000	09:45
Control Test	0.079	09:45
Air Blank	0.000	09:46
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

Benjamin Siddle
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLUMBIA COUNTY SO
Time of Inspection: 12:09

Date of Inspection: 10/12/2023

Serial Number: 80-000775
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG325603 Exp: 09/13/2025
0.000	0.049	0.078	0.195	0.078
0.000	0.049	0.077	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.078	0.195	0.079
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.050	0.079	0.199	0.080
0.000	0.048	0.079	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.079

Standard Deviations	0.0004	0.0010	0.0017	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0009 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin Siddoway BENJAMIN W SIDDOWAY
Signature and Printed Name

10/12/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000775 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000775</u>	UNCERTAINTY* ±	
Owning Agency:	<u>COLUMBIA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/12/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:09</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Digitally signed by Benjamin Siddoway
Date: 2023.10.12 13:54:53 -04'00'

10/12/2023

Date

BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality