

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD (LAKE SO)
Time of Inspection: 09:46

Date of Inspection: 01/09/2023

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT.

Not determined ^{NG} 01/09/2023

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

01/09/2023
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000836	Umatilla PD (Lake SO)	01/09/2023	TDG ML

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																												
0.047 to 0.053 ✓	0.077 to 0.083 ✓	0.194 to 0.206 ✓	0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																												
UMATILLA PD(LAKE SO) Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000836 01/09/2023 Software: 8100.27	UMATILLA PD(LAKE SO) Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000836 01/09/2023 Software: 8100.27	UMATILLA PD(LAKE SO) Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000836 01/09/2023 Software: 8100.27	Inadvertently printed to external printer. Results in acceptable range. Will attach. ML 01/09/2023																																																																																																												
<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:40</td></tr><tr><td>Control Test</td><td>0.051</td><td>11:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr><tr><td>Control Test</td><td>0.050</td><td>11:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:44</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0503</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	11:40	Control Test	0.051	11:41	Air Blank	0.000	11:42	Control Test	0.050	11:42	Air Blank	0.000	11:43	Control Test	0.050	11:44	Air Blank	0.000	11:44	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:46</td></tr><tr><td>Control Test</td><td>0.080</td><td>11:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:47</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:48</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	11:46	Control Test	0.080	11:47	Air Blank	0.000	11:47	Control Test	0.079	11:48	Air Blank	0.000	11:48	Control Test	0.078	11:49	Air Blank	0.000	11:50	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr><tr><td>Control Test</td><td>0.200</td><td>11:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:59</td></tr><tr><td>Control Test</td><td>0.199</td><td>12:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:00</td></tr><tr><td>Control Test</td><td>0.200</td><td>12:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:02</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.1997</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2892</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	11:58	Control Test	0.200	11:59	Air Blank	0.000	11:59	Control Test	0.199	12:00	Air Blank	0.000	12:00	Control Test	0.200	12:01	Air Blank	0.000	12:02	Control Test Stats			Average	0.1997		Std Dev	0.0006		Rel Std Dev(%)	0.2892		
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:40																																																																																																													
Control Test	0.051	11:41																																																																																																													
Air Blank	0.000	11:42																																																																																																													
Control Test	0.050	11:42																																																																																																													
Air Blank	0.000	11:43																																																																																																													
Control Test	0.050	11:44																																																																																																													
Air Blank	0.000	11:44																																																																																																													
Control Test Stats																																																																																																															
Average	0.0503																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	1.1471																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:46																																																																																																													
Control Test	0.080	11:47																																																																																																													
Air Blank	0.000	11:47																																																																																																													
Control Test	0.079	11:48																																																																																																													
Air Blank	0.000	11:48																																																																																																													
Control Test	0.078	11:49																																																																																																													
Air Blank	0.000	11:50																																																																																																													
Control Test Stats																																																																																																															
Average	0.0790																																																																																																														
Std Dev	0.0010																																																																																																														
Rel Std Dev(%)	1.2658																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	11:58																																																																																																													
Control Test	0.200	11:59																																																																																																													
Air Blank	0.000	11:59																																																																																																													
Control Test	0.199	12:00																																																																																																													
Air Blank	0.000	12:00																																																																																																													
Control Test	0.200	12:01																																																																																																													
Air Blank	0.000	12:02																																																																																																													
Control Test Stats																																																																																																															
Average	0.1997																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.2892																																																																																																														
ML Operator's Signature	ML Operator's Signature	ML Operator's Signature																																																																																																													

Comments:

UMATILLA PD(LAKE SO)
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000836
01/09/2023
Software: 8100.27

DBS

Test	g/210L	Time
Air Blank	0.000	11:34
Control Test	0.080	11:34
Air Blank	0.000	11:35
Control Test	0.080	11:35
Air Blank	0.000	11:36
Control Test	0.079	11:36
Air Blank	0.000	11:36
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

ML

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: UMATILLA PD(LAKE SO)
Time of Inspection: 15:20

Date of Inspection: 01/09/2023

Serial Number: 80-000836
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.200	0.079
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.079	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.079	0.200	0.078
0.000	0.050	0.079	0.200	0.078
0.000	0.050	0.080	0.200	0.078


Standard Deviations	0.0000	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/09/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000836, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000836</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>UMATILLA PD(LAKE SO)</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/09/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:20</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/09/2023

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1