

## INSTRUMENT PROCESSING SHEET

Agency Monroe CSO

S/N 80-000867

Florida Department of  
Law Enforcement

Date In 06/09/2023 DI Completion Date 06/14/2023

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

<b>Intake</b> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	<b>Quality Checks</b> By <u>TDG</u> Date <u>06/13/2023</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>260</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.261</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> By _____      Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5094</td> <td>202201C</td> </tr> <tr> <td>01/11/2024</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5095</td> <td>202201D</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5096</td> <td>202201E</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG223802</td> </tr> <tr> <td>08/26/2024</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C	01/11/2024	0.080	MP5095	202201D	01/18/2024	0.200	MP5096	202201E	01/18/2024	0.080 DGS	N/A	AG223802	08/26/2024	<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Maintenance</th> <th>By <u>TDG</u></th> </tr> </thead> <tbody> <tr> <td><input type="checkbox"/> Battery Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Dry Gas Regulator Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Breath Tube Replacement</td> <td></td> </tr> <tr> <td><input checked="" type="checkbox"/> Other <u>Tank Sensor Cal</u></td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> <tr> <td> </td> <td></td> </tr> </tbody> </table>	Maintenance	By <u>TDG</u>	<input type="checkbox"/> Battery Replacement		<input type="checkbox"/> Dry Gas Regulator Replacement		<input type="checkbox"/> Breath Tube Replacement		<input checked="" type="checkbox"/> Other <u>Tank Sensor Cal</u>																																	
Simulator	Serial #	Lot #/Exp																																																													
0.050	MP5094	202201C																																																													
		01/11/2024																																																													
0.080	MP5095	202201D																																																													
		01/18/2024																																																													
0.200	MP5096	202201E																																																													
		01/18/2024																																																													
0.080 DGS	N/A	AG223802																																																													
		08/26/2024																																																													
Maintenance	By <u>TDG</u>																																																														
<input type="checkbox"/> Battery Replacement																																																															
<input type="checkbox"/> Dry Gas Regulator Replacement																																																															
<input type="checkbox"/> Breath Tube Replacement																																																															
<input checked="" type="checkbox"/> Other <u>Tank Sensor Cal</u>																																																															
<b>Calibration Adjustment</b> By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> By <u>TDG</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1014</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5092</td> </tr> <tr> <td>Interferent</td> <td>MP5093</td> </tr> <tr> <td>0.050</td> <td>MP5094</td> </tr> <tr> <td>0.080</td> <td>MP5095</td> </tr> <tr> <td>0.200</td> <td>MP5096</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP5092																																																														
Interferent	MP5093																																																														
0.050	MP5094																																																														
0.080	MP5095																																																														
0.200	MP5096																																																														
Notes/Suggested Service: <u>Performed tank sensor cal prior to flow cal and baro pressure check on 6/13. (TDG)</u> _____ _____ _____ _____ _____ _____ _____ _____ _____		<b>Attachments</b> <div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input checked="" type="checkbox"/> Form 41  <input checked="" type="checkbox"/> Stability Checks  <input checked="" type="checkbox"/> Calibration Certificate  <input type="checkbox"/> Calibration Adjustment         </div> <div style="width: 48%;"> <input type="checkbox"/> Post-Stability Checks  <input type="checkbox"/> Flow Calibration  <input type="checkbox"/> Form 40  <input type="checkbox"/> Other _____         </div> </div>																																																													
_____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													
_____ _____ _____ _____ _____ _____ _____ _____ _____		<div style="display: flex; justify-content: space-between;"> <div style="width: 48%;">           Israel Soto      Digitally signed by Israel Soto  <small>Date: 2023.06.14 14:31:17 -0400</small> </div> <div style="width: 48%;">           Phil Nicodemo      Digitally signed by Phil Nicodemo  <small>Date: 2023.06.15 11:14:05 -0400</small> </div> </div> <div style="display: flex; justify-content: space-between; margin-top: 10px;"> <div style="width: 48%;">Tech Review / Date _____</div> <div style="width: 48%;">Admin Review / Date _____</div> </div>																																																													



Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000867	Monroe CSD	06/13/2023	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
<p>MONROE COUNTY SO Intoxilizer - Alcolon Analyzer Model: 8000 06/13/2023 Software: 8100.27</p> <p>SN 80-000867</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr><tr><td>Control Test</td><td>0.049</td><td>13:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr><tr><td>Control Test</td><td>0.049</td><td>13:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr><tr><td>Control Test</td><td>0.049</td><td>13:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	13:07	Control Test	0.049	13:08	Air Blank	0.000	13:08	Control Test	0.049	13:09	Air Blank	0.000	13:09	Control Test	0.049	13:10	Air Blank	0.000	13:11	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>MONROE COUNTY SO Intoxilizer - Alcolon Analyzer Model: 8000 06/13/2023 Software: 8100.27</p> <p>SN 80-000867</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr><tr><td>Control Test</td><td>0.077</td><td>13:16</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:18</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0777</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel. Std Dev(%)</td><td>0.7434</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	13:15	Control Test	0.077	13:16	Air Blank	0.000	13:16	Control Test	0.078	13:17	Air Blank	0.000	13:17	Control Test	0.078	13:18	Air Blank	0.000	13:18	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel. Std Dev(%)	0.7434		<p>MONROE COUNTY SO Intoxilizer - Alcolon Analyzer Model: 8000 06/13/2023 Software: 8100.27</p> <p>SN 80-000867</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr><tr><td>Control Test</td><td>0.197</td><td>13:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:23</td></tr><tr><td>Control Test</td><td>0.197</td><td>13:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr><tr><td>Control Test</td><td>0.197</td><td>13:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1970</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	13:22	Control Test	0.197	13:22	Air Blank	0.000	13:23	Control Test	0.197	13:24	Air Blank	0.000	13:24	Control Test	0.197	13:25	Air Blank	0.000	13:25	Control Test Stats			Average	0.1970		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>MONROE COUNTY SO Intoxilizer - Alcolon Analyzer Model: 8000 06/13/2023 Software: 8100.27</p> <p>SN 80-000867</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr><tr><td>Control Test</td><td>0.079</td><td>13:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table>	Test	g/210L	Time	Air Blank	0.000	12:58	Control Test	0.079	12:58	Air Blank	0.000	12:59	Control Test	0.079	12:59	Air Blank	0.000	13:00	Control Test	0.079	13:00	Air Blank	0.000	13:00	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:07																																																																																																																																																	
Control Test	0.049	13:08																																																																																																																																																	
Air Blank	0.000	13:08																																																																																																																																																	
Control Test	0.049	13:09																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test	0.049	13:10																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:15																																																																																																																																																	
Control Test	0.077	13:16																																																																																																																																																	
Air Blank	0.000	13:16																																																																																																																																																	
Control Test	0.078	13:17																																																																																																																																																	
Air Blank	0.000	13:17																																																																																																																																																	
Control Test	0.078	13:18																																																																																																																																																	
Air Blank	0.000	13:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:22																																																																																																																																																	
Control Test	0.197	13:22																																																																																																																																																	
Air Blank	0.000	13:23																																																																																																																																																	
Control Test	0.197	13:24																																																																																																																																																	
Air Blank	0.000	13:24																																																																																																																																																	
Control Test	0.197	13:25																																																																																																																																																	
Air Blank	0.000	13:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1970																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.079	12:58																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.079	12:59																																																																																																																																																	
Air Blank	0.000	13:00																																																																																																																																																	
Control Test	0.079	13:00																																																																																																																																																	
Air Blank	0.000	13:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
<p><i>ML</i></p> <p>Operator's Signature</p>	<p><i>ML</i></p> <p>Operator's Signature</p>	<p><i>ML</i></p> <p>Operator's Signature</p>	<p><i>ML</i></p> <p>Operator's Signature</p>																																																																																																																																																

Comments:

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO  
Time of Inspection: 11:40

Date of Inspection: 06/14/2023

Serial Number: 80-000867  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.048	0.077	0.199	0.078
0.000	0.048	0.078	0.200	0.079
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.199	0.078
0.000	0.050	0.078	0.200	0.078
0.000	0.050	0.078	0.200	0.078


Standard Deviations	0.0006	0.0003	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/14/2023  
Date





# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000867, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000867	UNCERTAINTY* $\pm$
Owning Agency:	MONROE COUNTY SO	0.050 g/210 L 0.004
Calibration Date:	06/14/2023	0.080 g/210 L 0.004
Calibration Time:	11:40	0.200 g/210 L 0.007
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/14/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality