



# Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
<p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000881 10/09/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:25</p> <p>Control Test 0.048 09:26</p> <p>Air Blank 0.000 09:26</p> <p>Control Test 0.048 09:27</p> <p>Air Blank 0.000 09:27</p> <p>Control Test 0.048 09:28</p> <p>Air Blank 0.000 09:29</p> <p>Control Test Stats</p> <p>Average 0.0480</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000881 10/09/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:33</p> <p>Control Test 0.077 09:34</p> <p>Air Blank 0.000 09:34</p> <p>Control Test 0.077 09:35</p> <p>Air Blank 0.000 09:35</p> <p>Control Test 0.077 09:36</p> <p>Air Blank 0.000 09:36</p> <p>Control Test Stats</p> <p>Average 0.0770</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000881 10/09/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:40</p> <p>Control Test 0.198 09:41</p> <p>Air Blank 0.000 09:42</p> <p>Control Test 0.200 09:42</p> <p>Air Blank 0.000 09:43</p> <p>Control Test 0.200 09:44</p> <p>Air Blank 0.000 09:44</p> <p>Control Test Stats</p> <p>Average 0.1993</p> <p>Std Dev 0.0012</p> <p>Rel Std Dev(%) 0.5793</p>	<p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000881 10/09/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:45</p> <p>Control Test 0.081 09:46</p> <p>Air Blank 0.000 09:46</p> <p>Control Test 0.080 09:46</p> <p>Air Blank 0.000 09:47</p> <p>Control Test 0.080 09:47</p> <p>Air Blank 0.000 09:48</p> <p>Control Test Stats</p> <p>Average 0.0803</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7187</p>
<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>



MIAMI DPEE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
10/10/2023  
SN 80-000891  
10:21:22

Sample % Abs (% Abs Ref)  
Sample #1 = 1.4580 (-0.0120)  
Sample #2 = 1.4640 (-0.0180)  
Sample #3 = 1.4490 (-0.0200)  
Sample #4 = 1.4450 (-0.0100)  
Avg % Abs = 1.4527 (-0.0160)  
STD DEV = 0.0100 (-0.0053)  
REL STD DEV = 0.690 (-33.072)

Auto Calibration  
Max Power Res Value = 58  
Auto Range Res Value = 29

Sol Value = 0.000 g/210L \*\*\*  
Fit Value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12688, Sum Io = 13867  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.9160 (-0.0260)  
Sample #2 = 1.9100 (-0.0120)  
Sample #3 = 1.9020 (-0.0100)  
Sample #4 = 1.9060 (-0.0540)  
Avg % Abs = 1.9060 (-0.0173)  
STD DEV = 0.0040 (-0.0336)  
REL STD DEV = 0.210 (-193.878)

Sample % Abs (% Abs Ref)  
Sample #1 = 3.4300 (-0.0110)  
Sample #2 = 3.4120 (-0.0030)  
Sample #3 = 3.4260 (-0.0020)  
Sample #4 = 3.4220 (-0.0260)  
Avg % Abs = 3.4200 (-0.0090)  
STD DEV = 0.0072 (-0.0149)  
REL STD DEV = 0.211 (-165.924)

Sol Value = 0.200 g/210L \*\*\*  
Fit Value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12583, Sum Io = 13811  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.5960 (-0.0260)  
Sample #2 = 3.6090 (-0.0250)  
Sample #3 = 3.5950 (-0.0130)  
Sample #4 = 3.6200 (-0.0050)  
Avg % Abs = 3.6080 (-0.0057)  
STD DEV = 0.0125 (-0.0190)  
REL STD DEV = 0.347 (-335.449)

Sol Value = 0.040 g/210L \*\*\*  
Fit Value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12616, Sum Io = 13830  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8140 (-0.0300)  
Sample #2 = 0.8330 (-0.0010)  
Sample #3 = 0.8260 (-0.0300)  
Sample #4 = 0.8440 (-0.0190)  
Avg % Abs = 0.8343 (-0.0160)  
STD DEV = 0.0091 (-0.0157)  
REL STD DEV = 1.088 (-98.226)

Sample % Abs (% Abs Ref)  
Sample #1 = 6.5380 (-0.0070)  
Sample #2 = 6.5200 (-0.0110)  
Sample #3 = 6.5230 (-0.0240)  
Sample #4 = 6.5060 (-0.0140)  
Avg % Abs = 6.5163 (-0.0163)  
STD DEV = 0.0091 (-0.0068)  
REL STD DEV = 0.139 (-41.675)

Sol Value = 0.300 g/210L \*\*\*  
Fit Value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12577, Sum Io = 13808  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.3060 (-0.0190)  
Sample #2 = 5.3330 (-0.0070)  
Sample #3 = 5.3110 (-0.0050)  
Sample #4 = 5.3030 (-0.0520)  
Avg % Abs = 5.3157 (-0.0213)  
STD DEV = 0.0155 (-0.0266)  
REL STD DEV = 0.292 (-124.579)

Sample % Abs (% Abs Ref)  
Sample #1 = 9.5190 (-0.0060)  
Sample #2 = 9.5190 (-0.0130)  
Sample #3 = 9.5190 (-0.0110)  
Sample #4 = 9.4980 (-0.0250)  
Avg % Abs = 9.5120 (-0.0163)  
STD DEV = 0.0121 (-0.0076)  
REL STD DEV = 0.127 (-46.358)

Sol Value = 0.200 g/210L \*\*\*  
Fit Value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12583, Sum Io = 13811  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.5960 (-0.0260)  
Sample #2 = 3.6090 (-0.0250)  
Sample #3 = 3.5950 (-0.0130)  
Sample #4 = 3.6200 (-0.0050)  
Avg % Abs = 3.6080 (-0.0057)  
STD DEV = 0.0125 (-0.0190)  
REL STD DEV = 0.347 (-335.449)

Optical Calibration  
Adjustment

By: TDG

Solution Stats Quadratic Fit Chan 2  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 -0.000 0.0001  
0.040 0.040 -0.0000  
0.100 0.100 -0.0004  
0.200 0.200 0.0005  
0.300 0.300 -0.0001

Sol Value = 0.000 g/210L \*\*\*  
Fit Value = 0.3610 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 3078.00  
Sample #2 = 3111.00  
Sample #3 = 3184.00  
Sample #4 = 3147.00  
Average Result = 3147.3333  
STD DEV = 36.5011  
REL STD DEV = 1.160

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample #1 = 3360.00  
Sample #2 = 3312.00  
Sample #3 = 3350.00  
Sample #4 = 3317.00  
Average Result = 3326.3333  
STD DEV = 20.6478  
REL STD DEV = 0.621

Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1014  
3 um H2O Adjust (mg/l\*10,000) = 662  
9 um H2O Adjust (mg/l\*10,000) = 483  
\*\*\*\*\* AUTO CAL PASS \*\*\*\*\*

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.108  
Std Dev = 0.01 Rel Std Dev = 6.18  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.834  
Std Dev = 0.01 Rel Std Dev = 1.09  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.906  
Std Dev = 0.00 Rel Std Dev = 0.21  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.608  
Std Dev = 0.01 Rel Std Dev = 0.35  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.316  
Std Dev = 0.02 Rel Std Dev = 0.29  
Zero Order Coef = -301.34  
First Order Coef = 2636.33  
Second Order Coef = 20.80  
Standard Deviation = 30.656026

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.107  
Std Dev = 0.01 Rel Std Dev = 12.15  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.453  
Std Dev = 0.01 Rel Std Dev = 0.69  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.420  
Std Dev = 0.01 Rel Std Dev = 0.21  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.516  
Std Dev = 0.01 Rel Std Dev = 0.14  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.512  
Std Dev = 0.01 Rel Std Dev = 0.13  
Zero Order Coef = -157.09  
First Order Coef = 1402.10  
Second Order Coef = 12.30  
Standard Deviation = 15.682371

Solution Stats Quadratic Fit Chan 1  
Act Fit Residual  
g/210L g/210L g/210L  
0.000 -0.000 0.0004  
0.040 0.040 -0.0002  
0.100 0.101 -0.0008  
0.200 0.199 0.0009  
0.300 0.300 -0.0003

# Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
<p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-000861</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:36</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:35	Control Test	0.049	11:36	Air Blank	0.000	11:36	Control Test	0.048	11:37	Air Blank	0.000	11:37	Control Test	0.049	11:38	Air Blank	0.000	11:39	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-000861</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:45</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:42	Control Test	0.078	11:43	Air Blank	0.000	11:43	Control Test	0.078	11:44	Air Blank	0.000	11:45	Control Test	0.078	11:45	Air Blank	0.000	11:46	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-000861</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test</td><td>0.197</td><td>11:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:52</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5822</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:49	Control Test	0.199	11:50	Air Blank	0.000	11:50	Control Test	0.197	11:51	Air Blank	0.000	11:52	Control Test	0.199	11:52	Air Blank	0.000	11:53	Control Test Stats			Average	0.1983		Std Dev	0.0012		Rel Std Dev(%)	0.5822		<p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-000861</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.081</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2500</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:32	Control Test	0.081	11:32	Air Blank	0.000	11:33	Control Test	0.080	11:33	Air Blank	0.000	11:33	Control Test	0.079	11:34	Air Blank	0.000	11:34	Control Test Stats			Average	0.0800		Std Dev	0.0010		Rel Std Dev(%)	1.2500	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:35																																																																																																																																																	
Control Test	0.049	11:36																																																																																																																																																	
Air Blank	0.000	11:36																																																																																																																																																	
Control Test	0.048	11:37																																																																																																																																																	
Air Blank	0.000	11:37																																																																																																																																																	
Control Test	0.049	11:38																																																																																																																																																	
Air Blank	0.000	11:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:42																																																																																																																																																	
Control Test	0.078	11:43																																																																																																																																																	
Air Blank	0.000	11:43																																																																																																																																																	
Control Test	0.078	11:44																																																																																																																																																	
Air Blank	0.000	11:45																																																																																																																																																	
Control Test	0.078	11:45																																																																																																																																																	
Air Blank	0.000	11:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:49																																																																																																																																																	
Control Test	0.199	11:50																																																																																																																																																	
Air Blank	0.000	11:50																																																																																																																																																	
Control Test	0.197	11:51																																																																																																																																																	
Air Blank	0.000	11:52																																																																																																																																																	
Control Test	0.199	11:52																																																																																																																																																	
Air Blank	0.000	11:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5822																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:32																																																																																																																																																	
Control Test	0.081	11:32																																																																																																																																																	
Air Blank	0.000	11:33																																																																																																																																																	
Control Test	0.080	11:33																																																																																																																																																	
Air Blank	0.000	11:33																																																																																																																																																	
Control Test	0.079	11:34																																																																																																																																																	
Air Blank	0.000	11:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2500																																																																																																																																																		
<p>Operator's Signature</p> <p><i>[Signature]</i></p>	<p>Operator's Signature</p> <p><i>[Signature]</i></p>	<p>Operator's Signature</p> <p><i>[Signature]</i></p>	<p>Operator's Signature</p> <p><i>[Signature]</i></p>																																																																																																																																																



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD  
Time of Inspection: 14:39

Date of Inspection: 10/10/2023

Serial Number: 80-000881  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.048	0.078	0.199	0.082
0.000	0.049	0.077	0.198	0.082
0.000	0.048	0.078	0.199	0.082
0.000	0.048	0.077	0.199	0.081
0.000	0.048	0.077	0.199	0.082
0.000	0.049	0.078	0.198	0.082
0.000	0.049	0.078	0.198	0.082
0.000	0.049	0.078	0.198	0.081
0.000	0.048	0.078	0.198	0.082
0.000	0.049	0.078	0.198	0.082


Standard Deviations	0.0005	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 TAYLOR D GUTSCHOW  
Signature and Printed Name

10/10/2023  
Date





# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000881, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000881</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>MIAMI DADE PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/10/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:39</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



10/10/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality