



INSTRUMENT PROCESSING SHEET

Agency Pasco County Sheriff's OfficeS/N 80-001074Florida Department of
Law EnforcementDate In 02-07-2023 DI Completion Date 02-07-2023☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date <u>02-07-2023</u>	Flow Calibration	By	Date																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Left keyboard tray hook is broken.		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>103</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks			<input type="checkbox"/> Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG113403 05-14-2023</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG113403 05-14-2023	Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____															
Simulator	Serial #	Lot #/Exp																																	
0.050	MP6291	202201C 01-11-2024																																	
0.080	MP6292	202201D 01-18-2024																																	
0.200	MP6293	202201E 01-18-2024																																	
0.080 DGS	N/A	AG113403 05-14-2023																																	
Calibration Adjustment By _____				Department Inspection By IS _____																															
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>30793</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>2023</u> Instrument <u>2020</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293																
Simulator	Serial Number																																		
0.000	MP6289																																		
Interferent	MP6290																																		
0.050	MP6291																																		
0.080	MP6292																																		
0.200	MP6293																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																															
Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2023.02.07 13:26:26 -05'00'</small>				Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.02.07 14:26:35 -05'00'</small>																															
Tech Review / Date				Admin Review / Date																															

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PASCO COUNTY SO

Time of Inspection: 12:35

Date of Inspection: 02/07/2023

Serial Number: 80-001074

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.049	0.080	0.202	0.080
0.000	0.049	0.080	0.201	0.081
0.000	0.049	0.080	0.202	0.080
0.000	0.049	0.080	0.202	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.049	0.080	0.202	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.050	0.080	0.202	0.080
0.000	0.049	0.080	0.202	0.080

Standard Deviations	0.0004	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

A F / M A: DID MA FIRST.

Performed MA test first, repeated test in
correct order. ✓

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

02/07/2023

Date

stability checks

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
02/07/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:04
Control Test	0.050	10:05
Air Blank	0.000	10:05
Control Test	0.049	10:06
Air Blank	0.000	10:06
Control Test	0.049	10:07
Air Blank	0.000	10:08
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
02/07/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:10
Control Test	0.080	10:10
Air Blank	0.000	10:11
Control Test	0.079	10:12
Air Blank	0.000	10:12
Control Test	0.079	10:13
Air Blank	0.000	10:13
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
02/07/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.202	10:16
Air Blank	0.000	10:17
Control Test	0.200	10:18
Air Blank	0.000	10:18
Control Test	0.201	10:19
Air Blank	0.000	10:20
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

Operator's Signature

PASCO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001074
02/07/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:22
Control Test	0.081	10:22
Air Blank	0.000	10:23
Control Test	0.081	10:23
Air Blank	0.000	10:23
Control Test	0.082	10:24
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

Operator's Signature

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001074, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001074</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>PASCO COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/07/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:35</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/07/2023

Date

Israel Soto

ISRAEL SOTO,
Department Inspector

Digitally signed by Israel

Soto

Date: 2023.02.07

13:11:55 -05'00'

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1