



INSTRUMENT PROCESSING SHEET

Agency FHP Troop ES/N 80-001120Florida Department of
Law EnforcementDate In 02/16/2023 DI Completion Date 03/01/2023☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>03/01/2023</u>	Flow Calibration	By	Date																																																																																																																																																																
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Keyboard does not work.</u> <u>Used another keyboard supplied</u> <u>by the agency. Agency wants to</u> <u>deregister and use as a trainer.</u> <u>Needs battery change, but since</u> <u>it will no longer be utilized as an</u> <u>evidential instrument, the battery</u> <u>will not be changed with consent</u> <u>from the Agency Inspector.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>210</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																																																																																																																																		
				<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5092</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5093</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5094</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024	<table border="1"><thead><tr><th>Maintenance</th><th>By</th></tr></thead><tbody><tr><td><input type="checkbox"/> Battery Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Other _____</td><td></td></tr><tr><td> </td><td></td></tr><tr><td> </td><td></td></tr><tr><td> </td><td></td></tr><tr><td> </td><td></td></tr></tbody></table>			Maintenance	By	<input type="checkbox"/> Battery Replacement		<input type="checkbox"/> Dry Gas Regulator Replacement		<input type="checkbox"/> Breath Tube Replacement		<input type="checkbox"/> Other _____																																																																																																																																								
Simulator	Serial #	Lot #/Exp																																																																																																																																																																					
0.050	MP5092	202201C 01/11/2024																																																																																																																																																																					
0.080	MP5093	202201D 01/18/2024																																																																																																																																																																					
0.200	MP5094	202201E 01/18/2024																																																																																																																																																																					
0.080 DGS	N/A	AG223802 08/26/2024																																																																																																																																																																					
Maintenance	By																																																																																																																																																																						
<input type="checkbox"/> Battery Replacement																																																																																																																																																																							
<input type="checkbox"/> Dry Gas Regulator Replacement																																																																																																																																																																							
<input type="checkbox"/> Breath Tube Replacement																																																																																																																																																																							
<input type="checkbox"/> Other _____																																																																																																																																																																							
<table border="1"><thead><tr><th colspan="4">Calibration Adjustment</th><th>By</th></tr></thead><tbody><tr><td colspan="4">Barometric Pressure Gauge _____ ID # _____</td><td></td></tr><tr><td>Simulator</td><td>Serial #</td><td>Lot #</td><td>Expiration</td><td></td></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td><td></td></tr><tr><td>0.040</td><td></td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td><td></td></tr><tr><td colspan="4"><input type="checkbox"/> Post Calibration Adjustment Stability Checks</td><td></td></tr><tr><td>Simulator</td><td>Serial #</td><td>Lot #</td><td>Expiration</td><td></td></tr><tr><td>0.050</td><td></td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td><td></td></tr></tbody></table>				Calibration Adjustment				By	Barometric Pressure Gauge _____ ID # _____					Simulator	Serial #	Lot #	Expiration		0.000		N/A	N/A		0.040					0.100					0.200					0.300					0.080 DGS	N/A				<input type="checkbox"/> Post Calibration Adjustment Stability Checks					Simulator	Serial #	Lot #	Expiration		0.050					0.080					0.200					0.080 DGS	N/A				<table border="1"><thead><tr><th colspan="2">Department Inspection</th><th>By</th><th>TDG</th></tr></thead><tbody><tr><td colspan="2">Barometric Pressure ID# <u>28663</u></td><td></td><td></td></tr><tr><td colspan="2">Gauge <u>1020</u> Instrument <u>1017</u></td><td></td><td></td></tr><tr><td colspan="2">Mouth Alcohol Solution Lot # <u>2021-D</u></td><td></td><td></td></tr><tr><td colspan="2">Acetone Stock Solution Lot # <u>2021-C</u></td><td></td><td></td></tr><tr><td>Simulator</td><td>Serial Number</td><td></td><td></td></tr><tr><td>0.000</td><td>MP5095</td><td></td><td></td></tr><tr><td>Interferent</td><td>MP5097</td><td></td><td></td></tr><tr><td>0.050</td><td>MP5092</td><td></td><td></td></tr><tr><td>0.080</td><td>MP5093</td><td></td><td></td></tr><tr><td>0.200</td><td>MP5094</td><td></td><td></td></tr><tr><td colspan="2"><table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td><input checked="" type="checkbox"/> Form 41</td><td><input type="checkbox"/> Post-Stability Checks</td></tr><tr><td><input checked="" type="checkbox"/> Stability Checks</td><td><input type="checkbox"/> Flow Calibration</td></tr><tr><td><input checked="" type="checkbox"/> Calibration Certificate</td><td><input checked="" type="checkbox"/> Form 40</td></tr><tr><td><input type="checkbox"/> Calibration Adjustment</td><td><input type="checkbox"/> Other _____</td></tr></tbody></table></td><td></td><td></td></tr><tr><td colspan="2"><input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC</td><td></td><td></td></tr><tr><td colspan="2"><input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC</td><td></td><td></td></tr><tr><td colspan="2"><input type="checkbox"/> Return to/Place into Evidentiary Use</td><td></td><td></td></tr><tr><td colspan="2"><input checked="" type="checkbox"/> Remain Out of Evidentiary Use</td><td></td><td></td></tr><tr><td colspan="2"><input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use</td><td></td><td></td></tr><tr><td colspan="2">Israel Soto <small>Digitally signed by Israel Soto Date: 2023.03.02 11:03:47 -0500</small></td><td>Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.03.07 09:13:55 -0500</small></td><td></td></tr><tr><td colspan="2">Tech Review / Date</td><td colspan="2">Admin Review / Date</td></tr></tbody></table>			Department Inspection		By	TDG	Barometric Pressure ID# <u>28663</u>				Gauge <u>1020</u> Instrument <u>1017</u>				Mouth Alcohol Solution Lot # <u>2021-D</u>				Acetone Stock Solution Lot # <u>2021-C</u>				Simulator	Serial Number			0.000	MP5095			Interferent	MP5097			0.050	MP5092			0.080	MP5093			0.200	MP5094			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td><input checked="" type="checkbox"/> Form 41</td><td><input type="checkbox"/> Post-Stability Checks</td></tr><tr><td><input checked="" type="checkbox"/> Stability Checks</td><td><input type="checkbox"/> Flow Calibration</td></tr><tr><td><input checked="" type="checkbox"/> Calibration Certificate</td><td><input checked="" type="checkbox"/> Form 40</td></tr><tr><td><input type="checkbox"/> Calibration Adjustment</td><td><input type="checkbox"/> Other _____</td></tr></tbody></table>		Attachments		<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks	<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration	<input checked="" type="checkbox"/> Calibration Certificate	<input checked="" type="checkbox"/> Form 40	<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____			<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC				<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC				<input type="checkbox"/> Return to/Place into Evidentiary Use				<input checked="" type="checkbox"/> Remain Out of Evidentiary Use				<input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use				Israel Soto <small>Digitally signed by Israel Soto Date: 2023.03.02 11:03:47 -0500</small>		Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.03.07 09:13:55 -0500</small>		Tech Review / Date		Admin Review / Date	
Calibration Adjustment				By																																																																																																																																																																			
Barometric Pressure Gauge _____ ID # _____																																																																																																																																																																							
Simulator	Serial #	Lot #	Expiration																																																																																																																																																																				
0.000		N/A	N/A																																																																																																																																																																				
0.040																																																																																																																																																																							
0.100																																																																																																																																																																							
0.200																																																																																																																																																																							
0.300																																																																																																																																																																							
0.080 DGS	N/A																																																																																																																																																																						
<input type="checkbox"/> Post Calibration Adjustment Stability Checks																																																																																																																																																																							
Simulator	Serial #	Lot #	Expiration																																																																																																																																																																				
0.050																																																																																																																																																																							
0.080																																																																																																																																																																							
0.200																																																																																																																																																																							
0.080 DGS	N/A																																																																																																																																																																						
Department Inspection		By	TDG																																																																																																																																																																				
Barometric Pressure ID# <u>28663</u>																																																																																																																																																																							
Gauge <u>1020</u> Instrument <u>1017</u>																																																																																																																																																																							
Mouth Alcohol Solution Lot # <u>2021-D</u>																																																																																																																																																																							
Acetone Stock Solution Lot # <u>2021-C</u>																																																																																																																																																																							
Simulator	Serial Number																																																																																																																																																																						
0.000	MP5095																																																																																																																																																																						
Interferent	MP5097																																																																																																																																																																						
0.050	MP5092																																																																																																																																																																						
0.080	MP5093																																																																																																																																																																						
0.200	MP5094																																																																																																																																																																						
<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td><input checked="" type="checkbox"/> Form 41</td><td><input type="checkbox"/> Post-Stability Checks</td></tr><tr><td><input checked="" type="checkbox"/> Stability Checks</td><td><input type="checkbox"/> Flow Calibration</td></tr><tr><td><input checked="" type="checkbox"/> Calibration Certificate</td><td><input checked="" type="checkbox"/> Form 40</td></tr><tr><td><input type="checkbox"/> Calibration Adjustment</td><td><input type="checkbox"/> Other _____</td></tr></tbody></table>		Attachments		<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks	<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration	<input checked="" type="checkbox"/> Calibration Certificate	<input checked="" type="checkbox"/> Form 40	<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____																																																																																																																																																												
Attachments																																																																																																																																																																							
<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks																																																																																																																																																																						
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration																																																																																																																																																																						
<input checked="" type="checkbox"/> Calibration Certificate	<input checked="" type="checkbox"/> Form 40																																																																																																																																																																						
<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____																																																																																																																																																																						
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC																																																																																																																																																																							
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC																																																																																																																																																																							
<input type="checkbox"/> Return to/Place into Evidentiary Use																																																																																																																																																																							
<input checked="" type="checkbox"/> Remain Out of Evidentiary Use																																																																																																																																																																							
<input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																																																																																																																																							
Israel Soto <small>Digitally signed by Israel Soto Date: 2023.03.02 11:03:47 -0500</small>		Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.03.07 09:13:55 -0500</small>																																																																																																																																																																					
Tech Review / Date		Admin Review / Date																																																																																																																																																																					

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 10:55

Date of Inspection: 03/01/2023

Serial Number: 80-001120
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT.

Not determined *MG*
3/1/2023

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

Signature and Printed Name TAYLOR D GUTSCHOW

03/01/2023
Date

FHP TROOP E MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001120
03/01/2023
Software: 8100.27

Test	g/210L	Time
Air Blānk	0.000	11:24
Control Test	0.049	11:25
Air Blank	0.000	11:26
Control Test	0.049	11:26
Air Blank	0.000	11:27
Control Test	0.049	11:28
Air Blank	0.000	11:28
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MG

Operator's Signature

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1120	FHP Troop E	03/01/2023	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																			
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																			
✓	✓	✓	✓																																																																																																			
<p><i>Inadvertently printed to the external printer. Results within range. Will attach.</i></p> <p><i>ML</i></p> <p><i>03/01/2023</i></p>	<p>FHP TROOP E MIAMI Intoxilyzer - Alcotest Analyzer Model 8000 03/01/2023 Software: 8100.27 SN 80-001120</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7370</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	11:32	Control Test	0.078	11:32	Air Blank	0.000	11:33	Control Test	0.079	11:34	Air Blank	0.000	11:34	Control Test	0.078	11:35	Air Blank	0.000	11:35	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel. Std Dev(%)	0.7370		<p>FHP TROOP E MIAMI Intoxilyzer - Alcotest Analyzer Model 8000 03/01/2023 Software: 8100.27 SN 80-001120</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr> <tr><td>Control Test</td><td>0.200</td><td>11:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:44</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1993</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2896</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	11:42	Control Test	0.200	11:43	Air Blank	0.000	11:43	Control Test	0.199	11:44	Air Blank	0.000	11:44	Control Test	0.199	11:45	Air Blank	0.000	11:46	Control Test Stats			Average	0.1993		Std Dev	0.0006		Rel. Std Dev(%)	0.2896		<p><i>065</i></p> <p>FHP TROOP E MIAMI Intoxilyzer - Alcotest Analyzer Model 8000 03/01/2023 Software: 8100.27 SN 80-001120</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:51</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7277</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	11:49	Control Test	0.079	11:49	Air Blank	0.000	11:50	Control Test	0.079	11:50	Air Blank	0.000	11:51	Control Test	0.080	11:51	Air Blank	0.000	11:52	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel. Std Dev(%)	0.7277	
Air Blank	0.000	11:32																																																																																																				
Control Test	0.078	11:32																																																																																																				
Air Blank	0.000	11:33																																																																																																				
Control Test	0.079	11:34																																																																																																				
Air Blank	0.000	11:34																																																																																																				
Control Test	0.078	11:35																																																																																																				
Air Blank	0.000	11:35																																																																																																				
Control Test Stats																																																																																																						
Average	0.0783																																																																																																					
Std Dev	0.0006																																																																																																					
Rel. Std Dev(%)	0.7370																																																																																																					
Air Blank	0.000	11:42																																																																																																				
Control Test	0.200	11:43																																																																																																				
Air Blank	0.000	11:43																																																																																																				
Control Test	0.199	11:44																																																																																																				
Air Blank	0.000	11:44																																																																																																				
Control Test	0.199	11:45																																																																																																				
Air Blank	0.000	11:46																																																																																																				
Control Test Stats																																																																																																						
Average	0.1993																																																																																																					
Std Dev	0.0006																																																																																																					
Rel. Std Dev(%)	0.2896																																																																																																					
Air Blank	0.000	11:49																																																																																																				
Control Test	0.079	11:49																																																																																																				
Air Blank	0.000	11:50																																																																																																				
Control Test	0.079	11:50																																																																																																				
Air Blank	0.000	11:51																																																																																																				
Control Test	0.080	11:51																																																																																																				
Air Blank	0.000	11:52																																																																																																				
Control Test Stats																																																																																																						
Average	0.0793																																																																																																					
Std Dev	0.0006																																																																																																					
Rel. Std Dev(%)	0.7277																																																																																																					

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 14:05

Date of Inspection: 03/01/2023

Serial Number: 80-001120
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.079	0.201	0.080
0.000	0.048	0.078	0.201	0.079
0.000	0.048	0.079	0.201	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080

Standard Deviations	0.0004	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/01/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001120, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001120</u>	UNCERTAINTY* \pm
Owning Agency:	<u>FHP TROOP E MIAMI</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/01/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:05</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


TAYLOR D GUTSCHOW,
Department Inspector

Date
03/01/2023

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality