



INSTRUMENT PROCESSING SHEET

Agency Santa Rosa County Sheriff's OfficeS/N 80-001303Florida Department of
Law EnforcementDate In 09-08-2023 DI Completion Date _____☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By ALL	Quality Checks	By IS	Date <u>09-11-2023</u>	Flow Calibration	By	Date																																								
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>slightly frayed breath hose</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>146</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202303K 03-29-2025</td></tr><tr><td>0.080</td><td>MP6292</td><td>202303L 03-29-2025</td></tr><tr><td>0.200</td><td>MP6293</td><td>202304C 04-05-2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5 04-05-2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202303K 03-29-2025	0.080	MP6292	202303L 03-29-2025	0.200	MP6293	202304C 04-05-2025	0.080 DGS	N/A	06723080A5 04-05-2025	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td><input type="checkbox"/> Battery Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Other _____</td><td></td><td></td></tr></tbody></table>			Maintenance		By	<input type="checkbox"/> Battery Replacement			<input type="checkbox"/> Dry Gas Regulator Replacement			<input type="checkbox"/> Breath Tube Replacement			<input type="checkbox"/> Other _____												
Simulator	Serial #	Lot #/Exp																																													
0.050	MP6291	202303K 03-29-2025																																													
0.080	MP6292	202303L 03-29-2025																																													
0.200	MP6293	202304C 04-05-2025																																													
0.080 DGS	N/A	06723080A5 04-05-2025																																													
Maintenance		By																																													
<input type="checkbox"/> Battery Replacement																																															
<input type="checkbox"/> Dry Gas Regulator Replacement																																															
<input type="checkbox"/> Breath Tube Replacement																																															
<input type="checkbox"/> Other _____																																															
Calibration Adjustment		By IS			Department Inspection																																										
Barometric Pressure Gauge <u>1009</u> ID # <u>28427</u>					Barometric Pressure ID# <u>30793</u> Gauge <u>1011</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>21410</td><td>09-30-2023</td></tr><tr><td>0.100</td><td>MP6296</td><td>22430</td><td>11-30-2024</td></tr><tr><td>0.200</td><td>MP6297</td><td>22400</td><td>10-12-2024</td></tr><tr><td>0.300</td><td>MP6298</td><td>23070</td><td>03-06-2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803</td><td>10-25-2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	21410	09-30-2023	0.100	MP6296	22430	11-30-2024	0.200	MP6297	22400	10-12-2024	0.300	MP6298	23070	03-06-2025	0.080 DGS	N/A	AG229803	10-25-2024				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td></td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	
Simulator	Serial #	Lot #	Expiration																																												
0.000	MP6294	N/A	N/A																																												
0.040	MP6295	21410	09-30-2023																																												
0.100	MP6296	22430	11-30-2024																																												
0.200	MP6297	22400	10-12-2024																																												
0.300	MP6298	23070	03-06-2025																																												
0.080 DGS	N/A	AG229803	10-25-2024																																												
Simulator	Serial Number																																														
0.000	MP6289																																														
Interferent	MP6290																																														
0.050	MP6291																																														
0.080	MP6292																																														
0.200																																															
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks					Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202303K</td><td>03-29-2025</td></tr><tr><td>0.080</td><td>MP6292</td><td>202303L</td><td>03-29-2025</td></tr><tr><td>0.200</td><td>MP6293</td><td>202304C</td><td>04-05-2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>04-05-2025</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202303K	03-29-2025	0.080	MP6292	202303L	03-29-2025	0.200	MP6293	202304C	04-05-2025	0.080 DGS	N/A	06723080A5	04-05-2025				<table border="1"><tbody><tr><td><input checked="" type="checkbox"/> Form 41</td><td><input checked="" type="checkbox"/> Post-Stability Checks</td></tr><tr><td><input checked="" type="checkbox"/> Stability Checks</td><td><input type="checkbox"/> Flow Calibration</td></tr><tr><td><input type="checkbox"/> Calibration Certificate</td><td><input type="checkbox"/> Form 40</td></tr><tr><td><input checked="" type="checkbox"/> Calibration Adjustment</td><td><input checked="" type="checkbox"/> Other <u>Form 51</u></td></tr></tbody></table>			<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Post-Stability Checks	<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40	<input checked="" type="checkbox"/> Calibration Adjustment	<input checked="" type="checkbox"/> Other <u>Form 51</u>												
Simulator	Serial #	Lot #	Expiration																																												
0.050	MP6291	202303K	03-29-2025																																												
0.080	MP6292	202303L	03-29-2025																																												
0.200	MP6293	202304C	04-05-2025																																												
0.080 DGS	N/A	06723080A5	04-05-2025																																												
<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Post-Stability Checks																																														
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration																																														
<input type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40																																														
<input checked="" type="checkbox"/> Calibration Adjustment	<input checked="" type="checkbox"/> Other <u>Form 51</u>																																														
Notes/Suggested Service: <u>Instrument had Ambient and Purge Fails during Stability Checks after area was cleared. Performed Optical Bench Calibration Adjustment. Instrument failed Department Inspection. Sending instrument to repair. IS</u>					<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																										
					Tech Review / Date _____ Admin Review / Date _____																																										

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SANTA ROSA COUNTY SO
Time of Inspection: 12:41

Date of Inspection: 09/15/2023

Serial Number: 80-001303
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.051	0.080 / 0.084		
0.000	0.051	0.081 / 0.084		
0.000	0.052	0.081 / 0.083		
0.000	0.051	0.083 / 0.083		
0.000	0.052	0.083 / 0.085		
0.000	0.050	0.085 / 0.085		
0.000	0.049	0.085 / 0.085		
0.000	0.048	0.087 / 0.086		
0.000	0.047	0.087 / 0.086		
0.000	0.047	0.086 / 0.085		

Standard Deviations	0.0019	0.0025 / 0.0010		
---------------------	--------	-----------------	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

Int Det:REMADE ACETONE SIMULATOR. 08: Control Outside Tolerance CHECKED SIM, Control Outside . Non-compliance:CONTROL OUTSIDE TOLERANCE.

Int. Remade acetone simulator, allowed to heat for 30 min prior to Detect. Repeat. Repeat test all Int Detect.

0.080 . Ensured 0.080g/210L simulator had no leaks and was g/210L connected properly. Repeat test still outside of tolerance

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

09/15/2023
Date

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	AMB*	12:59
Air Blank	0.000	12:59

*Ambient Fail

0.050, cleared area
with fan and
repeated

Operator's Signature

Stability Checks

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:01
Control Test	0.052	13:02
Air Blank	0.000	13:02
Control Test	0.048	13:03
Air Blank	0.000	13:04
Control Test	0.050	13:04
Air Blank	0.000	13:05
Control Test Stats		
Average	0.0500	
Std Dev	0.0020	
Rel Std Dev(%)	4.0000	

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:06
Control Test	0.079	13:07
Air Blank	0.000	13:07
Control Test	0.080	13:08
Air Blank	0.000	13:08
Control Test	0.079	13:09
Air Blank	0.000	13:10
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

wet

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	AMB*	13:11
Air Blank	PUR**	13:11

*Ambient Fail

**Purge Fail 0.200

Ensured instrument
area was clear,
repeated

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:12
Control Test	0.198	13:13
Air Blank	0.000	13:13
Control Test	0.196	13:14
Air Blank	0.000	13:15
Control Test	0.197	13:15
Air Blank	0.000	13:16
Control Test Stats		
Average	0.1970	
Std Dev	0.0010	
Rel Std Dev(%)	0.5076	

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:17
Control Test	0.082	13:17
Air Blank	0.000	13:18
Control Test	0.082	13:18
Air Blank	0.000	13:19
Control Test	0.080	13:19
Air Blank	0.000	13:20
Control Test Stats		
Average	0.0813	
Std Dev	0.0012	
Rel Std Dev(%)	1.4197	

Dry

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/15/2023 07:43:27

Optical Bench
Calibration Adjustment
✓

Auto Calibration

pg 1 of 2

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.2390	(0.2320)	0.2000	(0.1100)
Sample #2	0.1050	(0.2050)	0.1590	(0.0780)
Sample #3	0.0700	(0.2000)	0.1040	(0.1060)
Sample #4	0.0820	(-0.4030)	0.1250	(-0.2040)
Avg % Abs	0.0857	(0.0007)	0.1293	(-0.0067)
STD DEV	0.0178	(0.3496)	0.0278	(0.1715)
REL STD DEV	20.762	(52439.074)	21.460	(2572.023)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.6610	(-0.1090)	1.5030	(-0.0440)
Sample #2	0.8100	(-0.2450)	1.5050	(-0.1050)
Sample #3	0.7670	(-0.5480)	1.5110	(-0.2450)
Sample #4	0.8120	(-0.7760)	1.5460	(-0.3470)
Avg % Abs	0.7963	(-0.5230)	1.5207	(-0.2323)
STD DEV	0.0254	(0.2664)	0.0221	(0.1215)
REL STD DEV	3.193	(50.933)	1.456	(52.294)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	2.0050	(0.0980)	3.7340	(0.0320)
Sample #2	2.0030	(0.0950)	3.6510	(0.0550)
Sample #3	1.8600	(0.0800)	3.5870	(0.0670)
Sample #4	1.9240	(0.1140)	3.6440	(0.0640)
Avg % Abs	1.9290	(0.0963)	3.6273	(0.0620)
STD DEV	0.0716	(0.0170)	0.0351	(0.0062)
REL STD DEV	3.713	(17.688)	0.968	(10.073)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.8630	(-0.3270)	7.0430	(-0.1470)
Sample #2	3.9760	(-0.2100)	7.0720	(-0.0550)
Sample #3	3.7940	(-0.0330)	6.9770	(0.0400)
Sample #4	3.8510	(-0.0510)	6.9960	(0.0430)
Avg % Abs	3.8737	(-0.0980)	7.0150	(0.0093)
STD DEV	0.0931	(0.0974)	0.0503	(0.0557)
REL STD DEV	2.403	(99.399)	0.717	(597.155)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.4920	(0.3120)	10.1400	(0.1730)
Sample #2	5.4200	(0.2990)	10.0810	(0.1900)
Sample #3	5.4980	(0.3190)	10.1250	(0.2120)
Sample #4	5.5660	(0.2690)	10.1410	(0.1740)
Avg % Abs	5.4947	(0.2957)	10.1157	(0.1920)
STD DEV	0.0731	(0.0252)	0.0311	(0.0191)
REL STD DEV	1.330	(8.512)	0.307	(9.937)

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/15/2023 07:43:27

Optical Bench
Calibration Adjustment
✓

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>

Zero Order Coef -94.55
First Order Coef 2364.14
Second Order Coef 43.48

<<<<< 9um >>>>>

-132.11
1295.13
12.66

Act Fit Residual
(g/210L) (g/210L) (g/210L)
0.000 0.002 -0.0023
0.040 0.038 0.0019
0.100 0.097 0.0028
0.200 0.204 -0.0040
0.300 0.298 0.0016

Act Fit Residual
(g/210L) (g/210L) (g/210L)
0.000 0.001 -0.0007
0.040 0.039 0.0008
0.100 0.099 0.0006
0.200 0.201 -0.0011
0.300 0.300 0.0004

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 3392.00 3380.00
Sample #2 2988.00 3295.00
Sample #3 2871.00 3232.00
Sample #4 3026.00 3292.00
Avg 2961.6667 3273.0000
STD DEV 80.7857 35.5387
REL STD DEV 2.728 1.086
H2O adjust (mg/l*10k) 848 536

Barometric Pressure = 1009

*****CALIBRATION SUCCESSFUL*****

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3081.00
Sample #2 = 3052.00
Sample #3 = 2730.00
Sample #4 = 2992.00
Average Result = 2924.6667
STD DEV = 171.2347
REL STD DEV = 5.855

***** CHANNEL 2
Sample #1 = 3306.00
Sample #2 = 3309.00
Sample #3 = 3178.00
Sample #4 = 3279.00
Average Result = 3255.0000
STD DEV = 68.4617
REL STD DEV = 2.103

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3037.00
Sample #2 = 3039.00
Sample #3 = 3702.00
Sample #4 = 2887.00
Average Result = 3209.3333
STD DEV = 433.3778
REL STD DEV = 13.504

***** CHANNEL 2
Sample #1 = 3320.00
Sample #2 = 3307.00
Sample #3 = 3487.00
Sample #4 = 3260.00
Average Result = 3351.3333
STD DEV = 119.8179
REL STD DEV = 3.575

Dry 0.080g/210L test
was repeated. Tank was
running low, repeat
performed with full tank
✓

Post stability checks

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/15/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.050	09:45
Air Blank	0.000	09:46
Control Test	0.050	09:46
Air Blank	0.000	09:47
Control Test	0.050	09:47
Air Blank	0.000	09:48
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/15/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:49
Control Test	0.080	09:50
Air Blank	0.000	09:50
Control Test	0.081	09:51
Air Blank	0.000	09:51
Control Test	0.079	09:52
Air Blank	0.000	09:53
Control Test Stats		
Average	0.0800	
Std Dev	0.0010	
Rel Std Dev(%)	1.2500	

wet

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/15/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:55
Control Test	0.199	09:56
Air Blank	0.000	09:56
Control Test	0.198	09:57
Air Blank	0.000	09:58
Control Test	0.197	09:58
Air Blank	0.000	09:59
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001303
09/15/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:00
Control Test	0.079	10:00
Air Blank	0.000	10:00
Control Test	0.080	10:01
Air Blank	0.000	10:01
Control Test	0.082	10:02
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0803	
Std Dev	0.0015	
Rel Std Dev(%)	1.9015	

Dry

Operator's Signature

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Duron Nelson on 09-22-2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001303

Bill To Address:

Santa Rosa County Sheriff's Office

Florida

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Dept of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Instrument had Ambient and Purge Fails on initial testing. Optical Bench Calibration

Adjustment performed, 0.080 g/210L dry auto calibration had to be repeated twice due to

fault detect. Instrument ultimately passed calibration adjustment, but failed an Interferent

Detect and 0.080 wet g/210L test afterwards. Paperwork showing test results included.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Duron Nelson

Phone #: 850-981-2247

Email: dnelson@srsso.net

ATP Contact Name: Israel Soto

ATP Email: israelsoto@fdle.state.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Santa Rosa SO
Time of Inspection:

Date of Inspection:

Serial Number: 80-001303
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks: Compliance not determined for 2023, instrument at repair facility.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 Israel Soto

Signature and Printed Name

01-05-2024
Date