

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 1346	Sarasota CSO	02/01/2023	TDG <i>MG</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤0.003 of Wet
✓	✓	✓	✓
SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001346 02/01/2023 Software: 8100.27	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001346 02/01/2023 Software: 8100.27	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001346 02/01/2023 Software: 8100.27	DGS SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001346 02/01/2023 Software: 8100.27
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time
Air Blank 0.000 15:17	Air Blank 0.000 15:24	Air Blank 0.000 15:33	Air Blank 0.000 15:08
Control Test 0.049 15:18	Control Test 0.078 15:25	Control Test 0.198 15:34	Control Test 0.080 15:08
Air Blank 0.000 15:19	Air Blank 0.000 15:25	Air Blank 0.000 15:35	Air Blank 0.000 15:09
Control Test 0.048 15:19	Control Test 0.078 15:26	Control Test 0.198 15:35	Control Test 0.080 15:09
Air Blank 0.000 15:20	Air Blank 0.000 15:26	Air Blank 0.000 15:36	Air Blank 0.000 15:10
Control Test 0.048 15:20	Control Test 0.078 15:27	Control Test 0.198 15:37	Control Test 0.080 15:10
Air Blank 0.000 15:21	Air Blank 0.000 15:28	Air Blank 0.000 15:37	Air Blank 0.000 15:10
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0483	Average 0.0780	Average 0.1980	Average 0.0800
Std Dev 0.0006	Std Dev 0.0000	Std Dev 0.0000	Std Dev 0.0000
Rel Std Dev(%) 1.1945	Rel Std Dev(%) 0.0000	Rel Std Dev(%) 0.0000	Rel Std Dev(%) 0.0000
MG Operator's Signature	MG Operator's Signature	MG Operator's Signature	DGS Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 14:13

Date of Inspection: 02/07/2023

Serial Number: 80-001346
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG222203 Exp: 08/10/2024
0.000	0.049	0.074 / 0.075	0.199	0.080
0.000	0.049	0.075 / 0.075	0.200	0.081
0.000	0.049	0.074 / 0.075	0.199	0.080
0.000	0.050	0.074 / 0.075	0.199	0.080
0.000	0.050	0.075 / 0.075	0.200	0.080
0.000	0.049	0.074 / 0.075	0.199	0.080
0.000	0.050	0.075 / 0.076	0.199	0.081
0.000	0.049	0.075 / 0.076	0.199	0.080
0.000	0.050	0.075 / 0.076	0.199	0.080
0.000	0.050	0.075 / 0.076	0.199	0.081
Standard Deviations	0.0005	0.0005 / 0.0005	0.0004	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

checked sim seal and fitting to the instrument. Both were good. Repeated the 0.08 ARS test.

*ML
02/07/2023*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

02/07/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001346, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001346</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/07/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:13</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/07/2023

Date


TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

#1

ARS
(0.080)

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001346
02/08/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:15
Control Test	0.077	09:16
Air Blank	0.000	09:16
Control Test	0.077	09:17
Air Blank	0.000	09:17
Control Test	0.077	09:18
Air Blank	0.000	09:19
Control Test	0.077	09:19
Air Blank	0.000	09:20
Control Test	0.077	09:20
Air Blank	0.000	09:21
Control Test	0.077	09:22
Air Blank	0.000	09:22
Control Test	0.079	09:23
Air Blank	0.000	09:23
Control Test	0.079	09:24
Air Blank	0.000	09:25
Control Test	0.079	09:25
Air Blank	0.000	09:26
Control Test	0.078	09:26
Air Blank	0.000	09:27
Control Test	0.076	09:28
Air Blank	0.000	09:28
Control Test	0.076	09:29
Air Blank	0.000	09:29
Control Test	0.076	09:30
Air Blank	0.000	09:30
Control Test	0.076	09:31
Air Blank	0.000	09:32
Control Test	0.076	09:32
Air Blank	0.000	09:33
Control Test	0.076	09:34
Air Blank	0.000	09:34
Control Test	0.076	09:35
Air Blank	0.000	09:35
Control Test	0.076	09:36
Air Blank	0.000	09:36
Control Test	0.076	09:37
Air Blank	0.000	09:38
Control Test	0.077	09:38
Air Blank	0.000	09:39
Control Test Stats		
Average	0.0769	
Std Dev	0.0011	
Rel Std Dev(%)	1.3929	

80-001346
Extra Stabilities
02/08/2023

MB

#2

ARS
(0.080)

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001346
02/08/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:26
Control Test	0.077	11:26
Air Blank	0.000	11:27
Control Test	0.077	11:28
Air Blank	0.000	11:28
Control Test	0.077	11:29
Air Blank	0.000	11:29
Control Test	0.077	11:30
Air Blank	0.000	11:31
Control Test	0.077	11:31
Air Blank	0.000	11:32
Control Test	0.078	11:33
Air Blank	0.000	11:33
Control Test	0.077	11:34
Air Blank	0.000	11:34
Control Test	0.077	11:35
Air Blank	0.000	11:36
Control Test	0.077	11:36
Air Blank	0.000	11:37
Control Test	0.077	11:38
Air Blank	0.000	11:38
Control Test	0.077	11:39
Air Blank	0.000	11:39
Control Test	0.077	11:40
Air Blank	0.000	11:41
Control Test	0.077	11:41
Air Blank	0.000	11:42
Control Test	0.077	11:42
Air Blank	0.000	11:43
Control Test	0.077	11:44
Air Blank	0.000	11:44
Control Test	0.077	11:45
Air Blank	0.000	11:45
Control Test	0.077	11:46
Air Blank	0.000	11:47
Control Test	0.077	11:47
Air Blank	0.000	11:48
Control Test	0.077	11:49
Air Blank	0.000	11:49
Control Test	0.077	11:50
Air Blank	0.000	11:50
Control Test Stats		
Average	0.0771	
Std Dev	0.0002	
Rel Std Dev(%)	0.2902	

MB
Operator's Signature

MB
Operator's Signature

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001346
02/08/2023 09:40:59

Auto Calibration
Max Power Res Value = 57
Auto Range Res Value = 41

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12745, 9um Io = 13397

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1650 (-0.0090)
Sample #2 = 0.1600 (0.0120)
Sample #3 = 0.1540 (0.0180)
Sample #4 = 0.1660 (0.0380)
Avg % Abs = 0.1600 (0.0227)
STD DEV = 0.0060 (0.0136)
REL STD DEV = 3.750 (60.061)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1300 (0.0040)
Sample #2 = 0.1240 (0.0040)
Sample #3 = 0.1320 (0.0090)
Sample #4 = 0.1450 (0.0040)
Avg % Abs = 0.1337 (0.0057)
STD DEV = 0.0106 (0.0029)
REL STD DEV = 7.929 (50.943)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12735, 9um Io = 13390

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.9040 (-0.0090)
Sample #2 = 0.9020 (0.0080)
Sample #3 = 0.8900 (0.0280)
Sample #4 = 0.8980 (0.0280)
Avg % Abs = 0.8967 (0.0213)
STD DEV = 0.0061 (0.0115)
REL STD DEV = 0.681 (54.127)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.5150 (-0.0250)
Sample #2 = 1.5080 (-0.0030)
Sample #3 = 1.5130 (0.0000)
Sample #4 = 1.5090 (0.0070)
Avg % Abs = 1.5100 (0.0013)
STD DEV = 0.0026 (0.0051)
REL STD DEV = 0.175 (384.870)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12729, 9um Io = 13387

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.9930 (-0.0060)
Sample #2 = 1.9740 (0.0130)
Sample #3 = 1.9930 (0.0100)
Sample #4 = 1.9800 (0.0230)
Avg % Abs = 1.9823 (0.0153)
STD DEV = 0.0097 (0.0068)
REL STD DEV = 0.490 (44.393)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.5370 (-0.0170)
Sample #2 = 3.5280 (-0.0030)
Sample #3 = 3.5300 (-0.0050)
Sample #4 = 3.5230 (0.0090)
Avg % Abs = 3.5270 (0.0003)
STD DEV = 0.0036 (0.0076)
REL STD DEV = 0.102 (2271.564)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12723, 9um Io = 13384

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.7560 (-0.0080)
Sample #2 = 3.7410 (0.0150)
Sample #3 = 3.7440 (0.0280)
Sample #4 = 3.7680 (0.0280)
Avg % Abs = 3.7510 (0.0237)
STD DEV = 0.0148 (0.0075)
REL STD DEV = 0.395 (31.714)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 6.7310 (0.0000)
Sample #2 = 6.7280 (0.0200)
Sample #3 = 6.7190 (0.0270)
Sample #4 = 6.7270 (0.0270)
Avg % Abs = 6.7247 (0.0247)
STD DEV = 0.0049 (0.0040)
REL STD DEV = 0.073 (16.384)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12718, 9um Io = 13381

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 5.4490 (-0.0080)
Sample #2 = 5.4400 (0.0240)
Sample #3 = 5.4290 (0.0400)
Sample #4 = 5.4280 (0.0520)
Avg % Abs = 5.4323 (0.0387)
STD DEV = 0.0067 (0.0140)
REL STD DEV = 0.123 (136.330)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 9.7920 (-0.0100)
Sample #2 = 9.7330 (0.0420)
Sample #3 = 9.7260 (0.0560)
Sample #4 = 9.7230 (0.0520)
Avg % Abs = 9.7273 (0.0500)
STD DEV = 0.0051 (0.0072)
REL STD DEV = 0.053 (14.422)

Optical Calibration	
SN:	80-001346
Agency:	Sarasota CSO
Date:	02/08/2023
Quadratic Fit:	+/- 0.002g/210L ✓
By:	TDG <i>mg</i>

***** AUTO CAL DATA *****
***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.160
Std Dev = 0.01 Rel Std Dev = 3.75
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.897
Std Dev = 0.01 Rel Std Dev = 0.68
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.982
Std Dev = 0.01 Rel Std Dev = 0.49
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.751
Std Dev = 0.01 Rel Std Dev = 0.39
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.432
Std Dev = 0.01 Rel Std Dev = 0.12
Zero Order Coef = -397.86
First Order Coef = 2534.34
Second Order Coef = 30.82
Standard Deviation = 13.193066

***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.134
Std Dev = 0.01 Rel Std Dev = 7.93
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.510
Std Dev = 0.00 Rel Std Dev = 0.18
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.527
Std Dev = 0.00 Rel Std Dev = 0.10
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.725
Std Dev = 0.00 Rel Std Dev = 0.07
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.727
Std Dev = 0.01 Rel Std Dev = 0.05
Zero Order Coef = -172.65
First Order Coef = 1347.93
Second Order Coef = 14.19
Standard Deviation = 8.323061

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0001
0.100 0.100 0.0003
0.200 0.200 -0.0004
0.300 0.300 0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0002
0.100 0.100 0.0001
0.200 0.200 -0.0002
0.300 0.300 0.0001

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1 *****
Sample #1 = 2887.00
Sample #2 = 2782.00
Sample #3 = 2819.00
Sample #4 = 2794.00
Average Result = 2798.3333
STD DEV = 18.8768
REL STD DEV = 0.675

***** CHANNEL 2 *****
Sample #1 = 3350.00
Sample #2 = 3377.00
Sample #3 = 3373.00
Sample #4 = 3381.00
Average Result = 3377.0000
STD DEV = 4.0000
REL STD DEV = 0.118

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1025
3 um H2O Adjust (mg/l*10,000) = 1011
9 um H2O Adjust (mg/l*10,000) = 432
**** AUTO CAL PASS

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post-Cal)	80-00 1346	Sarasota CSO	02/08/2023	TDG MG

0.05g/210L			0.08g/210L			0.20g/210L			DGS 0.08g/210L			
0.047 to 0.053		✓	0.077 to 0.083		✓	0.194 to 0.206		✓	0.077 to 0.083	✓	≤0.003 of Wet	✓
SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001346 02/08/2023 Software: 8100.27			SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001346 02/08/2023 Software: 8100.27			SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001346 02/08/2023 Software: 8100.27			SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001346 02/08/2023 Software: 8100.27			
Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	
Air Blank	0.000	10:57	Air Blank	0.000	11:07	Air Blank	0.000	11:14	Air Blank	0.000	11:19	
Control Test	0.049	10:58	Control Test	0.078	11:07	Control Test	0.198	11:14	Control Test	0.080	11:19	
Air Blank	0.000	10:59	Air Blank	0.000	11:08	Air Blank	0.000	11:15	Air Blank	0.000	11:19	
Control Test	0.049	10:59	Control Test	0.078	11:09	Control Test	0.197	11:16	Control Test	0.080	11:20	
Air Blank	0.000	11:00	Air Blank	0.000	11:09	Air Blank	0.000	11:16	Air Blank	0.000	11:20	
Control Test	0.048	11:01	Control Test	0.077	11:10	Control Test	0.197	11:17	Control Test	0.080	11:20	
Air Blank	0.000	11:01	Air Blank	0.000	11:10	Air Blank	0.000	11:17	Air Blank	0.000	11:21	
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats			
Average	0.0487		Average	0.0777		Average	0.1973		Average	0.0800		
Std Dev	0.0006		Std Dev	0.0006		Std Dev	0.0006		Std Dev	0.0000		
Rel Std Dev(%)	1.1863		Rel Std Dev(%)	0.7434		Rel Std Dev(%)	0.2926		Rel Std Dev(%)	0.0000		
Operator's Signature			Operator's Signature			Operator's Signature			Operator's Signature			

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 13:54

Date of Inspection: 02/08/2023

Serial Number: 80-001346
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG222203 Exp: 08/10/2024
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.049	0.077	0.198	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.049	0.077	0.197	0.080
0.000	0.049	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080

Standard Deviations	0.0005	0.0004	0.0003	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/08/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001346, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001346</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/08/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:54</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/08/2023

Date


TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1