



Agency Clay County Sheriff's Office

S/N 80-001443

Date In 09-20-2023      DI Completion Date

☐ Ship   ☐ P/U   ☐ H/D   ☒ CMI   ☐ EE

FDLE/ATP Form 48 January 2022  
Issuing Authority: Alcohol Testing Program

Page 1 of 1

## Return Material Authorization

**Ship to:**



CMI, Inc.



Enforcement Electronics

Shipment to repair facility authorized by: Richard Patrone on 9/25/2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: \_\_\_\_\_

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001443

Bill To Address:

Richard Patrone

Clay County Sheriff's Office

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Department of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Instrument will not turn on.

**Please choose one of the following options:**

☐ 1. I \_\_\_\_\_, authorize all repairs.

☐ 2. I \_\_\_\_\_, authorize repairs up to \$\_\_\_\_\_.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Richard Patrone

Phone #: 904-529-5900

Email: rpatrone@claysheriff.com

ATP Contact Name: Benjamin Siddoway ATP Email: BenjaminSiddoway@fdle.state.fl.us



Agency Clay County Sheriff's Office

S/N 80-001443

Date In 02-28-2023 DI Completion Date 03-03-2023

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

## Calibration Adjustment

By\_

## ❑ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

## Department Inspection

By IS

## Attachments

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor  
Gutschow

Digitally signed by Taylor Gutschow  
Date: 2023.03.03 15:14:43 -05'00'

Phil Nicodemo

Digitally signed by Phil  
Nicodemo  
Date: 2023.03.07 10:15:12 -05'00'

Tech Review / Date

Admin Review / Date

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO

Time of Inspection: 15:03

Date of Inspection: 03/03/2023

Serial Number: 80-001443

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.047	0.077	0.200	0.078
0.000	0.048	0.078	0.200	0.078
0.000	0.048	0.077	0.200	0.078
0.000	0.048	0.077	0.200	0.078
0.000	0.048	0.078	0.200	0.079
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.078	0.201	0.078
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.079	0.200	0.078

Standard Deviations	0.0006	0.0007	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Israel Soto*

ISRAEL SOTO

Signature and Printed Name

03/03/2023  
Date

# Stability checks

CLAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001443  
03/02/2023  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:45
Control Test	0.048	11:46
Air Blank	0.000	11:47
Control Test	0.047	11:47
Air Blank	0.000	11:48
Control Test	0.047	11:48
Air Blank	0.000	11:49
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	

Operator's Signature

CLAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001443  
03/02/2023  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:51
Control Test	0.078	11:52
Air Blank	0.000	11:52
Control Test	0.079	11:53
Air Blank	0.000	11:54
Control Test	0.078	11:54
Air Blank	0.000	11:55
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Operator's Signature

CLAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001443  
03/02/2023  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:57
Control Test	0.200	11:58
Air Blank	0.000	11:58
Control Test	0.200	11:59
Air Blank	0.000	12:00
Control Test	0.200	12:00
Air Blank	0.000	12:01
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

CLAY COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001443  
03/02/2023  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:02
Control Test	0.078	12:02
Air Blank	0.000	12:03
Control Test	0.078	12:03
Air Blank	0.000	12:04
Control Test	0.078	12:04
Air Blank	0.000	12:04
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

wet

Dry





## Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road.  
Suite B1032  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001443, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001443</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>CLAY COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/03/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:03</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/03/2023

Date

Israel Soto

ISRAEL SOTO,  
Department Inspector

Digitally signed by Israel  
Soto  
Date: 2023.03.03 15:10:20  
-05'00'

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1