



INSTRUMENT PROCESSING SHEET

Agency Lakeland Police DepartmentS/N 80-001865Florida Department of
Law EnforcementDate In 08-08-2023 DI Completion Date 08-14-2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>ALL</u>	Quality Checks	By <u>BS</u>	Date <u>8/9/2023</u>	Flow Calibration	By _____	Date _____																																																														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>missing middle left foot</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>236</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																																
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5 4/5/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 3/29/2025	0.080	MP5089	202303L 3/29/2025	0.200	MP5090	202304C 4/5/2025	0.080 DGS	N/A	06723080A5 4/5/2025	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																																	
Simulator	Serial #	Lot #/Exp																																																																			
0.050	MP5088	202303K 3/29/2025																																																																			
0.080	MP5089	202303L 3/29/2025																																																																			
0.200	MP5090	202304C 4/5/2025																																																																			
0.080 DGS	N/A	06723080A5 4/5/2025																																																																			
Calibration Adjustment By _____				Department Inspection By <u>IS</u>																																																																	
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>30793</u> Gauge <u>1014</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table> Attachments <table border="1"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																																		
0.000		N/A	N/A																																																																		
0.040																																																																					
0.100																																																																					
0.200																																																																					
0.300																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial #	Lot #	Expiration																																																																		
0.050																																																																					
0.080																																																																					
0.200																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial Number																																																																				
0.000	MP6289																																																																				
Interferent	MP6290																																																																				
0.050	MP6291																																																																				
0.080	MP6292																																																																				
0.200	MP6293																																																																				
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																																																																				
Notes/Suggested Service: _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																																	
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2023.08.14 14:31:29 -0400				Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.08.18 13:17:16 -0400																																																																	
Tech Review / Date _____				Admin Review / Date _____																																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 12:40

Date of Inspection: 08/14/2023

Serial Number: 80-001865

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.049	0.080	0.198	0.080

Standard Deviations	0.0004	0.0005	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

08/14/2023

Date

Stability Checks

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001865
08/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:07
Control Test	0.050	11:07
Air Blank	0.000	11:08
Control Test	0.049	11:09
Air Blank	0.000	11:09
Control Test	0.049	11:10
Air Blank	0.000	11:10
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Benjamin Siddley
Operator's Signature

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001865
08/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:12
Control Test	0.078	11:13
Air Blank	0.000	11:14
Control Test	0.079	11:14
Air Blank	0.000	11:15
Control Test	0.078	11:16
Air Blank	0.000	11:16
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Benjamin Siddley
Operator's Signature

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001865
08/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:55
Control Test	0.200	10:56
Air Blank	0.000	10:56
Control Test	0.197	10:57
Air Blank	0.000	10:58
Control Test	0.197	10:58
Air Blank	0.000	10:59
Control Test Stats		
Average	0.1986	
Std Dev	0.0017	
Rel Std Dev(%)	0.8748	

Benjamin Siddley
Operator's Signature

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001865
08/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:02
Control Test	0.080	11:03
Air Blank	0.000	11:03
Control Test	0.079	11:04
Air Blank	0.000	11:04
Control Test	0.080	11:04
Air Blank	0.000	11:05
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

Benjamin Siddley
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001865, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001865</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LAKELAND PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/14/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:40</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2023.08.14
13:52:01 -04'00'

08/14/2023

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD
Time of Inspection: 09:35

Date of Inspection: 08/09/2023

Serial Number: 80-001865
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

N/A compliance not determined BS 8/9/2023

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin W Siddoway

BENJAMIN W SIDDOWAY
Signature and Printed Name

08/09/2023
Date