



INSTRUMENT PROCESSING SHEET

Agency Lakeland PDS/N 80-003945Florida Department of
Law EnforcementDate In 07/28/2023 DI Completion Date 08/08/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>08/07/2023</u>	Flow Calibration	By	Date																																								
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Missing back feet on DGS shelf</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>198</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024	Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																											
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5094	202201C 01/11/2024																																													
0.080	MP5095	202201D 01/18/2024																																													
0.200	MP5096	202201E 01/18/2024																																													
0.080 DGS	N/A	AG223802 08/26/2024																																													
Calibration Adjustment By _____		Department Inspection By <u>TDG</u>																																													
Barometric Pressure Gauge ID # _____		Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP5092																																														
Interferent	MP5093																																														
0.050	MP5094																																														
0.080	MP5095																																														
0.200	MP5096																																														
<input type="checkbox"/> Post Calibration Adjustment Stability Checks		Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																									
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____		Israel Soto <small>Digitally signed by Israel Soto Date: 2023.08.09 08:08:20 +0400</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.08.09 10:39:54 -0400</small> Tech Review / Date _____ Admin Review / Date _____																																													

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 13:44

Date of Inspection: 08/07/2023

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined *MG*
08/07/2023

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

08/07/2023
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 3945	Lakeland PD	08/07/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																			
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																			
✓	✓	✓	✓																																																																																																			
<div>Inadvertently printed to the external printer. Results within range. Will attach.</div> <div>MG 08/07/2023</div>	<div>LAKELAND PD Intoxilizer - Alcohol Analyzer Model 8000 08/07/2023 Software: 8100.27 SN 80-003945</div> <div>Test g/210L Time</div> <table><tr><td>Air Blank</td><td>0.000</td><td>13:21</td></tr><tr><td>Control Test</td><td>0.079</td><td>13:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:23</td></tr><tr><td>Control Test</td><td>0.076</td><td>13:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0793</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></table> <div>Operator's Signature MG</div>	Air Blank	0.000	13:21	Control Test	0.079	13:22	Air Blank	0.000	13:22	Control Test	0.078	13:23	Air Blank	0.000	13:23	Control Test	0.076	13:24	Air Blank	0.000	13:25	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7370		<div>LAKELAND PD Intoxilizer - Alcohol Analyzer Model 8000 08/07/2023 Software: 8100.27 SN 80-003945</div> <div>Test g/210L Time</div> <table><tr><td>Air Blank</td><td>0.000</td><td>13:26</td></tr><tr><td>Control Test</td><td>0.199</td><td>13:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:29</td></tr><tr><td>Control Test</td><td>0.199</td><td>13:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr><tr><td>Control Test</td><td>0.198</td><td>13:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1987</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr></table> <div>Operator's Signature MG</div>	Air Blank	0.000	13:26	Control Test	0.199	13:28	Air Blank	0.000	13:29	Control Test	0.199	13:30	Air Blank	0.000	13:30	Control Test	0.198	13:31	Air Blank	0.000	13:31	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		<div>LAKELAND PD Intoxilizer - Alcohol Analyzer Model 8000 08/07/2023 Software: 8100.27 SN 80-003945</div> <div>Test g/210L Time</div> <table><tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:34</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:34</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:35</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <div>Operator's Signature MG</div>	Air Blank	0.000	13:33	Control Test	0.080	13:33	Air Blank	0.000	13:33	Control Test	0.080	13:34	Air Blank	0.000	13:34	Control Test	0.080	13:35	Air Blank	0.000	13:35	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	13:21																																																																																																				
Control Test	0.079	13:22																																																																																																				
Air Blank	0.000	13:22																																																																																																				
Control Test	0.078	13:23																																																																																																				
Air Blank	0.000	13:23																																																																																																				
Control Test	0.076	13:24																																																																																																				
Air Blank	0.000	13:25																																																																																																				
Control Test Stats																																																																																																						
Average	0.0793																																																																																																					
Std Dev	0.0006																																																																																																					
Rel Std Dev(%)	0.7370																																																																																																					
Air Blank	0.000	13:26																																																																																																				
Control Test	0.199	13:28																																																																																																				
Air Blank	0.000	13:29																																																																																																				
Control Test	0.199	13:30																																																																																																				
Air Blank	0.000	13:30																																																																																																				
Control Test	0.198	13:31																																																																																																				
Air Blank	0.000	13:31																																																																																																				
Control Test Stats																																																																																																						
Average	0.1987																																																																																																					
Std Dev	0.0006																																																																																																					
Rel Std Dev(%)	0.2906																																																																																																					
Air Blank	0.000	13:33																																																																																																				
Control Test	0.080	13:33																																																																																																				
Air Blank	0.000	13:33																																																																																																				
Control Test	0.080	13:34																																																																																																				
Air Blank	0.000	13:34																																																																																																				
Control Test	0.080	13:35																																																																																																				
Air Blank	0.000	13:35																																																																																																				
Control Test Stats																																																																																																						
Average	0.0800																																																																																																					
Std Dev	0.0000																																																																																																					
Rel Std Dev(%)	0.0000																																																																																																					


Comments:

LAKELAND PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003945
08/07/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:14
Control Test	0.050	13:15
Air Blank	0.000	13:15
Control Test	0.050	13:16
Air Blank	0.000	13:17
Control Test	0.050	13:17
Air Blank	0.000	13:18
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

ML

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003945, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003945</u>	UNCERTAINTY* \pm
Owning Agency:	<u>LAKELAND PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/08/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:00</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/08/2023

Date



TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality