

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-005248	FWCC	08/10/2023	TDG MC

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
<p>FWCC Intoxilyzer - Alconol Analyzer Model 8000 08/10/2023 Software: 8100.27</p> <p>SN 80-005248</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:31</p> <p>Control Test 0.050 10:32</p> <p>Air Blank 0.000 10:32</p> <p>Control Test 0.049 10:33</p> <p>Air Blank 0.000 10:33</p> <p>Control Test 0.048 10:34</p> <p>Air Blank 0.000 10:35</p> <p>Control Test Stats</p> <p>Average 0.0490</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 2.0408</p>	<p>FWCC Intoxilyzer - Alconol Analyzer Model 8000 08/10/2023 Software: 8100.27</p> <p>SN 80-005248</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:37</p> <p>Control Test 0.079 10:38</p> <p>Air Blank 0.000 10:38</p> <p>Control Test 0.078 10:39</p> <p>Air Blank 0.000 10:40</p> <p>Control Test 0.078 10:40</p> <p>Air Blank 0.000 10:41</p> <p>Control Test Stats</p> <p>Average 0.0783</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7370</p>	<p>FWCC Intoxilyzer - Alconol Analyzer Model 8000 08/10/2023 Software: 8100.27</p> <p>SN 80-005248</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:44</p> <p>Control Test 0.204 10:44</p> <p>Air Blank 0.000 10:45</p> <p>Control Test 0.202 10:46</p> <p>Air Blank 0.000 10:46</p> <p>Control Test 0.201 10:47</p> <p>Air Blank 0.000 10:48</p> <p>Control Test Stats</p> <p>Average 0.2023</p> <p>Std Dev 0.0015</p> <p>Rel Std Dev(%) 0.7550</p>	<p>FWCC Intoxilyzer - Alconol Analyzer Model 8000 08/10/2023 Software: 8100.27</p> <p>SN 80-005248</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:48</p> <p>Control Test 0.077 10:49</p> <p>Air Blank 0.000 10:49</p> <p>Control Test 0.078 10:50</p> <p>Air Blank 0.000 10:50</p> <p>Control Test 0.078 10:50</p> <p>Air Blank 0.000 10:51</p> <p>Control Test Stats</p> <p>Average 0.0777</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7434</p>
<p>MC</p> <p>Operator's Signature</p>	<p>MC</p> <p>Operator's Signature</p>	<p>MC</p> <p>Operator's Signature</p>	<p>MC</p> <p>Operator's Signature</p>

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 12:56

Date of Inspection: 08/10/2023

Serial Number: 80-005248

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.080	0.204	0.076
0.000	0.049	0.079	0.203	0.077
0.000	0.049	0.079	0.203	0.077
0.000	0.049	0.079	0.203	0.076
0.000	0.048	0.079	0.203	0.076
0.000	0.049	0.079	0.203	0.076
0.000	0.049	0.079	0.203	0.076
0.000	0.049	0.079	0.204	0.077
0.000	0.049	0.079	0.203	0.076
0.000	0.049	0.079	0.203	0.076


Standard Deviations	0.0003	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

08/10/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005248, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005248</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/10/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:56</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/10/2023

Date



TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency FFWCCS/N 80-005248Florida Department of
Law EnforcementDate In 03/01/2023DI Completion Date 04/26/2023☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>04/24/2023</u>	Flow Calibration	By TDG	Date <u>04/24/2023</u>																																							
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Dropped off with DGS</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>99*</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.132*</u> (.139 - .169) 36 mm <u>0.152*</u> (.156 - .190) 53 mm <u>0.218*</u> (.228 - .278) 103 mm <u>0.392</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # <u>ATP106 (x2)</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>100 / 100</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP104 (x2)</u> 32 mm <u>0.132* / 0.140</u> (.139 - .169) 36 mm <u>0.152* / 0.152*</u> (.156 - .190) 53 mm <u>0.222* / 0.226*</u> (.228 - .278) 103 mm <u>0.496 / 0.492</u> (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP4864</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP4864	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP6286	202201C 01/11/2024																																												
0.080	MP4864	202201D 01/18/2024																																												
0.200	MP6288	202201E 01/18/2024																																												
0.080 DGS	N/A	AG223802 08/26/2024																																												
Calibration Adjustment		By _____		Department Inspection			By TDG _____																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP4864	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP6284																																													
Interferent	MP6285																																													
0.050	MP6286																																													
0.080	MP4864																																													
0.200	MP6288																																													
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Calibration (x2) <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Other <u>Form 51</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																						
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Notes/Suggested Service: <u>Checked breath tube screen and replaced o-rings on 3/7. Agency picked up power cord and DGS around this time. Attempted to replace battery on 4/14 but was unsuccessful. (TDG)</u>				Israel Soto <small>Digitally signed by Israel Soto Date: 2023.04.26 18:27:27 +04'00'</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.05.03 10:57:41 -04'00'</small>																																										
*R-value below 100. Flow values below nominal. Flow calibrations were unsuccessful; post-cal verifications failed. Will send to repair. (TDG)				Tech Review / Date _____ Admin Review / Date _____																																										

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FISH & WILDLIFE CC
Time of Inspection: 11:20

Date of Inspection: 04/24/2023

Serial Number: 80-005248
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT.

Not determined mg
04/24/2023

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

04/24/2023
Date

FISH & WILDLIFE CC

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-005248

04/24/2023

Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 5.566

2: Rate (Liters/min) = 15

SQRT(Diff)) = 10.391

3: Rate (Liters/min) = 30

SQRT(Diff)) = 20.516

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 642

Rounded Intercept = -330820

Correlation = 0.99622

Flow Calibration #1
MB 04/24/2023

FISH & WILDLIFE CC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005248
04/24/2023
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff)) = 5.477
2: Rate (Liters/min) = 15
SQRT(Diff)) = 10.098
3: Rate (Liters/min) = 30
SQRT(Diff)) = 20.441

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 638

Rounded Intercept = -294436

Correlation = 0.99477

Flow Calibration #2
MG 04/24/2023

FISH & WILDLIFE CC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005248
04/24/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:25
Control Test	0.050	12:25
Air Blank	0.000	12:26
Control Test	0.049	12:27
Air Blank	0.000	12:27
Control Test	0.049	12:28
Air Blank	0.000	12:29
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	



Operator's Signature

FISH & WILDLIFE CC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005248
04/24/2023
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.080	12:33
Air Blank	0.000	12:33
Control Test	0.079	12:34
Air Blank	0.000	12:34
Control Test	0.078	12:35
Air Blank	0.000	12:36
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

ML

Operator's Signature

FISH & WILDLIFE CC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005248
04/24/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:39
Control Test	0.205	12:40
Air Blank	0.000	12:40
Control Test	0.203	12:41
Air Blank	0.000	12:42
Control Test	0.203	12:42
Air Blank	0.000	12:43
Control Test Stats		
Average	0.2037	
Std Dev	0.0012	
Rel Std Dev(%)	0.5670	



Operator's Signature

FISH & WILDLIFE CC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005248
04/24/2023
Software: 8100.27

DGS

Test	g/210L	Time
Air Blank	0.000	12:16
Control Test	0.081	12:16
Air Blank	0.000	12:17
Control Test	0.081	12:17
Air Blank	0.000	12:17
Control Test	0.081	12:18
Air Blank	0.000	12:18
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

ML

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FISH & WILDLIFE CC
Time of Inspection: 12:55

Date of Inspection: 04/26/2023

Serial Number: 80-005248
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.078	0.204	0.080
0.000	0.049	0.079	0.204	0.081
0.000	0.049	0.078	0.203	0.080
0.000	0.049	0.078	0.203	0.080
0.000	0.049	0.078	0.203	0.080
0.000	0.049	0.078	0.203	0.080
0.000	0.049	0.078	0.203	0.080
0.000	0.049	0.078	0.203	0.080
0.000	0.050	0.078	0.203	0.080
0.000	0.050	0.078	0.203	0.080

Standard Deviations	0.0004	0.0003	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/26/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005248, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005248</u>	UNCERTAINTY* \pm
Owning Agency:	<u>FISH & WILDLIFE CC</u>	0.050 g/210 L 0.004
Calibration Date:	<u>04/26/2023</u>	0.080 g/210 L 0.004
Calibration Time:	<u>12:55</u>	0.200 g/210 L 0.007
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/26/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Captain Thomas Shipp on 4/14/2023

Items Returned: Instrument ☐ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005248

Bill To Address:

Florida Fish and Wildlife Conservation

Attn: Captain Shipp

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Instrument requires a battery change. The battery case is missing the divot used to remove the battery from its case. R-value is 100. Flow values are below optimal and could not be corrected with repeated flow calibrations.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Capt. Thomas Shipp

Phone #: 850-251-7278 Email: Thomas.Shipp@myfwc.com

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us