



INSTRUMENT PROCESSING SHEET

Agency Miami-Dade PDS/N 80-005551Florida Department of
Law EnforcementDate In 10/09/2023 DI Completion Date 10/10/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 10/10/2023	Flow Calibration	By	Date																																								
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>119</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024	Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____ _____																											
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5094	202201C 01/11/2024																																													
0.080	MP5095	202201D 01/18/2024																																													
0.200	MP5096	202201E 01/18/2024																																													
0.080 DGS	N/A	AG223802 08/26/2024																																													
Calibration Adjustment By TDG _____		Department Inspection By TDG _____																																													
Barometric Pressure Gauge <u>1014</u> ID # <u>28199</u>		Barometric Pressure ID# <u>26932</u> Gauge <u>1013</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>22460</td><td>12/28/2024</td></tr><tr><td>0.100</td><td>MP5099</td><td>22310</td><td>08/11/2024</td></tr><tr><td>0.200</td><td>MP5100</td><td>22050</td><td>02/07/2024</td></tr><tr><td>0.300</td><td>MP5101</td><td>22220</td><td>06/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203</td><td>08/10/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	22460	12/28/2024	0.100	MP5099	22310	08/11/2024	0.200	MP5100	22050	02/07/2024	0.300	MP5101	22220	06/15/2024	0.080 DGS	N/A	AG222203	08/10/2024	<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP4863</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP4863	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																												
0.000	MP5097	N/A	N/A																																												
0.040	MP5098	22460	12/28/2024																																												
0.100	MP5099	22310	08/11/2024																																												
0.200	MP5100	22050	02/07/2024																																												
0.300	MP5101	22220	06/15/2024																																												
0.080 DGS	N/A	AG222203	08/10/2024																																												
Simulator	Serial Number																																														
0.000	MP4863																																														
Interferent	MP5093																																														
0.050	MP5094																																														
0.080	MP5095																																														
0.200	MP5096																																														
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks		Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C</td><td>01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D</td><td>01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E</td><td>01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802</td><td>08/26/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5094	202201C	01/11/2024	0.080	MP5095	202201D	01/18/2024	0.200	MP5096	202201E	01/18/2024	0.080 DGS	N/A	AG223802	08/26/2024	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																									
Simulator	Serial #	Lot #	Expiration																																												
0.050	MP5094	202201C	01/11/2024																																												
0.080	MP5095	202201D	01/18/2024																																												
0.200	MP5096	202201E	01/18/2024																																												
0.080 DGS	N/A	AG223802	08/26/2024																																												
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		Benjamin Siddoway <small>Digitally signed by Benjamin Siddoway Date: 2023.10.10 15:25:02 -0400</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.10.10 15:30:32 -0400</small> Tech Review / Date _____ Admin Review / Date _____																																													

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>MIAMI DAGE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-005551 10/10/2023 Software: 8100.27</p> <p>0.047 to 0.053</p> <p>✓</p>	<p>MIAMI DAGE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-005551 10/10/2023 Software: 8100.27</p> <p>0.077 to 0.083</p> <p>✗</p>	<p>MIAMI DAGE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-005551 10/10/2023 Software: 8100.27</p> <p>0.194 to 0.206</p> <p>✓</p>	<p>MIAMI DAGE PD Intoxilyzer - Alconol Analyzer Model 8000 SN 80-005551 10/10/2023 Software: 8100.27</p> <p>0.077 to 0.083</p> <p>✓</p> <p>≤0.003 of Wet</p> <p>✗</p>
<p>Test g/210L Time</p> <p>Air Blank 0.000 09:45</p> <p>Control Test 0.049 09:45</p> <p>Air Blank 0.000 09:46</p> <p>Control Test 0.049 09:47</p> <p>Air Blank 0.000 09:47</p> <p>Control Test 0.048 09:48</p> <p>Air Blank 0.000 09:49</p> <p>Control Test Stats</p> <p>Average 0.0487</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1863</p>	<p>Test g/210L Time</p> <p>Air Blank 0.000 09:52</p> <p>Control Test 0.076 09:52</p> <p>Air Blank 0.000 09:53</p> <p>Control Test 0.076 09:54</p> <p>Air Blank 0.000 09:54</p> <p>Control Test 0.076 09:55</p> <p>Air Blank 0.000 09:56</p> <p>Control Test Stats</p> <p>Average 0.0760</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>Test g/210L Time</p> <p>Air Blank 0.000 10:00</p> <p>Control Test 0.199 10:00</p> <p>Air Blank 0.000 10:01</p> <p>Control Test 0.198 10:02</p> <p>Air Blank 0.000 10:02</p> <p>Control Test 0.199 10:03</p> <p>Air Blank 0.000 10:04</p> <p>Control Test Stats</p> <p>Average 0.1987</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2906</p>	<p>Test g/210L Time</p> <p>Air Blank 0.000 09:39</p> <p>Control Test 0.082 09:39</p> <p>Air Blank 0.000 09:39</p> <p>Control Test 0.080 09:40</p> <p>Air Blank 0.000 09:40</p> <p>Control Test 0.081 09:41</p> <p>Air Blank 0.000 09:41</p> <p>Control Test Stats</p> <p>Average 0.0810</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 1.2346</p>
<p>Operator's Signature</p> <p><i>MC</i></p>	<p>Operator's Signature</p> <p><i>MC</i></p>	<p>Operator's Signature</p> <p><i>MC</i></p>	<p>Operator's Signature</p> <p><i>MC</i></p>

MIAMI DQDE PD
Intoxilyzer - Alcohol Analyzer
Model 8000
10/10/2023
SN 80-005551
10:04:46

Auto Calibration

Max Power Res Value = 83
Auto Range Res Value = 67

Sol Value = 0.000 g/210L ***

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sum Io = 12623, Sum Io = 13437

Sample % Abs (% Abs Ref)
Sample #1 = 1.5580 (0.0000)
Sample #2 = 1.5590 (0.0200)
Sample #3 = 1.5570 (0.0220)
Sample #4 = 1.5380 (0.0260)
Avg % Abs = 1.5513 (0.0227)
STD DEV = 0.0116 (0.0031)
REL STD DEV = 0.747 (13.478)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12595, Sum Io = 13422

Sample % Abs (% Abs Ref)
Sample #1 = 1.9040 (-0.0070)
Sample #2 = 1.8510 (0.0220)
Sample #3 = 1.8630 (0.0340)
Sample #4 = 1.8300 (0.0750)
Avg % Abs = 1.8480 (0.0437)
STD DEV = 0.0167 (0.0278)
REL STD DEV = 0.904 (63.643)

Sample % Abs (% Abs Ref)
Sample #1 = 5.3230 (-0.0150)
Sample #2 = 5.1880 (0.1440)
Sample #3 = 5.1790 (0.1450)
Sample #4 = 5.1390 (0.1660)
Avg % Abs = 5.1687 (0.1383)
STD DEV = 0.0261 (0.0315)
REL STD DEV = 0.505 (22.795)

Sample % Abs (% Abs Ref)
Sample #1 = 9.9220 (-0.0020)
Sample #2 = 9.6790 (0.2190)
Sample #3 = 9.6420 (0.2780)
Sample #4 = 9.6150 (0.2980)
Avg % Abs = 9.6453 (0.2650)
STD DEV = 0.0321 (0.0411)
REL STD DEV = 0.333 (15.499)

Sample % Abs (% Abs Ref)
Sample #1 = 3.6010 (-0.0100)
Sample #2 = 3.5210 (0.0620)
Sample #3 = 3.5440 (0.0590)
Sample #4 = 3.5130 (0.0870)
Avg % Abs = 3.5260 (0.0693)
STD DEV = 0.0161 (0.0154)
REL STD DEV = 0.456 (22.173)

Sample % Abs (% Abs Ref)
Sample #1 = 3.6450 (-0.0160)
Sample #2 = 3.5300 (0.0830)
Sample #3 = 3.5140 (0.1120)
Sample #4 = 3.5390 (0.1060)
Avg % Abs = 3.5277 (0.1003)
STD DEV = 0.0127 (0.0153)
REL STD DEV = 0.359 (15.257)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12590, Sum Io = 13419

Sample % Abs (% Abs Ref)
Sample #1 = 1.9040 (-0.0070)
Sample #2 = 1.8510 (0.0220)
Sample #3 = 1.8630 (0.0340)
Sample #4 = 1.8300 (0.0750)
Avg % Abs = 1.8480 (0.0437)
STD DEV = 0.0167 (0.0278)
REL STD DEV = 0.904 (63.643)

Sample % Abs (% Abs Ref)
Sample #1 = 5.3230 (-0.0150)
Sample #2 = 5.1880 (0.1440)
Sample #3 = 5.1790 (0.1450)
Sample #4 = 5.1390 (0.1660)
Avg % Abs = 5.1687 (0.1383)
STD DEV = 0.0261 (0.0315)
REL STD DEV = 0.505 (22.795)

Sample % Abs (% Abs Ref)
Sample #1 = 9.9220 (-0.0020)
Sample #2 = 9.6790 (0.2190)
Sample #3 = 9.6420 (0.2780)
Sample #4 = 9.6150 (0.2980)
Avg % Abs = 9.6453 (0.2650)
STD DEV = 0.0321 (0.0411)
REL STD DEV = 0.333 (15.499)

Sample % Abs (% Abs Ref)
Sample #1 = 6.8510 (-0.0060)
Sample #2 = 6.6680 (0.1550)
Sample #3 = 6.6190 (0.1990)
Sample #4 = 6.6550 (0.1860)
Avg % Abs = 6.6473 (0.1800)
STD DEV = 0.0254 (0.0226)
REL STD DEV = 0.382 (12.559)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12584, Sum Io = 13416

Sample % Abs (% Abs Ref)
Sample #1 = 5.3230 (-0.0150)
Sample #2 = 5.1880 (0.1440)
Sample #3 = 5.1790 (0.1450)
Sample #4 = 5.1390 (0.1660)
Avg % Abs = 5.1687 (0.1383)
STD DEV = 0.0261 (0.0315)
REL STD DEV = 0.505 (22.795)

Sample % Abs (% Abs Ref)
Sample #1 = 9.9220 (-0.0020)
Sample #2 = 9.6790 (0.2190)
Sample #3 = 9.6420 (0.2780)
Sample #4 = 9.6150 (0.2980)
Avg % Abs = 9.6453 (0.2650)
STD DEV = 0.0321 (0.0411)
REL STD DEV = 0.333 (15.499)

Sample % Abs (% Abs Ref)
Sample #1 = 3.6010 (-0.0100)
Sample #2 = 3.5210 (0.0620)
Sample #3 = 3.5440 (0.0590)
Sample #4 = 3.5130 (0.0870)
Avg % Abs = 3.5260 (0.0693)
STD DEV = 0.0161 (0.0154)
REL STD DEV = 0.456 (22.173)

Sample % Abs (% Abs Ref)
Sample #1 = 3.6450 (-0.0160)
Sample #2 = 3.5300 (0.0830)
Sample #3 = 3.5140 (0.1120)
Sample #4 = 3.5390 (0.1060)
Avg % Abs = 3.5277 (0.1003)
STD DEV = 0.0127 (0.0153)
REL STD DEV = 0.359 (15.257)

Sol Value = 0.400 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12590, Sum Io = 13419

Sample % Abs (% Abs Ref)
Sample #1 = 1.9040 (-0.0070)
Sample #2 = 1.8510 (0.0220)
Sample #3 = 1.8630 (0.0340)
Sample #4 = 1.8300 (0.0750)
Avg % Abs = 1.8480 (0.0437)
STD DEV = 0.0167 (0.0278)
REL STD DEV = 0.904 (63.643)

Sample % Abs (% Abs Ref)
Sample #1 = 5.3230 (-0.0150)
Sample #2 = 5.1880 (0.1440)
Sample #3 = 5.1790 (0.1450)
Sample #4 = 5.1390 (0.1660)
Avg % Abs = 5.1687 (0.1383)
STD DEV = 0.0261 (0.0315)
REL STD DEV = 0.505 (22.795)

Sample % Abs (% Abs Ref)
Sample #1 = 9.9220 (-0.0020)
Sample #2 = 9.6790 (0.2190)
Sample #3 = 9.6420 (0.2780)
Sample #4 = 9.6150 (0.2980)
Avg % Abs = 9.6453 (0.2650)
STD DEV = 0.0321 (0.0411)
REL STD DEV = 0.333 (15.499)

Sample % Abs (% Abs Ref)
Sample #1 = 3.6010 (-0.0100)
Sample #2 = 3.5210 (0.0620)
Sample #3 = 3.5440 (0.0590)
Sample #4 = 3.5130 (0.0870)
Avg % Abs = 3.5260 (0.0693)
STD DEV = 0.0161 (0.0154)
REL STD DEV = 0.456 (22.173)

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0004
0.040 0.040 -0.0004
0.100 0.100 -0.0002
0.200 0.199 0.0005
0.300 0.300 -0.0002

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

Sample #1 = 3285.00
Sample #2 = 3246.00
Sample #3 = 3194.00
Sample #4 = 3281.00
Average Result = 3241.0000
STD DEV = 43.9204
REL STD DEV = 1.355

Sample #1 = 3329.00
Sample #2 = 3324.00
Sample #3 = 3317.00
Sample #4 = 3359.00
Average Result = 3333.3333
STD DEV = 22.5019
REL STD DEV = 0.675

Sample #1 = 3329.00
Sample #2 = 3324.00
Sample #3 = 3317.00
Sample #4 = 3359.00
Average Result = 3333.3333
STD DEV = 22.5019
REL STD DEV = 0.675

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l*10,000) = 568
9 um H2O Adjust (mg/l*10,000) = 476

***** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.001 0.0006
0.040 0.041 -0.0009
0.100 0.100 0.0000
0.200 0.200 0.0005
0.300 0.300 -0.0002

Sol Value = 0.000 mg/l or 0.000 g/210L
% Abs = 0.087
Std Dev = 0.02 Rel Std Dev = 24.92
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.821
Std Dev = 0.01 Rel Std Dev = 1.32
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.848
Std Dev = 0.02 Rel Std Dev = 0.90
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.528
Std Dev = 0.01 Rel Std Dev = 0.36
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.169
Std Dev = 0.03 Rel Std Dev = 0.50
Zero Order Coef = -258.51
First Order Coef = 2660.88
Second Order Coef = 29.96
Standard Deviation = 27.820765

Sol Value = 0.000 mg/l or 0.000 g/210L
% Abs = 0.161
Std Dev = 0.03 Rel Std Dev = 17.31
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.551
Std Dev = 0.01 Rel Std Dev = 0.75
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.526
Std Dev = 0.02 Rel Std Dev = 0.46
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.647
Std Dev = 0.03 Rel Std Dev = 0.38
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.645
Std Dev = 0.03 Rel Std Dev = 0.33
Zero Order Coef = -238.09
First Order Coef = 1372.25
Second Order Coef = 13.94
Standard Deviation = 19.135307

Sol Value = 0.000 mg/l or 0.000 g/210L
% Abs = 0.087
Std Dev = 0.02 Rel Std Dev = 24.92
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.821
Std Dev = 0.01 Rel Std Dev = 1.32
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.848
Std Dev = 0.02 Rel Std Dev = 0.90
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.528
Std Dev = 0.01 Rel Std Dev = 0.36
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.169
Std Dev = 0.03 Rel Std Dev = 0.50
Zero Order Coef = -258.51
First Order Coef = 2660.88
Second Order Coef = 29.96
Standard Deviation = 27.820765

Optical Calibration Adjustment

By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
<p>0.047 to 0.053</p> <p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-005551</p>	<p>0.077 to 0.083</p> <p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-005551</p>	<p>0.194 to 0.206</p> <p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-005551</p>	<p>0.077 to 0.083</p> <p>MIAMI DAGE PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-005551</p>
<p>Test</p> <p>g/210L</p> <p>Time</p> <p>Air Blank 0.000 11:20</p> <p>Control Test 0.049 11:21</p> <p>Air Blank 0.000 11:22</p> <p>Control Test 0.049 11:22</p> <p>Air Blank 0.000 11:23</p> <p>Control Test 0.049 11:24</p> <p>Air Blank 0.000 11:24</p> <p>Control Test Stats</p> <p>Average 0.0490</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>Test</p> <p>g/210L</p> <p>Time</p> <p>Air Blank 0.000 11:29</p> <p>Control Test 0.079 11:30</p> <p>Air Blank 0.000 11:31</p> <p>Control Test 0.079 11:31</p> <p>Air Blank 0.000 11:32</p> <p>Control Test 0.078 11:33</p> <p>Air Blank 0.000 11:33</p> <p>Control Test Stats</p> <p>Average 0.0787</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7339</p>	<p>Test</p> <p>g/210L</p> <p>Time</p> <p>Air Blank 0.000 11:35</p> <p>Control Test 0.203 11:37</p> <p>Air Blank 0.000 11:38</p> <p>Control Test 0.202 11:38</p> <p>Air Blank 0.000 11:39</p> <p>Control Test 0.201 11:40</p> <p>Air Blank 0.000 11:41</p> <p>Control Test Stats</p> <p>Average 0.2020</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 0.4950</p>	<p>Test</p> <p>g/210L</p> <p>Time</p> <p>Air Blank 0.000 11:15</p> <p>Control Test 0.079 11:16</p> <p>Air Blank 0.000 11:16</p> <p>Control Test 0.079 11:17</p> <p>Air Blank 0.000 11:17</p> <p>Control Test 0.079 11:17</p> <p>Air Blank 0.000 11:18</p> <p>Control Test Stats</p> <p>Average 0.0790</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>
<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD
Time of Inspection: 14:13

Date of Inspection: 10/10/2023

Serial Number: 80-005551
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.078	0.203	0.079
0.000	0.049	0.078	0.203	0.078
0.000	0.049	0.079	0.203	0.079
0.000	0.049	0.079	0.203	0.079
0.000	0.049	0.079	0.204	0.079
0.000	0.049	0.079	0.204	0.078
0.000	0.049	0.079	0.204	0.079
0.000	0.048	0.078	0.204	0.078
0.000	0.049	0.079	0.204	0.077
0.000	0.050	0.079	0.204	0.078

Standard Deviations	0.0004	0.0004	0.0005	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/10/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005551, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005551</u>	UNCERTAINTY* \pm
Owning Agency:	<u>MIAMI DADE PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/10/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:13</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



10/10/2023

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE PD
Time of Inspection: 08:29

Date of Inspection: 01/12/2023

Serial Number: 80-005551
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.079	0.202	0.080
0.000	0.049	0.079	0.202	0.080
0.000	0.049	0.080	0.202	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.202	0.079
0.000	0.049	0.079	0.202	0.080
0.000	0.050	0.080	0.202	0.079
0.000	0.050	0.079	0.202	0.079
0.000	0.049	0.079	0.202	0.080
0.000	0.049	0.080	0.202	0.080


Standard Deviations	0.0004	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

DAVID E REYES-RIVERA

01/12/2023
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-005551	Miami-Dade Police Department	01/12/2023	DERR <i>[Signature]</i>

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005551 01/12/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:45</td></tr> <tr><td>Control Test</td><td>0.050</td><td>06:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:46</td></tr> <tr><td>Control Test</td><td>0.050</td><td>06:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:47</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:45	Control Test	0.050	06:45	Air Blank	0.000	06:46	Control Test	0.050	06:47	Air Blank	0.000	06:47	Control Test	0.049	06:48	Air Blank	0.000	06:48	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005551 01/12/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:49</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:50</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:53</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:49	Control Test	0.080	06:50	Air Blank	0.000	06:50	Control Test	0.080	06:51	Air Blank	0.000	06:52	Control Test	0.081	06:52	Air Blank	0.000	06:53	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005551 01/12/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:54</td></tr> <tr><td>Control Test</td><td>0.202</td><td>06:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:55</td></tr> <tr><td>Control Test</td><td>0.201</td><td>06:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:57</td></tr> <tr><td>Control Test</td><td>0.200</td><td>06:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:58</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.2010</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.4975</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:54	Control Test	0.202	06:55	Air Blank	0.000	06:55	Control Test	0.201	06:56	Air Blank	0.000	06:57	Control Test	0.200	06:57	Air Blank	0.000	06:58	Control Test Stats			Average	0.2010		Std Dev	0.0010		Rel Std Dev(%)	0.4975		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005551 01/12/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:59</td></tr> <tr><td>Control Test</td><td>0.081</td><td>06:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr> <tr><td>Control Test</td><td>0.081</td><td>07:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:00</td></tr> <tr><td>Control Test</td><td>0.081</td><td>07:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:01</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>[Signature]</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	06:59	Control Test	0.081	06:59	Air Blank	0.000	07:00	Control Test	0.081	07:00	Air Blank	0.000	07:00	Control Test	0.081	07:01	Air Blank	0.000	07:01	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:45																																																																																																																																																	
Control Test	0.050	06:45																																																																																																																																																	
Air Blank	0.000	06:46																																																																																																																																																	
Control Test	0.050	06:47																																																																																																																																																	
Air Blank	0.000	06:47																																																																																																																																																	
Control Test	0.049	06:48																																																																																																																																																	
Air Blank	0.000	06:48																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0497																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:49																																																																																																																																																	
Control Test	0.080	06:50																																																																																																																																																	
Air Blank	0.000	06:50																																																																																																																																																	
Control Test	0.080	06:51																																																																																																																																																	
Air Blank	0.000	06:52																																																																																																																																																	
Control Test	0.081	06:52																																																																																																																																																	
Air Blank	0.000	06:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:54																																																																																																																																																	
Control Test	0.202	06:55																																																																																																																																																	
Air Blank	0.000	06:55																																																																																																																																																	
Control Test	0.201	06:56																																																																																																																																																	
Air Blank	0.000	06:57																																																																																																																																																	
Control Test	0.200	06:57																																																																																																																																																	
Air Blank	0.000	06:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2010																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.4975																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:59																																																																																																																																																	
Control Test	0.081	06:59																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test	0.081	07:00																																																																																																																																																	
Air Blank	0.000	07:00																																																																																																																																																	
Control Test	0.081	07:01																																																																																																																																																	
Air Blank	0.000	07:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005551, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005551</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>MIAMI DADE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/12/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:29</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/12/2023

Date


DAVID E REYES-RIVERA,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1