



Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 <i>6262</i>	<i>Volusia CSO</i>	<i>01/26/2023</i>	TDG <i>MG</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
<p>VOLUSIA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006262 01/26/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:27</td></tr><tr><td>Control Test</td><td>0.048</td><td>14:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:30</td></tr><tr><td>Control Test</td><td>0.048</td><td>14:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:32</td></tr><tr><td>Control Test</td><td>0.047</td><td>14:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:34</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0477</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2112</td><td></td></tr></table> <p>Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	14:27	Control Test	0.048	14:28	Air Blank	0.000	14:30	Control Test	0.048	14:31	Air Blank	0.000	14:32	Control Test	0.047	14:33	Air Blank	0.000	14:34	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel Std Dev(%)	1.2112		<p>VOLUSIA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006262 01/26/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:43</td></tr><tr><td>Control Test</td><td>0.077</td><td>14:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:44</td></tr><tr><td>Control Test</td><td>0.077</td><td>14:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:45</td></tr><tr><td>Control Test</td><td>0.077</td><td>14:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:46</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0770</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <p>Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	14:43	Control Test	0.077	14:44	Air Blank	0.000	14:44	Control Test	0.077	14:45	Air Blank	0.000	14:45	Control Test	0.077	14:46	Air Blank	0.000	14:46	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>VOLUSIA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006262 01/26/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:52</td></tr><tr><td>Control Test</td><td>0.197</td><td>14:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:53</td></tr><tr><td>Control Test</td><td>0.197</td><td>14:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr><tr><td>Control Test</td><td>0.196</td><td>14:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:55</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1967</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2936</td><td></td></tr></table> <p>Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	14:52	Control Test	0.197	14:52	Air Blank	0.000	14:53	Control Test	0.197	14:54	Air Blank	0.000	14:54	Control Test	0.196	14:55	Air Blank	0.000	14:55	Control Test Stats			Average	0.1967		Std Dev	0.0006		Rel Std Dev(%)	0.2936		<p>DGS</p> <p>VOLUSIA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006262 01/26/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr><tr><td>Control Test</td><td>0.081</td><td>14:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:57</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:58</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:59</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.4494</td><td></td></tr></table> <p>Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	14:56	Control Test	0.081	14:56	Air Blank	0.000	14:57	Control Test	0.079	14:57	Air Blank	0.000	14:58	Control Test	0.079	14:58	Air Blank	0.000	14:59	Control Test Stats			Average	0.0797		Std Dev	0.0012		Rel Std Dev(%)	1.4494	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:27																																																																																																																																																	
Control Test	0.048	14:28																																																																																																																																																	
Air Blank	0.000	14:30																																																																																																																																																	
Control Test	0.048	14:31																																																																																																																																																	
Air Blank	0.000	14:32																																																																																																																																																	
Control Test	0.047	14:33																																																																																																																																																	
Air Blank	0.000	14:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2112																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:43																																																																																																																																																	
Control Test	0.077	14:44																																																																																																																																																	
Air Blank	0.000	14:44																																																																																																																																																	
Control Test	0.077	14:45																																																																																																																																																	
Air Blank	0.000	14:45																																																																																																																																																	
Control Test	0.077	14:46																																																																																																																																																	
Air Blank	0.000	14:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:52																																																																																																																																																	
Control Test	0.197	14:52																																																																																																																																																	
Air Blank	0.000	14:53																																																																																																																																																	
Control Test	0.197	14:54																																																																																																																																																	
Air Blank	0.000	14:54																																																																																																																																																	
Control Test	0.196	14:55																																																																																																																																																	
Air Blank	0.000	14:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1967																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2936																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:56																																																																																																																																																	
Control Test	0.081	14:56																																																																																																																																																	
Air Blank	0.000	14:57																																																																																																																																																	
Control Test	0.079	14:57																																																																																																																																																	
Air Blank	0.000	14:58																																																																																																																																																	
Control Test	0.079	14:58																																																																																																																																																	
Air Blank	0.000	14:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.4494																																																																																																																																																		

Comments: Will perform a discretionary optical cal adjust to bring the 0.08 ARS and DGS tests into more nominal agreement. *MG 01/26/2023*



VOLUSIA COUNTY S.O.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006262  
02/01/2023 10:54:08

Auto Calibration  
Max Power Res Value = 115  
Auto Range Res Value = 77

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12736, Sum Io = 12680

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1200 (-0.0030)  
Sample #2 = 0.0850 (0.0400)  
Sample #3 = 0.0880 (0.0610)  
Sample #4 = 0.1320 (0.0640)  
Avg % Abs = 0.1017 (0.0550)  
STD DEV = 0.0263 (0.0131)  
REL STD DEV = 25.881 (23.776)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.2110 (-0.0100)  
Sample #2 = 0.2040 (-0.0130)  
Sample #3 = 0.2040 (0.0050)  
Sample #4 = 0.2300 (-0.0010)  
Avg % Abs = 0.2127 (-0.0030)  
STD DEV = 0.0150 (0.0092)  
REL STD DEV = 7.059 (305.505)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12720, Sum Io = 12674

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8320 (-0.0130)  
Sample #2 = 0.8000 (0.0300)  
Sample #3 = 0.7600 (0.0570)  
Sample #4 = 0.7900 (0.0480)  
Avg % Abs = 0.7833 (0.0450)  
STD DEV = 0.0208 (0.0137)  
REL STD DEV = 2.657 (30.551)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.6300 (-0.0230)  
Sample #2 = 1.6200 (0.0030)  
Sample #3 = 1.6030 (0.0200)  
Sample #4 = 1.6000 (0.0170)  
Avg % Abs = 1.6077 (0.0133)  
STD DEV = 0.0108 (0.0091)  
REL STD DEV = 0.671 (68.053)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12708, Sum Io = 12667

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.8270 (-0.0100)  
Sample #2 = 1.7880 (0.0140)  
Sample #3 = 1.8110 (0.0220)  
Sample #4 = 1.7830 (0.0500)  
Avg % Abs = 1.7940 (0.0287)  
STD DEV = 0.0149 (0.0189)  
REL STD DEV = 0.832 (65.942)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.6660 (-0.0200)  
Sample #2 = 3.6240 (0.0050)  
Sample #3 = 3.6090 (0.0320)  
Sample #4 = 3.5880 (0.0330)  
Avg % Abs = 3.6070 (0.0233)  
STD DEV = 0.0181 (0.0159)  
REL STD DEV = 0.501 (68.079)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12692, Sum Io = 12659

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.4670 (-0.0010)  
Sample #2 = 3.4080 (0.0420)  
Sample #3 = 3.4460 (0.0140)  
Sample #4 = 3.4400 (0.0550)  
Avg % Abs = 3.4313 (0.0370)  
STD DEV = 0.0204 (0.0210)  
REL STD DEV = 0.595 (56.628)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 6.8860 (0.0110)  
Sample #2 = 6.8470 (0.0500)  
Sample #3 = 6.8480 (0.0280)  
Sample #4 = 6.8600 (0.0370)  
Avg % Abs = 6.8517 (0.0383)  
STD DEV = 0.0072 (0.0111)  
REL STD DEV = 0.106 (28.853)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12687, Sum Io = 12863

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.0900 (-0.0070)  
Sample #2 = 5.0880 (0.0000)  
Sample #3 = 5.0660 (0.0120)  
Sample #4 = 5.0360 (0.0240)  
Avg % Abs = 5.0633 (0.0120)  
STD DEV = 0.0261 (0.0120)  
REL STD DEV = 0.516 (100.000)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 9.9830 (0.0190)  
Sample #2 = 9.9900 (0.0150)  
Sample #3 = 9.9650 (0.0290)  
Sample #4 = 9.9760 (0.0310)  
Avg % Abs = 9.9770 (0.0250)  
STD DEV = 0.0125 (0.0087)  
REL STD DEV = 0.126 (34.871)

Optical Calibration	
SN:	80-006262
Agency:	Volusia CSO
Date:	02/01/2023
Quadratic Fit:	+/- 0.002g/210L
By:	TDG NG

# \*\*\*\*\* AUTO CAL DATA \*\*\*\*\*

## Channel 1 Data

Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.102  
Std Dev = 0.03 Rel Std Dev = 25.88  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.783  
Std Dev = 0.02 Rel Std Dev = 2.66  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.794  
Std Dev = 0.01 Rel Std Dev = 0.83  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.431  
Std Dev = 0.02 Rel Std Dev = 0.60  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.063  
Std Dev = 0.03 Rel Std Dev = 0.52  
Zero Order Coef = -294.04  
First Order Coef = 2800.19  
Second Order Coef = 15.95  
Standard Deviation = 15.714912

## Channel 2 Data

Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.213  
Std Dev = 0.02 Rel Std Dev = 7.66  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.608  
Std Dev = 0.01 Rel Std Dev = 0.67  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.607  
Std Dev = 0.02 Rel Std Dev = 0.50  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.852  
Std Dev = 0.01 Rel Std Dev = 0.11  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.977  
Std Dev = 0.01 Rel Std Dev = 0.13  
Zero Order Coef = -302.60  
First Order Coef = 1369.26  
Second Order Coef = 9.34  
Standard Deviation = 11.283382

## Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0002
0.040	0.040	-0.0004
0.100	0.100	0.0001
0.200	0.200	0.0001
0.300	0.300	-0.0001

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Channel 1  
Sample #1 = 2960.00  
Sample #2 = 2969.00  
Sample #3 = 2934.00  
Sample #4 = 2872.00  
Average Result = 2925.0000  
STD DEV = 49.1223  
REL STD DEV = 1.679

Channel 2  
Sample #1 = 3175.00  
Sample #2 = 3060.00  
Sample #3 = 3116.00  
Sample #4 = 3103.00  
Average Result = 3093.0000  
STD DEV = 29.3087  
REL STD DEV = 0.948

## Dry Gas H2O Adjust Results \*\*\*\*\*

Barometric Pressure = 1023  
3 um H2O Adjust (mg/l\*10,000) = 884  
9 um H2O Adjust (mg/l\*10,000) = 716  
\*\*\*\* AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0002
0.040	0.040	-0.0001
0.100	0.100	-0.0004
0.200	0.200	0.0005
0.300	0.300	-0.0002

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post-Cal)	80-00 6262	Volusia CSO	02/01/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
065			
VOLUSIA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006262 02/01/2023 Software: 8100.27	VOLUSIA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006262 02/01/2023 Software: 8100.27	VOLUSIA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006262 02/01/2023 Software: 8100.27	VOLUSIA COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006262 02/01/2023 Software: 8100.27
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time
Air Blank 0.000 12:23	Air Blank 0.000 12:24	Air Blank 0.000 12:17	Air Blank 0.000 12:11
Control Test 0.050 12:24	Control Test 0.079 12:25	Control Test 0.200 12:18	Control Test 0.082 12:12
Air Blank 0.000 12:25	Air Blank 0.000 12:26	Air Blank 0.000 12:19	Air Blank 0.000 12:12
Control Test 0.049 12:25	Control Test 0.079 12:26	Control Test 0.200 12:19	Control Test 0.079 12:12
Air Blank 0.000 12:26	Air Blank 0.000 12:27	Air Blank 0.000 12:20	Air Blank 0.000 12:13
Control Test 0.049 12:26	Control Test 0.079 12:27	Control Test 0.200 12:21	Control Test 0.079 12:13
Air Blank 0.000 12:27	Air Blank 0.000 12:28	Air Blank 0.000 12:21	Air Blank 0.000 12:14
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0493	Average 0.0790	Average 0.2000	Average 0.0800
Std Dev 0.0006	Std Dev 0.0000	Std Dev 0.0000	Std Dev 0.0017
Rel Std Dev(%) 1.1703	Rel Std Dev(%) 0.0000	Rel Std Dev(%) 0.0000	Rel Std Dev(%) 2.1651
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Comments:



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: VOLUSIA COUNTY S.O.  
Time of Inspection: 14:52

Date of Inspection: 02/01/2023

Serial Number: 80-006262  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:00521080A2 Exp: 02/05/2023
0.000	0.049	0.078	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.050	0.078	0.199	0.081
0.000	0.050	0.078	0.199	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.078	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.050	0.078	0.199	0.081
0.000	0.050	0.079	0.199	0.081
0.000	0.050	0.079	0.199	0.081

Standard Deviations	0.0005	0.0005	0.0003	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/01/2023  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006262, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006262</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>VOLUSIA COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/01/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:52</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/01/2023

Date

  
TAYLOR D GUTSCHOW,  
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1