



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol

S/N 80-006637

Florida Department of Law Enforcement

Date In 09-05-2023

DI Completion Date 09-12-2023

Ship  P/U  H/D  CMI  EE

<b>Intake</b> By <u>ALL</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>missing both lower feet</u>	<b>Quality Checks</b> By <u>IS</u> Date <u>09-11-2023</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>229</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.253</u> (.228 - .278) 103 mm <u>0.531</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks	<b>Flow Calibration</b> By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)															
	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202303K 03-29-2025</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202303L 03-29-2025</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202304C 04-05-2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>06723080A5 04-05-2025</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202303K 03-29-2025	0.080	MP6292	202303L 03-29-2025	0.200	MP6293	202304C 04-05-2025	0.080 DGS	N/A	06723080A5 04-05-2025	<b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____
Simulator	Serial #	Lot #/Exp															
0.050	MP6291	202303K 03-29-2025															
0.080	MP6292	202303L 03-29-2025															
0.200	MP6293	202304C 04-05-2025															
0.080 DGS	N/A	06723080A5 04-05-2025															

<b>Calibration Adjustment</b> By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> By <u>IS</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1014</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP6289																																																												
Interferent	MP6290																																																												
0.050	MP6291																																																												
0.080	MP6292																																																												
0.200	MP6293																																																												

Notes/Suggested Service: _____ _____ _____ _____ _____ _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  <table border="0"> <tr> <td style="width: 50%;">           Benjamin Siddoway  <small>Digitally signed by Benjamin Siddoway Date: 2023.09.13 08:40:27 -04'00'</small> </td> <td style="width: 50%;">           Phil Nicodemo  <small>Digitally signed by Phil Nicodemo Date: 2023.09.13 13:12:40 -04'00'</small> </td> </tr> <tr> <td style="text-align: center;">Tech Review / Date</td> <td style="text-align: center;">Admin Review / Date</td> </tr> </table>	Benjamin Siddoway <small>Digitally signed by Benjamin Siddoway Date: 2023.09.13 08:40:27 -04'00'</small>	Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.09.13 13:12:40 -04'00'</small>	Tech Review / Date	Admin Review / Date
Benjamin Siddoway <small>Digitally signed by Benjamin Siddoway Date: 2023.09.13 08:40:27 -04'00'</small>	Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.09.13 13:12:40 -04'00'</small>				
Tech Review / Date	Admin Review / Date				

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 10:48

Date of Inspection: 09/12/2023

Serial Number: 80-006637  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.078	0.199	0.078
0.000	0.050	0.078	0.200	0.078
0.000	0.049	0.078	0.200	0.077
0.000	0.050	0.078	0.200	0.076
0.000	0.050	0.078	0.201	0.076
0.000	0.050	0.078	0.200	0.077
0.000	0.049	0.079	0.199	0.078

Standard Deviations	0.0005	0.0004	0.0006	0.0011
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


*Israel Soto*  
\_\_\_\_\_  
ISRAEL SOTO  
Signature and Printed Name

09/12/2023  
Date

# Stability checks

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006637  
09/11/2023  
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	13:49
Control Test	0.049	13:49
Air Blank	0.000	13:50
Control Test	0.049	13:51
Air Blank	0.000	13:51
Control Test	0.049	13:52
Air Blank	0.000	13:53
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

  
-----  
Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006637  
09/11/2023  
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	13:54
Control Test	0.079	13:55
Air Blank	0.000	13:55
Control Test	0.078	13:56
Air Blank	0.000	13:56
Control Test	0.079	13:57
Air Blank	0.000	13:58
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

wet

  
-----  
Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006637  
09/11/2023  
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	13:59
Control Test	0.197	14:00
Air Blank	0.000	14:01
Control Test	0.197	14:01
Air Blank	0.000	14:02
Control Test	0.197	14:03
Air Blank	0.000	14:03
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

  
-----  
Operator's Signature

FL HIGHWAY PATROL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006637  
09/11/2023  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:04
Control Test	0.079	14:04
Air Blank	0.000	14:05
Control Test	0.079	14:05
Air Blank	0.000	14:06
Control Test	0.079	14:06
Air Blank	0.000	14:07
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Dry

  
-----  
Operator's Signature



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road.  
Suite B1032  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006637, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006637</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/12/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:48</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

**Israel Soto**  
Soto  
Digitally signed by Israel Soto  
Date: 2023.09.12 11:32:09 -04'00'

**09/12/2023**

Date

**ISRAEL SOTO,**  
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality