



## INSTRUMENT PROCESSING SHEET

Agency Miami-Dade PDS/N 80-006895Florida Department of  
Law EnforcementDate In 03/10/2023 DI Completion Date 04/07/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>04/06/2023</u>	Flow Calibration	By	Date																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>111</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28663</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
					<b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____ _____																														
<b>Calibration Adjustment</b> By _____				<b>Department Inspection</b> By <u>TDG</u>																															
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28663</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1021</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP4864</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP4864	0.200	MP6288																
Simulator	Serial Number																																		
0.000	MP6284																																		
Interferent	MP6285																																		
0.050	MP6286																																		
0.080	MP4864																																		
0.200	MP6288																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																															
Israel Soto <small>Digitally signed by Israel Soto Date: 2023.04.10 11:50:20 -0400</small>				Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.04.11 09:34:06 -0400</small>																															
Tech Review / Date				Admin Review / Date																															

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 6895	Miami - Dade PD	04/06/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ✓	0.077 to 0.083 ✓	0.194 to 0.206 ✓	0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006895 04/06/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>13:37</td></tr><tr><td>Control Test</td><td>0.052</td><td>13:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr><tr><td>Control Test</td><td>0.052</td><td>13:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr><tr><td>Control Test</td><td>0.050</td><td>13:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0513</td><td></td></tr><tr><td>Std Dev</td><td>0.0012</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>2.2494</td><td></td></tr></table> <p>MG Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	13:37	Control Test	0.052	13:38	Air Blank	0.000	13:38	Control Test	0.052	13:39	Air Blank	0.000	13:39	Control Test	0.050	13:40	Air Blank	0.000	13:41	Control Test Stats			Average	0.0513		Std Dev	0.0012		Rel Std Dev(%)	2.2494		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006895 04/06/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>13:44</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:45</td></tr><tr><td>Control Test</td><td>0.081</td><td>13:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:46</td></tr><tr><td>Control Test</td><td>0.081</td><td>13:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:47</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0807</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr></table> <p>MG Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	13:44	Control Test	0.080	13:44	Air Blank	0.000	13:45	Control Test	0.081	13:45	Air Blank	0.000	13:46	Control Test	0.081	13:47	Air Blank	0.000	13:47	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157		<p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006895 04/06/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr><tr><td>Control Test</td><td>0.203</td><td>13:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr><tr><td>Control Test</td><td>0.200</td><td>13:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr><tr><td>Control Test</td><td>0.200</td><td>13:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:54</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2010</td><td></td></tr><tr><td>Std Dev</td><td>0.0017</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.8617</td><td></td></tr></table> <p>MG Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	13:51	Control Test	0.203	13:51	Air Blank	0.000	13:52	Control Test	0.200	13:53	Air Blank	0.000	13:53	Control Test	0.200	13:54	Air Blank	0.000	13:54	Control Test Stats			Average	0.2010		Std Dev	0.0017		Rel Std Dev(%)	0.8617		<p>DGS</p> <p>MIAMI-DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006895 04/06/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr><tr><td>Control Test</td><td>0.081</td><td>13:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:32</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></table> <p>MG Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	13:30	Control Test	0.081	13:31	Air Blank	0.000	13:31	Control Test	0.080	13:32	Air Blank	0.000	13:32	Control Test	0.080	13:32	Air Blank	0.000	13:33	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:37																																																																																																																																																	
Control Test	0.052	13:38																																																																																																																																																	
Air Blank	0.000	13:38																																																																																																																																																	
Control Test	0.052	13:39																																																																																																																																																	
Air Blank	0.000	13:39																																																																																																																																																	
Control Test	0.050	13:40																																																																																																																																																	
Air Blank	0.000	13:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0513																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	2.2494																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:44																																																																																																																																																	
Control Test	0.080	13:44																																																																																																																																																	
Air Blank	0.000	13:45																																																																																																																																																	
Control Test	0.081	13:45																																																																																																																																																	
Air Blank	0.000	13:46																																																																																																																																																	
Control Test	0.081	13:47																																																																																																																																																	
Air Blank	0.000	13:47																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:51																																																																																																																																																	
Control Test	0.203	13:51																																																																																																																																																	
Air Blank	0.000	13:52																																																																																																																																																	
Control Test	0.200	13:53																																																																																																																																																	
Air Blank	0.000	13:53																																																																																																																																																	
Control Test	0.200	13:54																																																																																																																																																	
Air Blank	0.000	13:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2010																																																																																																																																																		
Std Dev	0.0017																																																																																																																																																		
Rel Std Dev(%)	0.8617																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:30																																																																																																																																																	
Control Test	0.081	13:31																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test	0.080	13:32																																																																																																																																																	
Air Blank	0.000	13:32																																																																																																																																																	
Control Test	0.080	13:32																																																																																																																																																	
Air Blank	0.000	13:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

Comments:



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI-DADE PD  
Time of Inspection: 11:23

Date of Inspection: 04/07/2023

Serial Number: 80-006895  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.050	0.080	0.201	0.081
0.000	0.050	0.079	0.201	0.081
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.051	0.079	0.202	0.079
0.000	0.051	0.080	0.201	0.080
0.000	0.051	0.079	0.202	0.080
0.000	0.051	0.080	0.201	0.079
0.000	0.051	0.080	0.201	0.080


Standard Deviations	0.0005	0.0004	0.0004	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/07/2023  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006895, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006895</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MIAMI-DADE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/07/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:23</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/07/2023

Date

TAYLOR D GUTSCHOW,  
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1