



INSTRUMENT PROCESSING SHEET

Agency Broward CSOS/N 80-006925Florida Department of
Law EnforcementDate In 07/13/2023 DI Completion Date 07/17/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EETDG
7/21/2023

Intake	By TDG	Quality Checks	By TDG	Date 07/17/2023	Flow Calibration	By	Date																																							
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>130</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP5094	202201C 01/11/2024																																												
0.080	MP5095	202201D 01/18/2024																																												
0.200	MP5096	202201E 01/18/2024																																												
0.080 DGS	N/A	AG223802 08/26/2024																																												
Calibration Adjustment		By _____		Department Inspection			By <u>TDG</u>																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP5092																																													
Interferent	MP5093																																													
0.050	MP5094																																													
0.080	MP5095																																													
0.200	MP5096																																													
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																						
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Notes/Suggested Service: _____ _____ Admin Review: Indicated instrument returned from repair on IPS. (TDG 7/21/2023) _____ _____ _____ _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																										
				Israel Soto <small>Digitally signed by Israel Soto Date: 2023.07.18 08:00:59 +0400</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.07.21 13:43:01 -0400</small>																																										
				Tech Review / Date _____ Admin Review / Date _____																																										

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006925	Broward CSO	07/17/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 07/7/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:04</p> <p>Control Test 0.050 10:05</p> <p>Air Blank 0.000 10:05</p> <p>Control Test 0.050 10:06</p> <p>Air Blank 0.000 10:06</p> <p>Control Test 0.049 10:07</p> <p>Air Blank 0.000 10:08</p> <p>Control Test Stats</p> <p>Average 0.0497</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1625</p>	<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 07/17/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:10</p> <p>Control Test 0.079 10:11</p> <p>Air Blank 0.000 10:12</p> <p>Control Test 0.078 10:13</p> <p>Air Blank 0.000 10:13</p> <p>Control Test 0.080 10:14</p> <p>Air Blank 0.000 10:14</p> <p>Control Test Stats</p> <p>Average 0.0790</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 1.2658</p>	<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 07/17/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:18</p> <p>Control Test 0.202 10:18</p> <p>Air Blank 0.000 10:19</p> <p>Control Test 0.202 10:20</p> <p>Air Blank 0.000 10:20</p> <p>Control Test 0.202 10:21</p> <p>Air Blank 0.000 10:21</p> <p>Control Test Stats</p> <p>Average 0.2020</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 07/17/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:23</p> <p>Control Test 0.077 10:23</p> <p>Air Blank 0.000 10:24</p> <p>Control Test 0.078 10:24</p> <p>Air Blank 0.000 10:25</p> <p>Control Test 0.078 10:25</p> <p>Air Blank 0.000 10:25</p> <p>Control Test Stats</p> <p>Average 0.0777</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7434</p>
<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 12:58

Date of Inspection: 07/17/2023

Serial Number: 80-006925
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.079	0.202	0.077
0.000	0.049	0.079	0.202	0.078
0.000	0.050	0.079	0.202	0.078
0.000	0.050	0.079	0.202	0.077
0.000	0.050	0.079	0.203	0.078
0.000	0.050	0.079	0.202	0.077
0.000	0.049	0.079	0.203	0.078
0.000	0.050	0.079	0.203	0.077
0.000	0.050	0.079	0.203	0.077
0.000	0.050	0.079	0.203	0.077

Standard Deviations	0.0004	0.0000	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.



I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/17/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006925, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006925</u>	UNCERTAINTY* \pm
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/17/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:58</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/17/2023

Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Broward CSOS/N 80-006925Florida Department of
Law EnforcementDate In 02/07/2023

DI Completion Date _____

☐ Ship☐ P/U☐ H/D☒ CMI☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>02/16/2023</u>	Flow Calibration	By	Date																																						
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Missing two screws on the dry gas regulator valve</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>145</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																								
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5092</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5093</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5094</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024																												
Simulator	Serial #	Lot #/Exp																																											
0.050	MP5092	202201C 01/11/2024																																											
0.080	MP5093	202201D 01/18/2024																																											
0.200	MP5094	202201E 01/18/2024																																											
0.080 DGS	N/A	AG223802 08/26/2024																																											
Calibration Adjustment By <u>TDG</u> Barometric Pressure Gauge <u>1024 / 1024</u> ID # <u>28199</u>		Department Inspection By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5099</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5096</td><td>21410</td><td>09/30/2023</td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5099	N/A	N/A	0.040	MP5096	21410	09/30/2023	0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table>				Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial #	Lot #	Expiration																																										
0.000	MP5099	N/A	N/A																																										
0.040	MP5096	21410	09/30/2023																																										
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000																																													
Interferent																																													
0.050																																													
0.080																																													
0.200																																													
<input type="checkbox"/> Post Calibration Adjustment Stability Checks		Attachments <input type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustments <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Form 51</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: <u>Optical calibration adjustments were unsuccessful and could not be completed. Did not perform post-calibration Stability Checks or Department Inspection. Did not determine compliance with Chapter 11D-8, FAC. Will send to repair. (TDG)</u>		Tech Review / Date _____ Admin Review / Date _____																																											

Return Material Authorization

Ship to:

☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Anaya Frazier on 02/28/2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-006925

Bill To Address:

Broward CSO

Attn: Anaya Frazier

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Cannot conduct a successful optical calibration adjustment. Keeps giving "Fault Det" message during the 0.04 g/210 L portion.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Anaya Frazier

Phone #: 305-218-6752

Email: Anaya_Frazier@sheriff.org

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 6925	Broward CSO	02/16/2023	TDG MG

0.05g/210L			0.08g/210L			0.20g/210L			DGS 0.08g/210L																																																																																																																																																		
0.047 to 0.053			0.077 to 0.083			0.194 to 0.206			0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																		
<div>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 02/16/2023 Software: 8100.27</div> <table><tr><td>Test</td><td>g/210L</td><td>Time</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr><tr><td>Control Test</td><td>0.047</td><td>10:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr><tr><td>Control Test</td><td>0.046</td><td>11:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr><tr><td>Control Test</td><td>0.048</td><td>11:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0470</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>2.1277</td><td></td></tr></table> <div>Operator's Signature</div>			Test	g/210L	Time	Air Blank	0.000	10:58	Control Test	0.047	10:59	Air Blank	0.000	11:00	Control Test	0.046	11:00	Air Blank	0.000	11:01	Control Test	0.048	11:02	Air Blank	0.000	11:02	Control Test Stats			Average	0.0470		Std Dev	0.0010		Rel Std Dev(%)	2.1277		<div>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 02/16/2023 Software: 8100.27</div> <table><tr><td>Test</td><td>g/210L</td><td>Time</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:49</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr><tr><td>Control Test</td><td>0.077</td><td>10:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr><tr><td>Control Test</td><td>0.076</td><td>10:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0773</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.9753</td><td></td></tr></table> <div>Operator's Signature</div>			Test	g/210L	Time	Air Blank	0.000	10:49	Control Test	0.079	10:50	Air Blank	0.000	10:51	Control Test	0.077	10:51	Air Blank	0.000	10:52	Control Test	0.076	10:53	Air Blank	0.000	10:53	Control Test Stats			Average	0.0773		Std Dev	0.0015		Rel Std Dev(%)	1.9753		<div>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 02/16/2023 Software: 8100.27</div> <table><tr><td>Test</td><td>g/210L</td><td>Time</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr><tr><td>Control Test</td><td>0.199</td><td>11:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr><tr><td>Control Test</td><td>0.197</td><td>11:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr><tr><td>Control Test</td><td>0.194</td><td>11:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:10</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1967</td><td></td></tr><tr><td>Std Dev</td><td>0.0025</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2796</td><td></td></tr></table> <div>Operator's Signature</div>			Test	g/210L	Time	Air Blank	0.000	11:06	Control Test	0.199	11:07	Air Blank	0.000	11:07	Control Test	0.197	11:08	Air Blank	0.000	11:08	Control Test	0.194	11:09	Air Blank	0.000	11:10	Control Test Stats			Average	0.1967		Std Dev	0.0025		Rel Std Dev(%)	1.2796		<div>DGS BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006925 02/16/2023 Software: 8100.27</div> <table><tr><td>Test</td><td>g/210L</td><td>Time</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr><tr><td>Control Test</td><td>0.077</td><td>11:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr><tr><td>Control Test</td><td>0.082</td><td>11:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0793</td><td></td></tr><tr><td>Std Dev</td><td>0.0025</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>3.1722</td><td></td></tr></table> <div>Operator's Signature</div>			Test	g/210L	Time	Air Blank	0.000	11:11	Control Test	0.077	11:11	Air Blank	0.000	11:12	Control Test	0.082	11:12	Air Blank	0.000	11:12	Control Test	0.079	11:13	Air Blank	0.000	11:13	Control Test Stats			Average	0.0793		Std Dev	0.0025		Rel Std Dev(%)	3.1722	
Test	g/210L	Time																																																																																																																																																									
Air Blank	0.000	10:58																																																																																																																																																									
Control Test	0.047	10:59																																																																																																																																																									
Air Blank	0.000	11:00																																																																																																																																																									
Control Test	0.046	11:00																																																																																																																																																									
Air Blank	0.000	11:01																																																																																																																																																									
Control Test	0.048	11:02																																																																																																																																																									
Air Blank	0.000	11:02																																																																																																																																																									
Control Test Stats																																																																																																																																																											
Average	0.0470																																																																																																																																																										
Std Dev	0.0010																																																																																																																																																										
Rel Std Dev(%)	2.1277																																																																																																																																																										
Test	g/210L	Time																																																																																																																																																									
Air Blank	0.000	10:49																																																																																																																																																									
Control Test	0.079	10:50																																																																																																																																																									
Air Blank	0.000	10:51																																																																																																																																																									
Control Test	0.077	10:51																																																																																																																																																									
Air Blank	0.000	10:52																																																																																																																																																									
Control Test	0.076	10:53																																																																																																																																																									
Air Blank	0.000	10:53																																																																																																																																																									
Control Test Stats																																																																																																																																																											
Average	0.0773																																																																																																																																																										
Std Dev	0.0015																																																																																																																																																										
Rel Std Dev(%)	1.9753																																																																																																																																																										
Test	g/210L	Time																																																																																																																																																									
Air Blank	0.000	11:06																																																																																																																																																									
Control Test	0.199	11:07																																																																																																																																																									
Air Blank	0.000	11:07																																																																																																																																																									
Control Test	0.197	11:08																																																																																																																																																									
Air Blank	0.000	11:08																																																																																																																																																									
Control Test	0.194	11:09																																																																																																																																																									
Air Blank	0.000	11:10																																																																																																																																																									
Control Test Stats																																																																																																																																																											
Average	0.1967																																																																																																																																																										
Std Dev	0.0025																																																																																																																																																										
Rel Std Dev(%)	1.2796																																																																																																																																																										
Test	g/210L	Time																																																																																																																																																									
Air Blank	0.000	11:11																																																																																																																																																									
Control Test	0.077	11:11																																																																																																																																																									
Air Blank	0.000	11:12																																																																																																																																																									
Control Test	0.082	11:12																																																																																																																																																									
Air Blank	0.000	11:12																																																																																																																																																									
Control Test	0.079	11:13																																																																																																																																																									
Air Blank	0.000	11:13																																																																																																																																																									
Control Test Stats																																																																																																																																																											
Average	0.0793																																																																																																																																																										
Std Dev	0.0025																																																																																																																																																										
Rel Std Dev(%)	3.1722																																																																																																																																																										

Comments: The 0.05 and 0.08 ARS are outside the nominal range. Will perform an optical cal adjust. MG 02/16/2023

#1

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006925
02/23/2023 10:28:07

Auto Calibration
Max Power Res Value = 95
Auto Range Res Value = 79

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12632, 9um Io = 12115
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0770 (-0.0220)
Sample #2 = 0.0610 (0.0100)
Sample #3 = 0.0180 (0.0150)
Sample #4 = 0.0850 (0.0020)
Avg % Abs = 0.0547 (0.0090)
STD DEV = 0.0339 (0.0066)
REL STD DEV = 62.096 (72.860)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1010 (0.0630)
Sample #2 = 0.2490 (0.0440)
Sample #3 = 0.1500 (-0.0210)
Sample #4 = 0.0330 (0.0570)
Avg % Abs = 0.1440 (0.0267)
STD DEV = 0.1081 (0.0418)
REL STD DEV = 75.087 (156.709)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12623, 9um Io = 12111
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7600 (0.0000)
Sample #2 = 0.7360 (0.0080)
Sample #3 = 0.7360 (0.0160)
Sample #4 = 0.7690 (-0.0090)
Avg % Abs = 0.7470 (0.0050)
STD DEV = 0.0191 (0.0128)
REL STD DEV = 2.551 (255.343)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4430 (-0.0870)
Sample #2 = 1.4190 (-0.1590)
Sample #3 = 1.4250 (-0.1600)
Sample #4 = 1.4980 (-0.2520)
Avg % Abs = 1.4473 (-0.1903)
STD DEV = 0.0440 (0.0534)
REL STD DEV = 3.039 (28.060)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12617, 9um Io = 12116

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7390 (-0.0100)
Sample #2 = 0.7280 (0.0010)
Sample #3 = 0.7340 (0.0030)
Sample #4 = 0.7960 (-0.0200)
Avg % Abs = 0.7527 (-0.0053)
STD DEV = 0.0376 (0.0127)
REL STD DEV = 5.002 (238.894)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.3950 (-0.0340)
Sample #2 = 1.4230 (-0.1170)
Sample #3 = 1.3830 (-0.0660)
Sample #4 = 1.3780 (-0.1300)
Avg % Abs = 1.3947 (-0.1043)
STD DEV = 0.0247 (0.0338)
REL STD DEV = 1.768 (32.423)

**** AUTO CAL FAIL

#2

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006925
02/23/2023 11:48:02

Auto Calibration
Max Power Res Value = 96
Auto Range Res Value = 80

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12702, 9um Io = 12169
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0590 (-0.0280)
Sample #2 = 0.0090 (0.0030)
Sample #3 = 0.0090 (0.0060)
Sample #4 = 0.0290 (0.0020)
Avg % Abs = 0.0157 (0.0037)
STD DEV = 0.0115 (0.0021)
REL STD DEV = 73.704 (56.773)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0380 (0.0540)
Sample #2 = 0.1050 (0.0560)
Sample #3 = -0.0210 (-0.0770)
Sample #4 = 0.0810 (-0.0830)
Avg % Abs = 0.0550 (-0.0347)
STD DEV = 0.0669 (0.0786)
REL STD DEV = 121.642 (226.664)

Optical Calibration 1+2	
SN:	80-006925
Agency:	Broward CSO
Date:	02/23/2023
Quadratic Fit:	+/- 0.002g/210L
By:	TDG ML

Could not conduct successful cal adjust. Kept receiving "Fault Det" message on the 0.04 solution and could not pass. Tried 2 separate adjustments w/ the same results.
ML 02/23/2023

#2

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12700, 9um Io = 12210

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7950 (-0.0160)
Sample #2 = 0.7540 (-0.0140)
Sample #3 = 0.7850 (-0.0060)
Sample #4 = 0.7130 (0.0350)
Avg % Abs = 0.7507 (0.0050)
STD DEV = 0.0361 (0.0263)
REL STD DEV = 4.811 (525.738)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.7890 (-0.0430)
Sample #2 = 1.3790 (0.1350)
Sample #3 = 1.1560 (0.3910)
Sample #4 = 1.2750 (0.3590)
Avg % Abs = 1.2700 (0.2950)
STD DEV = 0.1116 (0.1395)
REL STD DEV = 8.786 (47.283)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12697, 9um Io = 12185
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7840 (-0.0130)
Sample #2 = 0.7350 (0.0130)
Sample #3 = 0.7180 (0.0240)
Sample #4 = 0.7850 (-0.0060)
Avg % Abs = 0.7460 (0.0103)
STD DEV = 0.0348 (0.0152)
REL STD DEV = 4.669 (146.872)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.3500 (-0.0430)
Sample #2 = 1.2860 (-0.0720)
Sample #3 = 1.2950 (-0.1470)
Sample #4 = 1.3980 (-0.2610)
Avg % Abs = 1.3263 (-0.1600)
STD DEV = 0.0622 (0.0952)
REL STD DEV = 4.692 (59.480)

#2

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12692, 9um Io = 12211

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7260 (-0.0050)
Sample #2 = 0.7770 (-0.0210)
Sample #3 = 0.7480 (0.0080)
Sample #4 = 0.7300 (0.0100)
Avg % Abs = 0.7517 (-0.0010)
STD DEV = 0.0237 (0.0173)
REL STD DEV = 3.155 (1734.935)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.3750 (0.0490)
Sample #2 = 1.4290 (0.0550)
Sample #3 = 1.1520 (0.0920)
Sample #4 = 1.4350 (-0.0300)
Avg % Abs = 1.3387 (0.0390)
STD DEV = 0.1617 (0.0626)
REL STD DEV = 12.078 (160.395)

**** AUTO CAL FAIL