



S/N 80-007373

Date In 11/02/2023 DI Completion Date 11/08/2023

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 11/08/2023																			
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <div style="display:flex; justify-content:space-between;"> <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf </div> <div style="display:flex; justify-content:space-between;"> <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight </div> Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>233</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.253</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse; margin-top: 5px;"> <thead> <tr> <th style="width:15%;">Simulator</th><th style="width:25%;">Serial #</th><th style="width:60%;">Lot #/Exp</th></tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td><td rowspan="2">MP5094</td><td>202303K</td></tr> <tr> <td>03/29/2025</td></tr> <tr> <td rowspan="2">0.080</td><td rowspan="2">MP5095</td><td>202303L</td></tr> <tr> <td>03/29/2025</td></tr> <tr> <td rowspan="2">0.200</td><td rowspan="2">MP5096</td><td>202304C</td></tr> <tr> <td>04/05/2025</td></tr> <tr> <td rowspan="2">0.080 DGS</td><td rowspan="2">N/A</td><td>AG223802</td></tr> <tr> <td>08/26/2024</td></tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP5094	202303K	03/29/2025	0.080	MP5095	202303L	03/29/2025	0.200	MP5096	202304C	04/05/2025	0.080 DGS	N/A	AG223802	08/26/2024
Simulator	Serial #	Lot #/Exp																					
0.050	MP5094	202303K																					
		03/29/2025																					
0.080	MP5095	202303L																					
		03/29/2025																					
0.200	MP5096	202304C																					
		04/05/2025																					
0.080 DGS	N/A	AG223802																					
		08/26/2024																					





Flow Calibration	By	Date
Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) _____ Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)		

Maintenance	By TDG	Date
<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Replaced the internal printer paper on 11/8 after Stability Checks</u> _____ _____ _____		

Calibration Adjustment	By																												
Barometric Pressure Gauge _____ ID # _____																													
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th style="width:20%;">Simulator</th><th style="width:20%;">Serial #</th><th style="width:20%;">Lot #</th><th style="width:40%;">Expiration</th></tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			
Simulator	Serial #	Lot #	Expiration																										
0.000		N/A	N/A																										
0.040																													
0.100																													
0.200																													
0.300																													
0.080 DGS	N/A																												
<input type="checkbox"/> Post Calibration Adjustment Stability Checks																													
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th style="width:20%;">Simulator</th><th style="width:20%;">Serial #</th><th style="width:20%;">Lot #</th><th style="width:40%;">Expiration</th></tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A											
Simulator	Serial #	Lot #	Expiration																										
0.050																													
0.080																													
0.200																													
0.080 DGS	N/A																												

Department Inspection	By TDG												
Barometric Pressure ID# <u>26932</u>													
Gauge <u>1018</u> Instrument <u>1018</u>													
Mouth Alcohol Solution Lot # <u>2023-A</u>													
Acetone Stock Solution Lot # <u>2022-B</u>													
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th style="width:50%;">Simulator</th><th style="width:50%;">Serial Number</th></tr> </thead> <tbody> <tr><td>0.000</td><td>MP5092</td></tr> <tr><td>Interferent</td><td>MP5093</td></tr> <tr><td>0.050</td><td>MP5094</td></tr> <tr><td>0.080</td><td>MP5095</td></tr> <tr><td>0.200</td><td>MP5096</td></tr> </tbody> </table>	Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096	
Simulator	Serial Number												
0.000	MP5092												
Interferent	MP5093												
0.050	MP5094												
0.080	MP5095												
0.200	MP5096												
Attachments													
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use													
Benjamin Siddoway	Phil Nicodemmo												
Tech Review / Date	Admin Review / Date												

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
≤0.003 of Wet			≤0.003 of Wet																																																																																																																																																
✓			✓																																																																																																																																																
<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2023 Software: 8100.27</p> <p>SN 80-007373</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	10:21	Control Test	0.049	10:22	Air Blank	0.000	10:22	Control Test	0.049	10:23	Air Blank	0.000	10:23	Control Test	0.048	10:24	Air Blank	0.000	10:25	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2023 Software: 8100.27</p> <p>SN 80-007373</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	10:28	Control Test	0.079	10:29	Air Blank	0.000	10:29	Control Test	0.079	10:30	Air Blank	0.000	10:30	Control Test	0.079	10:31	Air Blank	0.000	10:32	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2023 Software: 8100.27</p> <p>SN 80-007373</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	10:35	Control Test	0.199	10:35	Air Blank	0.000	10:36	Control Test	0.199	10:37	Air Blank	0.000	10:37	Control Test	0.199	10:38	Air Blank	0.000	10:38	Control Test Stats			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2023 Software: 8100.27</p> <p>SN 80-007373</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	10:03	Control Test	0.081	10:03	Air Blank	0.000	10:03	Control Test	0.081	10:04	Air Blank	0.000	10:04	Control Test	0.081	10:05	Air Blank	0.000	10:05	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:21																																																																																																																																																	
Control Test	0.049	10:22																																																																																																																																																	
Air Blank	0.000	10:22																																																																																																																																																	
Control Test	0.049	10:23																																																																																																																																																	
Air Blank	0.000	10:23																																																																																																																																																	
Control Test	0.048	10:24																																																																																																																																																	
Air Blank	0.000	10:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:28																																																																																																																																																	
Control Test	0.079	10:29																																																																																																																																																	
Air Blank	0.000	10:29																																																																																																																																																	
Control Test	0.079	10:30																																																																																																																																																	
Air Blank	0.000	10:30																																																																																																																																																	
Control Test	0.079	10:31																																																																																																																																																	
Air Blank	0.000	10:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:35																																																																																																																																																	
Control Test	0.199	10:35																																																																																																																																																	
Air Blank	0.000	10:36																																																																																																																																																	
Control Test	0.199	10:37																																																																																																																																																	
Air Blank	0.000	10:37																																																																																																																																																	
Control Test	0.199	10:38																																																																																																																																																	
Air Blank	0.000	10:38																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:03																																																																																																																																																	
Control Test	0.081	10:03																																																																																																																																																	
Air Blank	0.000	10:03																																																																																																																																																	
Control Test	0.081	10:04																																																																																																																																																	
Air Blank	0.000	10:04																																																																																																																																																	
Control Test	0.081	10:05																																																																																																																																																	
Air Blank	0.000	10:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 13:04

Date of Inspection: 11/08/2023

Serial Number: 80-007373
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.048	0.079	0.199	0.081
0.000	0.049	0.078	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.081
0.000	0.048	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.079	0.199	0.081

Standard Deviations	0.0004	0.0003	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/08/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007373, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007373</u>	UNCERTAINTY* \pm
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/08/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:04</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence. ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



Date 11/08/2023 TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Anaya Frazier on 05/18/2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-007373

Bill To Address:

Broward County Sheriff's Office

Attn: Anaya Frazier

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Conducted two optical calibration adjustments but could not get post-cal stabilities to pass (see attached opticals and stabilities). Records have been uploaded by FDLE.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Anaya Frazier

Phone #: 305-218-6752

Email: Anaya_Frazier@sheriff.org

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities <i># mg</i>	80-00 <i>7373</i>	<i>Broward CSO</i>	<i>05/18/2023</i>	TDG <i>mg</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 X ≤0.003 of Wet																																																																																																																																																
<div>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr><tr><td>Control Test</td><td>0.048</td><td>10:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr><tr><td>Control Test</td><td>0.051</td><td>10:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr><tr><td>Control Test</td><td>0.053</td><td>10:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0507</td><td></td></tr><tr><td>Std Dev</td><td>0.0025</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>4.9670</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	10:10	Control Test	0.048	10:10	Air Blank	0.000	10:11	Control Test	0.051	10:12	Air Blank	0.000	10:12	Control Test	0.053	10:13	Air Blank	0.000	10:13	Control Test Stats			Average	0.0507		Std Dev	0.0025		Rel Std Dev(%)	4.9670		<div>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr><tr><td>Control Test</td><td>0.082</td><td>10:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.9015</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	10:18	Control Test	0.080	10:18	Air Blank	0.000	10:19	Control Test	0.079	10:20	Air Blank	0.000	10:20	Control Test	0.082	10:21	Air Blank	0.000	10:21	Control Test Stats			Average	0.0803		Std Dev	0.0015		Rel Std Dev(%)	1.9015		<div>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr><tr><td>Control Test</td><td>0.199</td><td>10:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr><tr><td>Control Test</td><td>0.200</td><td>10:28</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr><tr><td>Control Test</td><td>0.201</td><td>10:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.2000</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.5000</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	10:26	Control Test	0.199	10:27	Air Blank	0.000	10:27	Control Test	0.200	10:28	Air Blank	0.000	10:28	Control Test	0.201	10:29	Air Blank	0.000	10:30	Control Test Stats			Average	0.2000		Std Dev	0.0010		Rel Std Dev(%)	0.5000		<div>DGS BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr><tr><td>Control Test</td><td>0.085</td><td>10:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr><tr><td>Control Test</td><td>0.081</td><td>10:06</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0820</td><td></td></tr><tr><td>Std Dev</td><td>0.0026</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>3.2265</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	10:04	Control Test	0.085	10:04	Air Blank	0.000	10:05	Control Test	0.080	10:05	Air Blank	0.000	10:06	Control Test	0.081	10:06	Air Blank	0.000	10:06	Control Test Stats			Average	0.0820		Std Dev	0.0026		Rel Std Dev(%)	3.2265	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test	0.048	10:10																																																																																																																																																	
Air Blank	0.000	10:11																																																																																																																																																	
Control Test	0.051	10:12																																																																																																																																																	
Air Blank	0.000	10:12																																																																																																																																																	
Control Test	0.053	10:13																																																																																																																																																	
Air Blank	0.000	10:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0507																																																																																																																																																		
Std Dev	0.0025																																																																																																																																																		
Rel Std Dev(%)	4.9670																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:18																																																																																																																																																	
Control Test	0.080	10:18																																																																																																																																																	
Air Blank	0.000	10:19																																																																																																																																																	
Control Test	0.079	10:20																																																																																																																																																	
Air Blank	0.000	10:20																																																																																																																																																	
Control Test	0.082	10:21																																																																																																																																																	
Air Blank	0.000	10:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	1.9015																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:26																																																																																																																																																	
Control Test	0.199	10:27																																																																																																																																																	
Air Blank	0.000	10:27																																																																																																																																																	
Control Test	0.200	10:28																																																																																																																																																	
Air Blank	0.000	10:28																																																																																																																																																	
Control Test	0.201	10:29																																																																																																																																																	
Air Blank	0.000	10:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2000																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:04																																																																																																																																																	
Control Test	0.085	10:04																																																																																																																																																	
Air Blank	0.000	10:05																																																																																																																																																	
Control Test	0.080	10:05																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test	0.081	10:06																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0820																																																																																																																																																		
Std Dev	0.0026																																																																																																																																																		
Rel Std Dev(%)	3.2265																																																																																																																																																		

Comments:

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007373
05/18/2023 10:31:19
Auto Calibration
Max Power Res Value = 89
Auto Range Res Value = 66

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12689, 9um lo = 12633

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.1370 (0.0040)
Sample #2 = 0.1650 (0.0100)
Sample #3 = 0.1360 (0.0450)
Sample #4 = 0.1270 (0.0540)
Avg % Abs = 0.1427 (0.0363)
STD DEV = 0.0199 (0.0232)
REL STD DEV = 13.919 (63.977)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1940 (-0.0130)
Sample #2 = 0.1510 (-0.0520)
Sample #3 = 0.1570 (0.0300)
Sample #4 = 0.0810 (-0.0020)
Avg % Abs = 0.1297 (-0.0080)
STD DEV = 0.0423 (0.0413)
REL STD DEV = 32.586 (516.599)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12675, 9um lo = 12630

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.8430 (-0.0150)
Sample #2 = 0.8600 (-0.0080)
Sample #3 = 0.8420 (0.0100)
Sample #4 = 0.8630 (0.0060)
Avg % Abs = 0.8550 (0.0027)
STD DEV = 0.0114 (0.0095)
REL STD DEV = 1.328 (354.436)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5180 (0.0070)
Sample #2 = 1.5530 (0.0560)
Sample #3 = 1.4760 (0.1150)
Sample #4 = 1.4330 (0.1100)
Avg % Abs = 1.4873 (0.0937)
STD DEV = 0.0608 (0.0327)
REL STD DEV = 4.088 (34.928)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12671, 9um lo = 12614

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.8620 (-0.0100)
Sample #2 = 0.8590 (0.0060)
Sample #3 = 0.8730 (0.0000)
Sample #4 = 0.8640 (0.0160)
Avg % Abs = 0.8653 (0.0073)
STD DEV = 0.0071 (0.0081)
REL STD DEV = 0.820 (110.221)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5070 (-0.0200)
Sample #2 = 1.4390 (0.0050)
Sample #3 = 1.4870 (-0.0400)
Sample #4 = 1.4770 (-0.0560)
Avg % Abs = 1.4677 (-0.0303)
STD DEV = 0.0253 (0.0316)
REL STD DEV = 1.725 (104.268)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12665, 9um lo = 12624

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.9060 (0.0040)
Sample #2 = 1.8920 (0.0180)
Sample #3 = 1.9200 (0.0080)
Sample #4 = 1.9060 (0.0280)
Avg % Abs = 1.9060 (0.0180)
STD DEV = 0.0140 (0.0100)
REL STD DEV = 0.735 (55.556)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.4450 (0.0100)
Sample #2 = 3.4470 (0.0400)
Sample #3 = 3.4180 (0.0880)
Sample #4 = 3.4660 (0.0480)
Avg % Abs = 3.4437 (0.0587)
STD DEV = 0.0242 (0.0257)
REL STD DEV = 0.702 (43.835)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12660, 9um lo = 12627

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 3.6390 (-0.0140)
Sample #2 = 3.6410 (0.0000)
Sample #3 = 3.6080 (0.0190)
Sample #4 = 3.6130 (0.0120)
Avg % Abs = 3.6207 (0.0103)
STD DEV = 0.0178 (0.0096)
REL STD DEV = 0.491 (92.991)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 6.5350 (0.0220)
Sample #2 = 6.5240 (0.0570)
Sample #3 = 6.5320 (0.0930)
Sample #4 = 6.5060 (0.0690)
Avg % Abs = 6.5207 (0.0730)
STD DEV = 0.0133 (0.0183)
REL STD DEV = 0.204 (25.110)

Optical Calibration #1
SN: 80-007373
Agency: Broward CSD
Date: 05/18/2023
Quadratic Fit: +/- 0.002g/210L ✓
By: TDG ML

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12653, 9um lo = 12644

Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 5.3020 (-0.0070)
Sample #2 = 5.2670 (0.0220)
Sample #3 = 5.2890 (0.0140)
Sample #4 = 5.2960 (0.0220)
Avg % Abs = 5.2840 (0.0193)
STD DEV = 0.0151 (0.0046)
REL STD DEV = 0.286 (23.890)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 9.5080 (-0.0340)
Sample #2 = 9.4800 (0.1050)
Sample #3 = 9.4660 (0.0950)
Sample #4 = 9.5570 (0.0680)
Avg % Abs = 9.5010 (0.0893)
STD DEV = 0.0490 (0.0191)
REL STD DEV = 0.516 (21.425)

AUTO CAL DATA
Channel 1
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.143

Std Dev = 0.02 Rel Std Dev = 13.92
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.865
Std Dev = 0.01 Rel Std Dev = 0.82
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.906
Std Dev = 0.01 Rel Std Dev = 0.73
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.621
Std Dev = 0.02 Rel Std Dev = 0.49
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.284
Std Dev = 0.02 Rel Std Dev = 0.29
Zero Order Coef = -388.44
First Order Coef = 2651.49
Second Order Coef = 23.79
Standard Deviation = 11.830306

Channel 2
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.130
Std Dev = 0.04 Rel Std Dev = 32.59
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.468
Std Dev = 0.03 Rel Std Dev = 1.73
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.444
Std Dev = 0.02 Rel Std Dev = 0.70
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.521
Std Dev = 0.01 Rel Std Dev = 0.20
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.501
Std Dev = 0.05 Rel Std Dev = 0.52
Zero Order Coef = -185.44
First Order Coef = 1399.87
Second Order Coef = 13.06
Standard Deviation = 19.987240

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0002
0.040	0.040	-0.0004
0.100	0.100	0.0002
0.200	0.200	0.0000
0.300	0.300	-0.0000

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	0.0002
0.100	0.101	-0.0006
0.200	0.199	0.0005
0.300	0.300	-0.0002

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 2897.00
 Sample #2 = 2844.00
 Sample #3 = 2831.00
 Sample #4 = 2833.00
 Average Result = 2836.0000
 STD DEV = 7.0000
 REL STD DEV = 0.247

 ***** CHANNEL 2
 Sample #1 = 3292.00
 Sample #2 = 3366.00
 Sample #3 = 3320.00
 Sample #4 = 3281.00
 Average Result = 3322.3333
 STD DEV = 42.5480
 REL STD DEV = 1.281

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1014
 3 um H2O Adjust (mg/l x 10,000) = 973
 9 um H2O Adjust (mg/l x 10,000) = 487
 ***** AUTO CAL PASS

A "Fault Det Remn
 Sol?" message was
 generated after running
 the 0.04 g/210L cal
 adjust solution.

ML
 5/18/2023

Optical Calibration #1	
SN:	80-00 7373
Agency:	Broward CSO
Date:	05/18/2023
Quadratic Fit:	+/- 0.002g/210L ✓
By:	TDG ML

Page 2/2

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post-Cal) #1	80-00 7373	Broward CSO	05/18/2023	TDG MC

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
<div>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:02</td></tr><tr><td>Control Test</td><td>0.050</td><td>12:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr><tr><td>Control Test</td><td>0.048</td><td>12:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr><tr><td>Control Test</td><td>0.045</td><td>12:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0477</td><td></td></tr><tr><td>Std Dev</td><td>0.0025</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>5.2796</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	12:02	Control Test	0.050	12:03	Air Blank	0.000	12:03	Control Test	0.048	12:04	Air Blank	0.000	12:05	Control Test	0.045	12:05	Air Blank	0.000	12:06	Control Test Stats			Average	0.0477		Std Dev	0.0025		Rel Std Dev(%)	5.2796		<div>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr><tr><td>Control Test</td><td>0.075</td><td>12:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:11</td></tr><tr><td>Control Test</td><td>0.081</td><td>12:12</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:12</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0031</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>3.9001</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	12:08	Control Test	0.079	12:09	Air Blank	0.000	12:10	Control Test	0.075	12:10	Air Blank	0.000	12:11	Control Test	0.081	12:12	Air Blank	0.000	12:12	Control Test Stats			Average	0.0783		Std Dev	0.0031		Rel Std Dev(%)	3.9001		<div>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr><tr><td>Control Test</td><td>0.197</td><td>12:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr><tr><td>Control Test</td><td>0.201</td><td>12:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.1987</td><td></td></tr><tr><td>Std Dev</td><td>0.0021</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.0478</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	12:15	Control Test	0.197	12:15	Air Blank	0.000	12:16	Control Test	0.201	12:17	Air Blank	0.000	12:17	Control Test	0.198	12:18	Air Blank	0.000	12:18	Control Test Stats			Average	0.1987		Std Dev	0.0021		Rel Std Dev(%)	1.0478		<div>DGS BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</div> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:56</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:57</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr><tr><td>Control Test</td><td>0.079</td><td>11:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:59</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	11:56	Control Test	0.078	11:56	Air Blank	0.000	11:57	Control Test	0.078	11:57	Air Blank	0.000	11:58	Control Test	0.079	11:58	Air Blank	0.000	11:59	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:02																																																																																																																																																	
Control Test	0.050	12:03																																																																																																																																																	
Air Blank	0.000	12:03																																																																																																																																																	
Control Test	0.048	12:04																																																																																																																																																	
Air Blank	0.000	12:05																																																																																																																																																	
Control Test	0.045	12:05																																																																																																																																																	
Air Blank	0.000	12:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0025																																																																																																																																																		
Rel Std Dev(%)	5.2796																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:08																																																																																																																																																	
Control Test	0.079	12:09																																																																																																																																																	
Air Blank	0.000	12:10																																																																																																																																																	
Control Test	0.075	12:10																																																																																																																																																	
Air Blank	0.000	12:11																																																																																																																																																	
Control Test	0.081	12:12																																																																																																																																																	
Air Blank	0.000	12:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0031																																																																																																																																																		
Rel Std Dev(%)	3.9001																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:15																																																																																																																																																	
Control Test	0.197	12:15																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test	0.201	12:17																																																																																																																																																	
Air Blank	0.000	12:17																																																																																																																																																	
Control Test	0.198	12:18																																																																																																																																																	
Air Blank	0.000	12:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0021																																																																																																																																																		
Rel Std Dev(%)	1.0478																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:56																																																																																																																																																	
Control Test	0.078	11:56																																																																																																																																																	
Air Blank	0.000	11:57																																																																																																																																																	
Control Test	0.078	11:57																																																																																																																																																	
Air Blank	0.000	11:58																																																																																																																																																	
Control Test	0.079	11:58																																																																																																																																																	
Air Blank	0.000	11:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

Comments:

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007373
05/18/2023 12:19:41

Auto Calibration
Max Power Res Value = 89
Auto Range Res Value = 66

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12650, 9um lo = 12625

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1620 (-0.0200)
Sample #2 = 0.1340 (0.0240)
Sample #3 = 0.1100 (0.0480)
Sample #4 = 0.1100 (0.0630)
Avg % Abs = 0.1180 (0.0450)
STD DEV = 0.0139 (0.0197)
REL STD DEV = 11.743 (43.716)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1060 (-0.0010)
Sample #2 = 0.1040 (0.0060)
Sample #3 = -0.2400 (0.1200)
Sample #4 = 0.0820 (-0.0890)
Avg % Abs = -0.0180 (0.0123)
STD DEV = 0.1926 (0.1046)
REL STD DEV = 1069.845 (848.464)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12639, 9um lo = 12616

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8620 (-0.0160)
Sample #2 = 0.8240 (0.0060)
Sample #3 = 0.8450 (0.0010)
Sample #4 = 0.8620 (0.0000)
Avg % Abs = 0.8437 (0.0023)
STD DEV = 0.0190 (0.0032)
REL STD DEV = 2.256 (137.766)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.3790 (-0.0670)
Sample #2 = 1.4140 (0.0760)
Sample #3 = 1.5850 (-0.0690)
Sample #4 = 1.4470 (0.0900)
Avg % Abs = 1.4820 (0.0323)
STD DEV = 0.0907 (0.0880)
REL STD DEV = 6.121 (272.276)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12636, 9um lo = 12590
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8390 (-0.0140)
Sample #2 = 0.8440 (-0.0060)
Sample #3 = 0.8190 (0.0110)
Sample #4 = 0.8260 (0.0060)
Avg % Abs = 0.8297 (0.0037)
STD DEV = 0.0129 (0.0087)
REL STD DEV = 1.554 (238.279)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5420 (-0.1160)
Sample #2 = 1.5840 (-0.1310)
Sample #3 = 1.5100 (-0.2220)
Sample #4 = 1.4830 (-0.1600)
Avg % Abs = 1.5257 (-0.1710)
STD DEV = 0.0523 (0.0465)
REL STD DEV = 3.427 (27.185)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12634, 9um lo = 12601

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8600 (-0.0240)
Sample #2 = 0.8460 (0.0000)
Sample #3 = 0.8490 (-0.0030)
Sample #4 = 0.8210 (0.0180)
Avg % Abs = 0.8387 (0.0050)
STD DEV = 0.0154 (0.0114)
REL STD DEV = 1.833 (227.156)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.4910 (0.0000)
Sample #2 = 1.4910 (0.0260)
Sample #3 = 1.4530 (0.0830)
Sample #4 = 1.4650 (0.0900)
Avg % Abs = 1.4697 (0.0663)
STD DEV = 0.0194 (0.0351)
REL STD DEV = 1.322 (52.922)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12631, 9um lo = 12591

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.9290 (-0.0140)
Sample #2 = 1.8930 (0.0070)
Sample #3 = 1.9230 (-0.0060)
Sample #4 = 1.8910 (0.0160)
Avg % Abs = 1.9023 (0.0057)
STD DEV = 0.0179 (0.0111)
REL STD DEV = 0.942 (195.184)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.4200 (0.0110)
Sample #2 = 3.4240 (0.0110)
Sample #3 = 3.4450 (-0.0330)
Sample #4 = 3.4420 (-0.0020)
Avg % Abs = 3.4370 (-0.0080)
STD DEV = 0.0114 (0.0226)
REL STD DEV = 0.330 (282.566)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12628, 9um lo = 12586
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6050 (-0.0100)
Sample #2 = 3.6120 (-0.0010)
Sample #3 = 3.6390 (-0.0130)
Sample #4 = 3.6360 (-0.0180)
Avg % Abs = 3.6290 (-0.0107)
STD DEV = 0.0148 (0.0087)
REL STD DEV = 0.408 (81.908)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.5250 (0.0230)
Sample #2 = 6.5490 (0.0470)
Sample #3 = 6.5430 (0.0690)
Sample #4 = 6.5870 (0.0230)
Avg % Abs = 6.5597 (0.0463)
STD DEV = 0.0239 (0.0230)
REL STD DEV = 0.364 (49.656)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um lo = 12628, 9um lo = 12581
Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.2640 (-0.0120)
Sample #2 = 5.2790 (0.0020)
Sample #3 = 5.2580 (0.0150)
Sample #4 = 5.2360 (0.0220)
Avg % Abs = 5.2577 (0.0130)
STD DEV = 0.0215 (0.0101)
REL STD DEV = 0.409 (78.068)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 9.4930 (-0.0380)
Sample #2 = 9.4400 (0.0360)
Sample #3 = 9.4900 (0.0580)
Sample #4 = 9.5100 (0.0280)
Avg % Abs = 9.4800 (0.0407)
STD DEV = 0.0361 (0.0155)
REL STD DEV = 0.380 (38.201)

***** AUTO CAL DATA *****

Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.118
Std Dev = 0.01 Rel Std Dev = 11.74
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.839
Std Dev = 0.02 Rel Std Dev = 1.83
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.902
Std Dev = 0.02 Rel Std Dev = 0.94
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.629
Std Dev = 0.01 Rel Std Dev = 0.41
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.258
Std Dev = 0.02 Rel Std Dev = 0.41
Zero Order Coef = -292.81
First Order Coef = 2578.31
Second Order Coef = 36.68
Standard Deviation = 16.856998

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = -0.018
Std Dev = 0.19 Rel Std Dev = 1069.84
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.470
Std Dev = 0.02 Rel Std Dev = 1.32
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.437
Std Dev = 0.01 Rel Std Dev = 0.33
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.560
Std Dev = 0.02 Rel Std Dev = 0.36
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.480
Std Dev = 0.04 Rel Std Dev = 0.38
Zero Order Coef = -12.56
First Order Coef = 1315.73
Second Order Coef = 20.44
Standard Deviation = 38.445946

Optical Calibration #2	
SN:	80-00 7373
Agency:	Broward CSO
Date:	05/18/2023
Quadratic Fit:	+/- 0.002g/210L ✓
By:	TDG [Signature]

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0003
0.040	0.040	0.0002
0.100	0.100	0.0004
0.200	0.200	-0.0005
0.300	0.300	0.0002

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.001	0.0008
0.040	0.041	-0.0013
0.100	0.100	0.0002
0.200	0.199	0.0005
0.300	0.300	-0.0002

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 2817.00
 Sample #2 = 2804.00
 Sample #3 = 2836.00
 Sample #4 = 2851.00
 Average Result = 2830.3333
 STD DEV = 24.0069
 REL STD DEV = 0.848

 ***** CHANNEL 2
 Sample #1 = 3246.00
 Sample #2 = 3223.00
 Sample #3 = 3269.00
 Sample #4 = 3281.00
 Average Result = 3257.6667
 STD DEV = 30.6159
 REL STD DEV = 0.940

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1015
 3 um H2O Adjust (mg/l*10,000) = 979
 9 um H2O Adjust (mg/l*10,000) = 552
 ***** AUTO CAL PASS

Several "Fault Det Remn Sol?"
 messages were generated while
 running the 0.04 g/210L cal
 adjust solution.

ML
 5/18/2023

Optical Calibration #2	
SN:	80-00 7373
Agency:	Broward CSD
Date:	05/18/2023
Quadratic Fit:	+/- 0.002g/210L ✓
By:	TDG ML

Page 2/2

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post-Cal #2)	80-007373	Broward CSU	05/18/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ✓	0.077 to 0.083 ✗	0.194 to 0.206 ✓	0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr><tr><td>Control Test</td><td>0.049</td><td>13:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr><tr><td>Control Test</td><td>0.047</td><td>13:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:26</td></tr><tr><td>Control Test</td><td>0.050</td><td>13:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:27</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0487</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>3.1388</td><td></td></tr></table> <p>Operator's Signature <i>MB</i></p>	Test	g/210L	Time	Air Blank	0.000	13:24	Control Test	0.049	13:24	Air Blank	0.000	13:25	Control Test	0.047	13:26	Air Blank	0.000	13:26	Control Test	0.050	13:27	Air Blank	0.000	13:27	Control Test Stats			Average	0.0487		Std Dev	0.0015		Rel Std Dev(%)	3.1388		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr><tr><td>Control Test</td><td>0.085</td><td>14:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr><tr><td>Control Test</td><td>0.076</td><td>14:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr><tr><td>Control Test</td><td>0.087</td><td>14:11</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0827</td><td></td></tr><tr><td>Std Dev</td><td>0.0059</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>7.0881</td><td></td></tr></table> <p>Operator's Signature <i>MB</i></p>	Test	g/210L	Time	Air Blank	0.000	14:08	Control Test	0.085	14:09	Air Blank	0.000	14:09	Control Test	0.076	14:10	Air Blank	0.000	14:11	Control Test	0.087	14:11	Air Blank	0.000	14:12	Control Test Stats			Average	0.0827		Std Dev	0.0059		Rel Std Dev(%)	7.0881		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:42</td></tr><tr><td>Control Test</td><td>0.196</td><td>14:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:43</td></tr><tr><td>Control Test</td><td>0.195</td><td>14:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:45</td></tr><tr><td>Control Test</td><td>0.196</td><td>14:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:46</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.1957</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr></table> <p>Operator's Signature <i>MB</i></p>	Test	g/210L	Time	Air Blank	0.000	14:42	Control Test	0.196	14:43	Air Blank	0.000	14:43	Control Test	0.195	14:44	Air Blank	0.000	14:45	Control Test	0.196	14:45	Air Blank	0.000	14:46	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951		<p><i>DGS</i></p> <p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007373 05/18/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr><tr><td>Control Test</td><td>0.081</td><td>13:32</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:32</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:33</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></table> <p>Operator's Signature <i>MB</i></p>	Test	g/210L	Time	Air Blank	0.000	13:31	Control Test	0.080	13:31	Air Blank	0.000	13:31	Control Test	0.081	13:32	Air Blank	0.000	13:32	Control Test	0.080	13:33	Air Blank	0.000	13:33	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:24																																																																																																																																																	
Control Test	0.049	13:24																																																																																																																																																	
Air Blank	0.000	13:25																																																																																																																																																	
Control Test	0.047	13:26																																																																																																																																																	
Air Blank	0.000	13:26																																																																																																																																																	
Control Test	0.050	13:27																																																																																																																																																	
Air Blank	0.000	13:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	3.1388																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:08																																																																																																																																																	
Control Test	0.085	14:09																																																																																																																																																	
Air Blank	0.000	14:09																																																																																																																																																	
Control Test	0.076	14:10																																																																																																																																																	
Air Blank	0.000	14:11																																																																																																																																																	
Control Test	0.087	14:11																																																																																																																																																	
Air Blank	0.000	14:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0827																																																																																																																																																		
Std Dev	0.0059																																																																																																																																																		
Rel Std Dev(%)	7.0881																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:42																																																																																																																																																	
Control Test	0.196	14:43																																																																																																																																																	
Air Blank	0.000	14:43																																																																																																																																																	
Control Test	0.195	14:44																																																																																																																																																	
Air Blank	0.000	14:45																																																																																																																																																	
Control Test	0.196	14:45																																																																																																																																																	
Air Blank	0.000	14:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1957																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2951																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test	0.080	13:31																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test	0.081	13:32																																																																																																																																																	
Air Blank	0.000	13:32																																																																																																																																																	
Control Test	0.080	13:33																																																																																																																																																	
Air Blank	0.000	13:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

Comments:

80-007373
Extra Stabilities
MG

0.05 Solution

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007373
05/18/2023
Software: 8100.27

0.04 Solution

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007373
05/18/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:36
Control Test	0.040	13:37
Air Blank	0.000	13:38
Control Test	0.042	13:38
Air Blank	0.000	13:39
Control Test	0.041	13:39
Air Blank	0.000	13:40
Control Test	0.039	13:41
Air Blank	0.000	13:41
Control Test	0.041	13:42
Air Blank	0.000	13:42
Control Test	0.042	13:43
Air Blank	0.000	13:44
Control Test	0.044	13:44
Air Blank	0.000	13:45
Control Test	0.041	13:46
Air Blank	0.000	13:46
Control Test	0.044	13:47
Air Blank	0.000	13:47
Control Test	0.041	13:48
Air Blank	0.000	13:49
Control Test	0.042	13:49
Air Blank	0.000	13:50
Control Test	0.039	13:50
Air Blank	0.000	13:51
Control Test	0.041	13:52
Air Blank	0.000	13:52
Control Test	INT*	13:53
Air Blank	0.000	13:53
Control Test	0.040	13:54
Air Blank	AMB**	13:55
Air Blank	0.000	13:55

*Interferent Detect
**Ambient Fail

MG

Operator's Signature

Test	g/210L	Time
Air Blank	0.000	14:15
Control Test	0.051	14:16
Air Blank	0.000	14:16
Control Test	0.049	14:17
Air Blank	0.000	14:17
Control Test	0.054	14:18
Air Blank	0.000	14:19
Control Test	0.045	14:19
Air Blank	0.000	14:20
Control Test	0.055	14:20
Air Blank	0.000	14:21
Control Test	0.049	14:22
Air Blank	0.000	14:22
Control Test	0.052	14:23
Air Blank	0.000	14:23
Control Test	0.055	14:24
Air Blank	0.000	14:25
Control Test	0.052	14:25
Air Blank	0.000	14:26
Control Test	0.047	14:27
Air Blank	0.000	14:27
Control Test	0.053	14:28
Air Blank	0.000	14:28
Control Test	0.049	14:29
Air Blank	0.000	14:30
Control Test	0.049	14:30
Air Blank	0.000	14:31
Control Test	0.050	14:32
Air Blank	0.000	14:32
Control Test	0.048	14:33
Air Blank	0.000	14:33
Control Test	0.056	14:34
Air Blank	0.000	14:35
Control Test	0.049	14:35
Air Blank	0.000	14:36
Control Test	0.049	14:36
Air Blank	0.000	14:37
Control Test	0.049	14:38
Air Blank	0.000	14:38
Control Test	0.050	14:38
Air Blank	0.000	14:39
Control Test Stats		
Average	0.0506	
Std Dev	0.0029	
Rel Std Dev(%)	5.7218	

MG

Operator's Signature