



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol, Troop LS/N 80-007458Florida Department of
Law EnforcementDate In 06/07/2023DI Completion Date 06/09/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>06/08/2023</u>	Flow Calibration	By	Date																																							
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>207</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.476</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP4864</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP4864	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024	<table border="1"><thead><tr><th>Maintenance</th><th>By</th></tr></thead><tbody><tr><td><input type="checkbox"/> Battery Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Other _____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr></tbody></table>			Maintenance	By	<input type="checkbox"/> Battery Replacement		<input type="checkbox"/> Dry Gas Regulator Replacement		<input type="checkbox"/> Breath Tube Replacement		<input type="checkbox"/> Other _____		_____		_____		_____		_____		_____					
Simulator	Serial #	Lot #/Exp																																												
0.050	MP6286	202201C 01/11/2024																																												
0.080	MP4864	202201D 01/18/2024																																												
0.200	MP6288	202201E 01/18/2024																																												
0.080 DGS	N/A	AG223802 08/26/2024																																												
Maintenance	By																																													
<input type="checkbox"/> Battery Replacement																																														
<input type="checkbox"/> Dry Gas Regulator Replacement																																														
<input type="checkbox"/> Breath Tube Replacement																																														
<input type="checkbox"/> Other _____																																														

Calibration Adjustment		By _____		Department Inspection			By TDG _____																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>26932</u> Gauge <u>1007</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP5092																																													
Interferent	MP5093																																													
0.050	MP5094																																													
0.080	MP5095																																													
0.200	MP5096																																													
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																						
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Notes/Suggested Service: <u>Added plastic caps for return to agency. (TDG)</u> _____ _____ _____ _____ _____ _____ _____ _____ _____				Israel Soto <small>Digitally signed by Israel Soto Date: 2023.06.12 10:47:30 +0000</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.06.13 10:17:22 -04'00'</small> Tech Review / Date _____ Admin Review / Date _____																																										

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007458	FHP Troop L	06/08/2023	TDG NB

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 06/08/2023 SN 80-007458 Software: 8100.27</div> <div>Test g/210L Time</div> <table><tr><td>Air Blank</td><td>0.000</td><td>13:59</td></tr><tr><td>Control Test</td><td>0.048</td><td>13:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:00</td></tr><tr><td>Control Test</td><td>0.049</td><td>14:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:01</td></tr><tr><td>Control Test</td><td>0.048</td><td>14:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0483</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr></table>	Air Blank	0.000	13:59	Control Test	0.048	13:59	Air Blank	0.000	14:00	Control Test	0.049	14:01	Air Blank	0.000	14:01	Control Test	0.048	14:02	Air Blank	0.000	14:02	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 06/08/2023 SN 80-007458 Software: 8100.27</div> <div>Test g/210L Time</div> <table><tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr><tr><td>Control Test</td><td>0.078</td><td>14:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr><tr><td>Control Test</td><td>0.079</td><td>14:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr><tr><td>Control Test</td><td>0.078</td><td>14:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></table>	Air Blank	0.000	14:06	Control Test	0.078	14:07	Air Blank	0.000	14:07	Control Test	0.079	14:08	Air Blank	0.000	14:08	Control Test	0.078	14:09	Air Blank	0.000	14:10	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 06/08/2023 SN 80-007458 Software: 8100.27</div> <div>Test g/210L Time</div> <table><tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr><tr><td>Control Test</td><td>0.200</td><td>14:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr><tr><td>Control Test</td><td>0.199</td><td>14:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr><tr><td>Control Test</td><td>0.199</td><td>14:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1993</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2896</td><td></td></tr></table>	Air Blank	0.000	14:12	Control Test	0.200	14:13	Air Blank	0.000	14:13	Control Test	0.199	14:14	Air Blank	0.000	14:14	Control Test	0.199	14:15	Air Blank	0.000	14:16	Control Test Stats			Average	0.1993		Std Dev	0.0006		Rel Std Dev(%)	0.2896		<div>FHP TROOP L Intoxilyzer - Alcohol Analyzer Model 8000 06/08/2023 SN 80-007458 Software: 8100.27</div> <div>Test g/210L Time</div> <table><tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr><tr><td>Control Test</td><td>0.080</td><td>14:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr><tr><td>Control Test</td><td>0.080</td><td>14:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr><tr><td>Control Test</td><td>0.080</td><td>14:18</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:19</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table>	Air Blank	0.000	14:16	Control Test	0.080	14:17	Air Blank	0.000	14:17	Control Test	0.080	14:17	Air Blank	0.000	14:18	Control Test	0.080	14:18	Air Blank	0.000	14:19	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	13:59																																																																																																																																					
Control Test	0.048	13:59																																																																																																																																					
Air Blank	0.000	14:00																																																																																																																																					
Control Test	0.049	14:01																																																																																																																																					
Air Blank	0.000	14:01																																																																																																																																					
Control Test	0.048	14:02																																																																																																																																					
Air Blank	0.000	14:02																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0483																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1945																																																																																																																																						
Air Blank	0.000	14:06																																																																																																																																					
Control Test	0.078	14:07																																																																																																																																					
Air Blank	0.000	14:07																																																																																																																																					
Control Test	0.079	14:08																																																																																																																																					
Air Blank	0.000	14:08																																																																																																																																					
Control Test	0.078	14:09																																																																																																																																					
Air Blank	0.000	14:10																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0783																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7370																																																																																																																																						
Air Blank	0.000	14:12																																																																																																																																					
Control Test	0.200	14:13																																																																																																																																					
Air Blank	0.000	14:13																																																																																																																																					
Control Test	0.199	14:14																																																																																																																																					
Air Blank	0.000	14:14																																																																																																																																					
Control Test	0.199	14:15																																																																																																																																					
Air Blank	0.000	14:16																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1993																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2896																																																																																																																																						
Air Blank	0.000	14:16																																																																																																																																					
Control Test	0.080	14:17																																																																																																																																					
Air Blank	0.000	14:17																																																																																																																																					
Control Test	0.080	14:17																																																																																																																																					
Air Blank	0.000	14:18																																																																																																																																					
Control Test	0.080	14:18																																																																																																																																					
Air Blank	0.000	14:19																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
<div>NW</div> <div>Operator's Signature</div>	<div>NW</div> <div>Operator's Signature</div>	<div>NW</div> <div>Operator's Signature</div>	<div>NW</div> <div>Operator's Signature</div>																																																																																																																																				

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP L
Time of Inspection: 14:52

Date of Inspection: 06/09/2023

Serial Number: 80-007458
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.078	0.203	0.080
0.000	0.049	0.078	0.203	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.078	0.203	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.079	0.203	0.080
0.000	0.049	0.079	0.203	0.081


Standard Deviations	0.0000	0.0004	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name TAYLOR D GUTSCHOW

06/09/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007458, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007458</u>	UNCERTAINTY* \pm
Owning Agency:	<u>FHP TROOP L</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/09/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:52</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

06/09/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

Service • Integrity • Respect • Quality