



INSTRUMENT PROCESSING SHEET

Agency Florida Fish and Wildlife ConservationS/N 80-007499Florida Department of
Law EnforcementDate In 06-19-2023DI Completion Date 06-19-2023☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date	06-19-2023	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Agency Inspector indicated instrument gives intermittent Ambient Fails during breath tests.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>175</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks																																		
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202303L 03-29-2025</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10-25-2024</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202303L 03-29-2025	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG229803 10-25-2024			<table border="1"><thead><tr><th>Maintenance</th><th>By</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other</td><td></td></tr></tbody></table>		Maintenance	By	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other					
Simulator	Serial #	Lot #/Exp																																		
0.050	MP6291	202201C 01-11-2024																																		
0.080	MP6292	202303L 03-29-2025																																		
0.200	MP6293	202201E 01-18-2024																																		
0.080 DGS	N/A	AG229803 10-25-2024																																		
Maintenance	By																																			
☐ Battery Replacement																																				
☐ Dry Gas Regulator Replacement																																				
☐ Breath Tube Replacement																																				
☐ Other																																				
Calibration Adjustment		By IS		Department Inspection				By IS																												
Barometric Pressure Gauge <u>1012</u> ID # <u>30793</u>				Barometric Pressure ID# <u>28421</u>																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>22460</td><td>12-28-2024</td></tr><tr><td>0.100</td><td>MP6296</td><td>22430</td><td>11-30-2024</td></tr><tr><td>0.200</td><td>MP6297</td><td>22400</td><td>10-12-2024</td></tr><tr><td>0.300</td><td>MP6298</td><td>23070</td><td>03-06-2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>04-05-2025</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	22460	12-28-2024	0.100	MP6296	22430	11-30-2024	0.200	MP6297	22400	10-12-2024	0.300	MP6298	23070	03-06-2025	0.080 DGS	N/A	06723080A5	04-05-2025			Gauge <u>1009</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>				
Simulator	Serial #	Lot #	Expiration																																	
0.000	MP6294	N/A	N/A																																	
0.040	MP6295	22460	12-28-2024																																	
0.100	MP6296	22430	11-30-2024																																	
0.200	MP6297	22400	10-12-2024																																	
0.300	MP6298	23070	03-06-2025																																	
0.080 DGS	N/A	06723080A5	04-05-2025																																	
☒ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293			Attachments														
Simulator	Serial Number																																			
0.000	MP6289																																			
Interferent	MP6290																																			
0.050	MP6291																																			
0.080	MP6292																																			
0.200	MP6293																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C</td><td>01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202303L</td><td>03-29-2025</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E</td><td>01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803</td><td>10-25-2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202201C	01-11-2024	0.080	MP6292	202303L	03-29-2025	0.200	MP6293	202201E	01-18-2024	0.080 DGS	N/A	AG229803	10-25-2024			☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other												
Simulator	Serial #	Lot #	Expiration																																	
0.050	MP6291	202201C	01-11-2024																																	
0.080	MP6292	202303L	03-29-2025																																	
0.200	MP6293	202201E	01-18-2024																																	
0.080 DGS	N/A	AG229803	10-25-2024																																	
Notes/Suggested Service: <u>Performed Optical Bench</u> <u>Calibration Adjustment to address reported intermittent</u> <u>Ambient Fails. IS</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																
				Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2023.06.19 15:48:07 -04'00'				Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.06.19 15:53:42 -04'00'																												
				Tech Review / Date				Admin Review / Date																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 14:13

Date of Inspection: 06/19/2023

Serial Number: 80-007499

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
/ 0.000	0.048	0.079	0.198	0.078
/ 0.000	0.048	0.079	0.198	0.079
/ 0.000	0.048	0.078	0.198	0.078
/ 0.000	0.049	0.079	0.198	0.078
/ 0.000	0.048	0.078	0.198	0.078
/ 0.000	0.049	0.079	0.198	0.079
/ 0.000	0.049	0.079	0.198	0.079
/ 0.000	0.049	0.079	0.198	0.078
/ 0.000	0.049	0.079	0.198	0.078
/ 0.000	0.049	0.079	0.198	0.079

Standard Deviations	0.0005	0.0004	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

00: Ambient Fail MOUTH ALCOHOL TOO CLOSE.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

06/19/2023
Date

Stability checks

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007499
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:36
Control Test	0.049	11:37
Air Blank	0.000	11:37
Control Test	0.049	11:38
Air Blank	0.000	11:39
Control Test	0.048	11:39
Air Blank	0.000	11:40
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	

Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007499
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:41
Control Test	0.080	11:41
Air Blank	0.000	11:42
Control Test	0.079	11:42
Air Blank	0.000	11:43
Control Test	0.079	11:44
Air Blank	0.000	11:44
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007499
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:45
Control Test	0.199	11:46
Air Blank	0.000	11:46
Control Test	0.198	11:47
Air Blank	0.000	11:48
Control Test	0.198	11:48
Air Blank	0.000	11:49
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007499
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:49
Control Test	0.079	11:50
Air Blank	0.000	11:50
Control Test	0.080	11:50
Air Blank	0.000	11:51
Control Test	0.080	11:51
Air Blank	0.000	11:52
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Operator's Signature

wet

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007499 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007499</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/19/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:13</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2023.06.19 15:20:57 -04'00'

06/19/2023

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

FFWCC

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-007499

06/19/2023

11:52:40

Optical Bench
Calibration Adjustment

Auto Calibration

pg 1 of 2

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.1140	(-0.0040)	0.1290	(-0.0230)
Sample #2	0.0900	(0.0440)	0.1030	(0.0130)
Sample #3	0.1060	(0.0600)	0.1070	(0.0230)
Sample #4	0.1100	(0.0790)	0.1200	(0.0170)
Avg % Abs	0.1020	(0.0610)	0.1100	(0.0177)
STD DEV	0.0106	(0.0175)	0.0089	(0.0050)
REL STD DEV	10.375	(28.724)	8.080	(28.490)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8260	(0.0000)	1.4370	(-0.0050)
Sample #2	0.8210	(0.0260)	1.4270	(0.0130)
Sample #3	0.8170	(0.0460)	1.4140	(0.0200)
Sample #4	0.8380	(0.0490)	1.4290	(0.0340)
Avg % Abs	0.8253	(0.0403)	1.4233	(0.0223)
STD DEV	0.0112	(0.0125)	0.0081	(0.0107)
REL STD DEV	1.351	(31.000)	0.572	(47.878)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.9420	(-0.0070)	3.3710	(-0.0040)
Sample #2	1.9070	(0.0360)	3.3470	(0.0170)
Sample #3	1.9140	(0.0470)	3.3710	(0.0110)
Sample #4	1.9170	(0.0500)	3.3780	(0.0170)
Avg % Abs	1.9127	(0.0443)	3.3653	(0.0150)
STD DEV	0.0051	(0.0074)	0.0163	(0.0035)
REL STD DEV	0.268	(16.627)	0.483	(23.094)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.6680	(-0.0090)	6.4110	(-0.0080)
Sample #2	3.6700	(0.0100)	6.3890	(0.0190)
Sample #3	3.6510	(0.0240)	6.3890	(0.0190)
Sample #4	3.6360	(0.0390)	6.3720	(0.0310)
Avg % Abs	3.6523	(0.0243)	6.3833	(0.0230)
STD DEV	0.0170	(0.0145)	0.0098	(0.0069)
REL STD DEV	0.467	(59.601)	0.154	(30.123)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.3610	(-0.0070)	9.3170	(-0.0060)
Sample #2	5.3530	(0.0230)	9.2750	(0.0310)
Sample #3	5.3670	(0.0300)	9.3000	(0.0290)
Sample #4	5.3610	(0.0390)	9.2990	(0.0350)
Avg % Abs	5.3603	(0.0307)	9.2913	(0.0317)
STD DEV	0.0070	(0.0080)	0.0142	(0.0031)
REL STD DEV	0.131	(26.155)	0.152	(9.648)

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007499
06/19/2023 11:52:40

Optical Bench
Calibration Adjustment

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -262.63
First Order Coef 2593.22
Second Order Coef 22.67

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0000
0.040	0.040	0.0002
0.100	0.100	-0.0004
0.200	0.200	0.0003
0.300	0.300	-0.0001

<<<< 9um >>>>

-157.54
1420.99
14.45

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0000
0.040	0.040	0.0002
0.100	0.101	-0.0006
0.200	0.200	0.0005
0.300	0.300	-0.0001

<<<< 3um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 3173.00 3410.00
Sample #2 3147.00 3410.00
Sample #3 3149.00 3389.00
Sample #4 3147.00 3432.00
Avg 3147.6667 3410.3333
STD DEV 1.1547 21.5019
REL STD DEV 0.037 0.630
H2O adjust (mg/l*10k) 662 399

Barometric Pressure = 1011

*****CALIBRATION SUCCESSFUL*****

Post stability checks

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007499
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:29
Control Test	0.049	12:30
Air Blank	0.000	12:30
Control Test	0.048	12:31
Air Blank	0.000	12:31
Control Test	0.048	12:32
Air Blank	0.000	12:33
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007499
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:34
Control Test	0.079	12:34
Air Blank	0.000	12:35
Control Test	0.079	12:36
Air Blank	0.000	12:36
Control Test	0.078	12:37
Air Blank	0.000	12:37
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

wet

Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007499
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:38
Control Test	0.198	12:39
Air Blank	0.000	12:39
Control Test	0.198	12:40
Air Blank	0.000	12:41
Control Test	0.198	12:41
Air Blank	0.000	12:42
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007499
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:42
Control Test	0.079	12:43
Air Blank	0.000	12:43
Control Test	0.078	12:44
Air Blank	0.000	12:44
Control Test	0.079	12:44
Air Blank	0.000	12:45
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Dry

Operator's Signature