

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Miami-Dade Police Department

Serial Number: 80-000881

Time of Inspection: 0354

Date of Inspection: 11/01/2024

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		X
Diagnostic Check (Pre-Inspection): OK		X
Alcohol Free Subject Test: 0.000		X
Mouth Alcohol Test: Slope Not Met		X
Interferent Detect Test: Interferent Detect		X
Diagnostic Check (Post-Inspection): OK		X

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A

Number of Simulators Used: N/A

Remarks: I was unable to enter the inspection mode to conduct an agency inspection because 80-000881's display screen repeatedly went blank. The instrument was taken out of service, and will be sent to CMI for a service evaluation.

AI R. Myrtil present.

The above instrument complies () does not comply (✓) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Ofc. Richard Closius,30-3195

Signature and Printed Name

Date