

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Jacksonville Sheriff's Office

Serial Number: 80-001743

Time of Inspection: 945

Date of Inspection: 01/17/2024

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		NO
Diagnostic Check (Pre-Inspection): OK		NO
Alcohol Free Subject Test: 0.000		NO
Mouth Alcohol Test: Slope Not Met		NO
Interferent Detect Test: Interferent Detect		NO
Diagnostic Check (Post-Inspection): OK		NO

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: 0

Instrument would not come to (Ready Mode). Instrument removed from Evidentiary use at the PTDF. Unable to perform Agency Inspection due to Diagnostic Failure of the Temperature Regulation system in Instrument #80-001743. Notified my Departmental Inspector of Issue and removal of Instrument from Evidentiary use. Sent back to manufacturer(CMI)for inspection and maintenance. Then FDLE for Departmental Inspection.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


69016

Robert D Lightsey

Signature and Printed Name

Date

JACKSONVILLE SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001743
01/17/2024
Software: 8100.27

DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

2 LARRY