FLORIDA DEPARTMENT OF LAW ENFORCEMENT ALCOHOL TESTING PROGRAM BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000 Instrument Registered To: HILLSBOROUGH CO SO Instrument Serial Number: 80-003387 Software: 8100.27

Date of Test: 02/04/2024

Date of Last Agency Inspection: 02/02/2024

Observation Period Began: 04:16

Subject's Name: RAMIPO CRUZ

DOB: 04/11/2001 Sex: M

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time	
	Diagnostics Check	OK	04:48	
	Air Blank	0.000	04:48	
	Control Test	0.081	04:49	
	Air Blank	0.000	04:49	
	Subject Sample #1	0.143	04:51	
	Air Blank	0.000	04:51	
	Air Blank	0.000	04:53	
	Subject Sample #2	0.140	04:54	
	Air Blank	0.000	04:55	
	Control Test	0.081	04:55	
	Air Blank	0.000	04:56	
	Diagnostics Check	OK	04:56	

Cylinder Lot: 08223080A4 Exp: 06/05/2025

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.